

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q1429

Client: Tris Pharma, Inc.

Analytical Method: SW624.1

Datafile : VX045046.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits	RPD
									High	
VX0226WBSD01	Ethyl Acetate	20	23.8	ug/L	119	12		68	140	20
	Isopropyl Acetate	20	21.7	ug/L	109	3		70	130	20
	Acetone	100	110	ug/L	110	17		41	148	20
	Methylene Chloride	20	18.6	ug/L	93	4		60	140	20
	n-amyl Acetate	20	21.1	ug/L	106	7		70	130	20