

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: TRIS02
 Lab Code: CHEM Case No.: Q1429 SAS No.: Q1429 SDG NO.: Q1429
 Lab File ID: VX045044.D Date Analyzed: 02/26/2025
 Instrument ID: MSVOA_X Time Analyzed: 10:05
 GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	19230	4.89	102945	6.75	95036	10.05
UPPER LIMIT	38460	5.385	205890	7.251	190072	10.549
LOWER LIMIT	9615	4.385	51472.5	6.251	47518	9.549
EPA SAMPLE NO.						
OUTFALL-DSN-001	16649	4.89	86265	6.76	78548	10.05
OUTFALL-DSN-002	16856	4.90	90840	6.76	81481	10.06
VX0226WBL01	18408	4.90	98459	6.76	88243	10.06
VX0226WBS01	20343	4.89	110538	6.75	100826	10.05
VX0226WBSD01	16155	4.90	85912	6.76	79610	10.06

IS1 = Bromochloromethane
 IS2 = 1,4-Difluorobenzene
 IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.