

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: GFEL01
 Lab Code: CHEM Case No.: Q1430 SAS No.: Q1430 SDG NO.: Q1430
 Lab File ID: VL042075.D Date Analyzed: 03/03/2025
 Instrument ID: MSVOA_L Time Analyzed: 12:15
 GC Column: RTX-1 ID: 0.32 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	173325	2.78	409365	3.95	350236	8.88
UPPER LIMIT	242655	3.11	573111	4.28	490330	9.21
LOWER LIMIT	103995	2.45	245619	3.62	210142	8.55
EPA SAMPLE NO.						
SV1	160113	2.79	354605	3.96	302162	8.88
SV1DUP	157351	2.78	351741	3.96	297676	8.88
SV2	163018	2.79	364560	3.96	307983	8.88
SV3	161283	2.79	360438	3.96	301876	8.88
SV4	159834	2.79	351160	3.96	298456	8.88
VL0303ABL01	156551	2.79	349743	3.96	302717	8.88
VL0303ABS01	157956	2.79	356972	3.96	309763	8.88

IS1 = Bromochloromethane
 IS2 = 1,4-Difluorobenzene
 IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +40% of internal standard area
 AREA LOWER LIMIT = -40% of internal standard area
 RT UPPER LIMIT = +0.33 minutes of internal standard RT
 RT LOWER LIMIT = -0.33 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.