

### CALIBRATION VERIFICATION SUMMARY

Contract: ATCE02

Lab Code: CHEM Case No.: Q1433 SAS No.: Q1433 SDG NO.: Q1433

Continuing Calib Date: 02/27/2025 Initial Calibration Date(s): 02/20/2025 02/21/2025

Continuing Calib Time: 11:08 Initial Calibration Time(s): 16:46 01:02

GC Column: ZB-MR1 ID: 0.32 (mm)

| COMPOUND             |     | CCAL<br>RT | AVG<br>RT | RT WINDOW |      | DIFF<br>RT |
|----------------------|-----|------------|-----------|-----------|------|------------|
|                      |     |            |           | FROM      | TO   |            |
| Aroclor-1248-1       | (1) | 4.79       | 4.79      | 4.69      | 4.89 | 0.00       |
| Aroclor-1248-2       | (2) | 5.03       | 5.03      | 4.93      | 5.13 | 0.00       |
| Aroclor-1248-3       | (3) | 5.24       | 5.25      | 5.15      | 5.35 | 0.01       |
| Aroclor-1248-4       | (4) | 5.60       | 5.60      | 5.50      | 5.70 | 0.00       |
| Aroclor-1248-5       | (5) | 5.64       | 5.64      | 5.54      | 5.74 | 0.00       |
| Tetrachloro-m-xylene |     | 3.70       | 3.70      | 3.60      | 3.80 | 0.00       |
| Decachlorobiphenyl   |     | 8.76       | 8.76      | 8.66      | 8.86 | 0.00       |

### CALIBRATION VERIFICATION SUMMARY

Contract: ATCE02

Lab Code: CHEM Case No.: Q1433 SAS No.: Q1433 SDG NO.: Q1433

Continuing Calib Date: 02/27/2025 Initial Calibration Date(s): 02/20/2025 02/21/2025

Continuing Calib Time: 11:08 Initial Calibration Time(s): 16:46 01:02

GC Column: ZB-MR2 ID: 0.32 (mm)

| COMPOUND             |     | CCAL<br>RT | AVG<br>RT | RT WINDOW |      | DIFF<br>RT |
|----------------------|-----|------------|-----------|-----------|------|------------|
|                      |     |            |           | FROM      | TO   |            |
| Aroclor-1248-1       | (1) | 4.78       | 4.78      | 4.68      | 4.88 | 0.00       |
| Aroclor-1248-2       | (2) | 5.01       | 5.01      | 4.91      | 5.11 | 0.00       |
| Aroclor-1248-3       | (3) | 5.06       | 5.06      | 4.96      | 5.16 | 0.01       |
| Aroclor-1248-4       | (4) | 5.23       | 5.23      | 5.13      | 5.33 | 0.00       |
| Aroclor-1248-5       | (5) | 5.62       | 5.62      | 5.52      | 5.72 | 0.00       |
| Tetrachloro-m-xylene |     | 3.69       | 3.69      | 3.59      | 3.79 | 0.00       |
| Decachlorobiphenyl   |     | 8.70       | 8.71      | 8.61      | 8.81 | 0.01       |