

CALIBRATION VERIFICATION SUMMARY

Contract: ATCE02

Lab Code: CHEM Case No.: Q1433 SAS No.: Q1433 SDG NO.: Q1433

Continuing Calib Date: 03/11/2025 Initial Calibration Date(s): 02/20/2025 02/21/2025

Continuing Calib Time: 22:43 Initial Calibration Time(s): 16:46 01:02

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1254-1	(1)	5.60	5.60	5.50	5.70	0.01
Aroclor-1254-2	(2)	5.74	5.75	5.65	5.85	0.01
Aroclor-1254-3	(3)	6.15	6.16	6.06	6.26	0.01
Aroclor-1254-4	(4)	6.38	6.39	6.29	6.49	0.01
Aroclor-1254-5	(5)	6.80	6.81	6.71	6.91	0.01
Tetrachloro-m-xylene		3.69	3.70	3.60	3.80	0.01
Decachlorobiphenyl		8.75	8.76	8.66	8.86	0.01

CALIBRATION VERIFICATION SUMMARY

Contract: ATCE02

Lab Code: CHEM Case No.: Q1433 SAS No.: Q1433 SDG NO.: Q1433

Continuing Calib Date: 03/11/2025 Initial Calibration Date(s): 02/20/2025 02/21/2025

Continuing Calib Time: 22:43 Initial Calibration Time(s): 16:46 01:02

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1254-1	(1)	5.58	5.58	5.48	5.68	0.00
Aroclor-1254-2	(2)	5.72	5.73	5.63	5.83	0.01
Aroclor-1254-3	(3)	6.13	6.13	6.03	6.23	0.00
Aroclor-1254-4	(4)	6.35	6.36	6.26	6.46	0.01
Aroclor-1254-5	(5)	6.77	6.77	6.67	6.87	0.00
Tetrachloro-m-xylene		3.69	3.69	3.59	3.79	0.00
Decachlorobiphenyl		8.70	8.71	8.61	8.81	0.01