

CALIBRATION VERIFICATION SUMMARY

Contract: ATCE02

Lab Code: CHEM Case No.: Q1433 SAS No.: Q1433 SDG NO.: Q1433

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 02/20/2025 02/20/2025

Client Sample No.: CCAL10 Date Analyzed: 03/11/2025

Lab Sample No.: AR1660CCC500 Data File : PO109790.D Time Analyzed: 17:17

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	4.787	4.691	4.891	552.860	500.000	10.6
Aroclor-1016-2	4.807	4.711	4.911	558.550	500.000	11.7
Aroclor-1016-3	4.863	4.767	4.967	557.620	500.000	11.5
Aroclor-1016-4	4.984	4.888	5.088	568.430	500.000	13.7
Aroclor-1016-5	5.241	5.145	5.345	573.290	500.000	14.7
Aroclor-1260-1	6.282	6.188	6.388	516.360	500.000	3.3
Aroclor-1260-2	6.471	6.377	6.577	512.850	500.000	2.6
Aroclor-1260-3	6.839	6.746	6.946	505.790	500.000	1.2
Aroclor-1260-4	7.100	7.005	7.205	494.440	500.000	-1.1
Aroclor-1260-5	7.341	7.247	7.447	503.620	500.000	0.7
Decachlorobiphenyl	8.749	8.657	8.857	41.590	50.000	-16.8
Tetrachloro-m-xylene	3.694	3.598	3.798	54.570	50.000	9.1

CALIBRATION VERIFICATION SUMMARY

Contract: ATCE02

Lab Code: CHEM Case No.: Q1433 SAS No.: Q1433 SDG NO.: Q1433

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 02/20/2025 02/20/2025

Client Sample No.: CCAL10 Date Analyzed: 03/11/2025

Lab Sample No.: AR1660CCC500 Data File : PO109790.D Time Analyzed: 17:17

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	4.775	4.678	4.878	591.620	500.000	18.3
Aroclor-1016-2	4.795	4.697	4.897	606.190	500.000	21.2
Aroclor-1016-3	4.970	4.872	5.072	605.710	500.000	21.1
Aroclor-1016-4	5.012	4.914	5.114	572.430	500.000	14.5
Aroclor-1016-5	5.225	5.127	5.327	585.590	500.000	17.1
Aroclor-1260-1	6.258	6.161	6.361	545.930	500.000	9.2
Aroclor-1260-2	6.445	6.348	6.548	555.200	500.000	11.0
Aroclor-1260-3	6.599	6.502	6.702	541.160	500.000	8.2
Aroclor-1260-4	7.071	6.974	7.174	530.450	500.000	6.1
Aroclor-1260-5	7.311	7.214	7.414	537.920	500.000	7.6
Decachlorobiphenyl	8.703	8.607	8.807	44.280	50.000	-11.4
Tetrachloro-m-xylene	3.693	3.594	3.794	58.120	50.000	16.2