

CALIBRATION VERIFICATION SUMMARY

Contract: ATCE02

Lab Code: CHEM Case No.: Q1434 SAS No.: Q1434 SDG NO.: Q1434

Continuing Calib Date: 02/27/2025 Initial Calibration Date(s): 02/20/2025 02/21/2025

Continuing Calib Time: 10:32 Initial Calibration Time(s): 16:46 01:02

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW FROM TO		DIFF RT
Aroclor-1016-1	(1)	4.79	4.79	4.69	4.89	0.00
Aroclor-1016-2	(2)	4.81	4.81	4.71	4.91	0.00
Aroclor-1016-3	(3)	4.87	4.87	4.77	4.97	0.00
Aroclor-1016-4	(4)	4.99	4.99	4.89	5.09	0.00
Aroclor-1016-5	(5)	5.25	5.25	5.15	5.35	0.01
Aroclor-1260-1	(1)	6.29	6.29	6.19	6.39	0.00
Aroclor-1260-2	(2)	6.48	6.48	6.38	6.58	0.01
Aroclor-1260-3	(3)	6.84	6.85	6.75	6.95	0.01
Aroclor-1260-4	(4)	7.10	7.11	7.01	7.21	0.01
Aroclor-1260-5	(5)	7.35	7.35	7.25	7.45	0.01
Tetrachloro-m-xylene		3.70	3.70	3.60	3.80	0.00
Decachlorobiphenyl		8.76	8.76	8.66	8.86	0.00

CALIBRATION VERIFICATION SUMMARY

Contract: ATCE02

Lab Code: CHEM Case No.: Q1434 SAS No.: Q1434 SDG NO.: Q1434

Continuing Calib Date: 02/27/2025 Initial Calibration Date(s): 02/20/2025 02/21/2025

Continuing Calib Time: 10:32 Initial Calibration Time(s): 16:46 01:02

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW FROM TO		DIFF RT
Aroclor-1016-1	(1)	4.78	4.78	4.68	4.88	0.00
Aroclor-1016-2	(2)	4.80	4.80	4.70	4.90	0.00
Aroclor-1016-3	(3)	4.97	4.97	4.87	5.07	0.00
Aroclor-1016-4	(4)	5.01	5.01	4.91	5.11	0.00
Aroclor-1016-5	(5)	5.23	5.23	5.13	5.33	0.00
Aroclor-1260-1	(1)	6.26	6.26	6.16	6.36	0.00
Aroclor-1260-2	(2)	6.45	6.45	6.35	6.55	0.00
Aroclor-1260-3	(3)	6.60	6.60	6.50	6.70	0.00
Aroclor-1260-4	(4)	7.07	7.07	6.97	7.17	0.00
Aroclor-1260-5	(5)	7.31	7.31	7.21	7.41	0.00
Tetrachloro-m-xylene		3.69	3.69	3.59	3.79	0.00
Decachlorobiphenyl		8.71	8.71	8.61	8.81	0.01