

### CALIBRATION VERIFICATION SUMMARY

Contract: ATCE02

Lab Code: CHEM Case No.: Q1434 SAS No.: Q1434 SDG NO.: Q1434

Continuing Calib Date: 03/07/2025 Initial Calibration Date(s): 02/24/2025 02/24/2025

Continuing Calib Time: 15:45 Initial Calibration Time(s): 14:59 22:17

GC Column: ZB-MR1 ID: 0.32 (mm)

| COMPOUND             |     | CCAL<br>RT | AVG<br>RT | RT WINDOW<br>FROM   TO |       | DIFF<br>RT |
|----------------------|-----|------------|-----------|------------------------|-------|------------|
| Aroclor-1016-1       | (1) | 5.68       | 5.68      | 5.58                   | 5.78  | 0.00       |
| Aroclor-1016-2       | (2) | 5.70       | 5.70      | 5.60                   | 5.80  | 0.00       |
| Aroclor-1016-3       | (3) | 5.76       | 5.77      | 5.67                   | 5.87  | 0.01       |
| Aroclor-1016-4       | (4) | 5.86       | 5.86      | 5.76                   | 5.96  | 0.00       |
| Aroclor-1016-5       | (5) | 6.15       | 6.16      | 6.06                   | 6.26  | 0.01       |
| Aroclor-1260-1       | (1) | 7.27       | 7.28      | 7.18                   | 7.38  | 0.01       |
| Aroclor-1260-2       | (2) | 7.53       | 7.53      | 7.43                   | 7.63  | 0.00       |
| Aroclor-1260-3       | (3) | 7.89       | 7.89      | 7.79                   | 7.99  | 0.00       |
| Aroclor-1260-4       | (4) | 8.11       | 8.11      | 8.01                   | 8.21  | 0.00       |
| Aroclor-1260-5       | (5) | 8.43       | 8.43      | 8.33                   | 8.53  | 0.00       |
| Tetrachloro-m-xylene |     | 4.53       | 4.53      | 4.43                   | 4.63  | 0.00       |
| Decachlorobiphenyl   |     | 10.25      | 10.26     | 10.16                  | 10.36 | 0.01       |

### CALIBRATION VERIFICATION SUMMARY

Contract: ATCE02

Lab Code: CHEM Case No.: Q1434 SAS No.: Q1434 SDG NO.: Q1434

Continuing Calib Date: 03/07/2025 Initial Calibration Date(s): 02/24/2025 02/24/2025

Continuing Calib Time: 15:45 Initial Calibration Time(s): 14:59 22:17

GC Column: ZB-MR2 ID: 0.32 (mm)

| COMPOUND             |     | CCAL<br>RT | AVG<br>RT | RT WINDOW<br>FROM   TO |      | DIFF<br>RT |
|----------------------|-----|------------|-----------|------------------------|------|------------|
| Aroclor-1016-1       | (1) | 4.92       | 4.92      | 4.82                   | 5.02 | 0.00       |
| Aroclor-1016-2       | (2) | 4.94       | 4.94      | 4.84                   | 5.04 | 0.00       |
| Aroclor-1016-3       | (3) | 5.11       | 5.12      | 5.02                   | 5.22 | 0.01       |
| Aroclor-1016-4       | (4) | 5.16       | 5.16      | 5.06                   | 5.26 | 0.00       |
| Aroclor-1016-5       | (5) | 5.37       | 5.37      | 5.27                   | 5.47 | 0.00       |
| Aroclor-1260-1       | (1) | 6.41       | 6.41      | 6.31                   | 6.51 | 0.00       |
| Aroclor-1260-2       | (2) | 6.60       | 6.60      | 6.50                   | 6.70 | 0.00       |
| Aroclor-1260-3       | (3) | 6.75       | 6.75      | 6.65                   | 6.85 | 0.00       |
| Aroclor-1260-4       | (4) | 7.22       | 7.23      | 7.13                   | 7.33 | 0.01       |
| Aroclor-1260-5       | (5) | 7.46       | 7.47      | 7.37                   | 7.57 | 0.01       |
| Tetrachloro-m-xylene |     | 3.83       | 3.83      | 3.73                   | 3.93 | 0.00       |
| Decachlorobiphenyl   |     | 8.88       | 8.89      | 8.79                   | 8.99 | 0.01       |