

CALIBRATION VERIFICATION SUMMARY

Contract: ATCE02

Lab Code: CHEM Case No.: Q1434 SAS No.: Q1434 SDG NO.: Q1434

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 02/20/2025 02/20/2025

Client Sample No.: CCAL04 Date Analyzed: 02/28/2025

Lab Sample No.: AR1660CCC500 Data File : PO109600.D Time Analyzed: 18:46

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	4.790	4.691	4.891	536.470	500.000	7.3
Aroclor-1016-2	4.810	4.711	4.911	541.680	500.000	8.3
Aroclor-1016-3	4.866	4.767	4.967	537.630	500.000	7.5
Aroclor-1016-4	4.987	4.888	5.088	539.760	500.000	8.0
Aroclor-1016-5	5.245	5.145	5.345	539.990	500.000	8.0
Aroclor-1260-1	6.287	6.188	6.388	515.020	500.000	3.0
Aroclor-1260-2	6.475	6.377	6.577	506.140	500.000	1.2
Aroclor-1260-3	6.844	6.746	6.946	497.740	500.000	-0.5
Aroclor-1260-4	7.104	7.005	7.205	496.350	500.000	-0.7
Aroclor-1260-5	7.346	7.247	7.447	504.120	500.000	0.8
Decachlorobiphenyl	8.754	8.657	8.857	45.080	50.000	-9.8
Tetrachloro-m-xylene	3.697	3.598	3.798	52.340	50.000	4.7

CALIBRATION VERIFICATION SUMMARY

Contract: ATCE02

Lab Code: CHEM Case No.: Q1434 SAS No.: Q1434 SDG NO.: Q1434

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 02/20/2025 02/20/2025

Client Sample No.: CCAL04 Date Analyzed: 02/28/2025

Lab Sample No.: AR1660CCC500 Data File : PO109600.D Time Analyzed: 18:46

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.777	4.678	4.878	562.880	500.000	12.6
Aroclor-1016-2	4.796	4.697	4.897	578.530	500.000	15.7
Aroclor-1016-3	4.972	4.872	5.072	571.850	500.000	14.4
Aroclor-1016-4	5.014	4.914	5.114	536.120	500.000	7.2
Aroclor-1016-5	5.228	5.127	5.327	564.540	500.000	12.9
Aroclor-1260-1	6.261	6.161	6.361	542.060	500.000	8.4
Aroclor-1260-2	6.448	6.348	6.548	548.120	500.000	9.6
Aroclor-1260-3	6.601	6.502	6.702	534.630	500.000	6.9
Aroclor-1260-4	7.073	6.974	7.174	528.860	500.000	5.8
Aroclor-1260-5	7.313	7.214	7.414	547.960	500.000	9.6
Decachlorobiphenyl	8.706	8.607	8.807	48.890	50.000	-2.2
Tetrachloro-m-xylene	3.695	3.594	3.794	55.890	50.000	11.8