

CALIBRATION VERIFICATION SUMMARY

Contract: ATCE02

Lab Code: CHEM Case No.: Q1434 SAS No.: Q1434 SDG NO.: Q1434

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 02/20/2025 02/20/2025

Client Sample No.: CCAL06 Date Analyzed: 03/03/2025

Lab Sample No.: AR1660CCC500 Data File : PO109622.D Time Analyzed: 09:57

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	4.788	4.691	4.891	519.230	500.000	3.8
Aroclor-1016-2	4.808	4.711	4.911	524.620	500.000	4.9
Aroclor-1016-3	4.864	4.767	4.967	515.470	500.000	3.1
Aroclor-1016-4	4.984	4.888	5.088	517.890	500.000	3.6
Aroclor-1016-5	5.242	5.145	5.345	497.160	500.000	-0.6
Aroclor-1260-1	6.283	6.188	6.388	491.050	500.000	-1.8
Aroclor-1260-2	6.473	6.377	6.577	490.900	500.000	-1.8
Aroclor-1260-3	6.841	6.746	6.946	485.220	500.000	-3.0
Aroclor-1260-4	7.101	7.005	7.205	483.850	500.000	-3.2
Aroclor-1260-5	7.343	7.247	7.447	498.530	500.000	-0.3
Decachlorobiphenyl	8.749	8.657	8.857	42.920	50.000	-14.2
Tetrachloro-m-xylene	3.695	3.598	3.798	49.450	50.000	-1.1

CALIBRATION VERIFICATION SUMMARY

Contract: ATCE02

Lab Code: CHEM Case No.: Q1434 SAS No.: Q1434 SDG NO.: Q1434

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 02/20/2025 02/20/2025

Client Sample No.: CCAL06 Date Analyzed: 03/03/2025

Lab Sample No.: AR1660CCC500 Data File : PO109622.D Time Analyzed: 09:57

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.774	4.678	4.878	557.720	500.000	11.5
Aroclor-1016-2	4.794	4.697	4.897	564.870	500.000	13.0
Aroclor-1016-3	4.970	4.872	5.072	557.960	500.000	11.6
Aroclor-1016-4	5.011	4.914	5.114	533.400	500.000	6.7
Aroclor-1016-5	5.224	5.127	5.327	524.540	500.000	4.9
Aroclor-1260-1	6.257	6.161	6.361	526.620	500.000	5.3
Aroclor-1260-2	6.444	6.348	6.548	535.190	500.000	7.0
Aroclor-1260-3	6.598	6.502	6.702	525.430	500.000	5.1
Aroclor-1260-4	7.069	6.974	7.174	521.840	500.000	4.4
Aroclor-1260-5	7.309	7.214	7.414	548.860	500.000	9.8
Decachlorobiphenyl	8.700	8.607	8.807	46.160	50.000	-7.7
Tetrachloro-m-xylene	3.692	3.594	3.794	53.040	50.000	6.1