

CALIBRATION VERIFICATION SUMMARY

Contract: ATCE02

Lab Code: CHEM Case No.: Q1434 SAS No.: Q1434 SDG NO.: Q1434

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 02/20/2025 02/20/2025

Client Sample No.: CCAL07 Date Analyzed: 03/03/2025

Lab Sample No.: AR1660CCC500 Data File : PO109629.D Time Analyzed: 12:46

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	4.789	4.691	4.891	511.610	500.000	2.3
Aroclor-1016-2	4.808	4.711	4.911	512.590	500.000	2.5
Aroclor-1016-3	4.865	4.767	4.967	503.600	500.000	0.7
Aroclor-1016-4	4.985	4.888	5.088	507.280	500.000	1.5
Aroclor-1016-5	5.242	5.145	5.345	485.810	500.000	-2.8
Aroclor-1260-1	6.284	6.188	6.388	506.570	500.000	1.3
Aroclor-1260-2	6.473	6.377	6.577	496.900	500.000	-0.6
Aroclor-1260-3	6.841	6.746	6.946	489.320	500.000	-2.1
Aroclor-1260-4	7.101	7.005	7.205	490.870	500.000	-1.8
Aroclor-1260-5	7.343	7.247	7.447	502.320	500.000	0.5
Decachlorobiphenyl	8.749	8.657	8.857	42.340	50.000	-15.3
Tetrachloro-m-xylene	3.696	3.598	3.798	47.770	50.000	-4.5

CALIBRATION VERIFICATION SUMMARY

Contract: ATCE02

Lab Code: CHEM Case No.: Q1434 SAS No.: Q1434 SDG NO.: Q1434

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 02/20/2025 02/20/2025

Client Sample No.: CCAL07 Date Analyzed: 03/03/2025

Lab Sample No.: AR1660CCC500 Data File : PO109629.D Time Analyzed: 12:46

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.776	4.678	4.878	553.750	500.000	10.8
Aroclor-1016-2	4.795	4.697	4.897	566.020	500.000	13.2
Aroclor-1016-3	4.971	4.872	5.072	557.770	500.000	11.6
Aroclor-1016-4	5.013	4.914	5.114	487.530	500.000	-2.5
Aroclor-1016-5	5.225	5.127	5.327	523.730	500.000	4.7
Aroclor-1260-1	6.258	6.161	6.361	545.160	500.000	9.0
Aroclor-1260-2	6.446	6.348	6.548	543.930	500.000	8.8
Aroclor-1260-3	6.599	6.502	6.702	532.970	500.000	6.6
Aroclor-1260-4	7.071	6.974	7.174	527.000	500.000	5.4
Aroclor-1260-5	7.311	7.214	7.414	548.740	500.000	9.7
Decachlorobiphenyl	8.702	8.607	8.807	45.750	50.000	-8.5
Tetrachloro-m-xylene	3.694	3.594	3.794	52.200	50.000	4.4