

CALIBRATION VERIFICATION SUMMARY

Contract: ATCE02

Lab Code: CHEM Case No.: Q1436 SAS No.: Q1436 SDG NO.: Q1436

Continuing Calib Date: 03/06/2025 Initial Calibration Date(s): 02/24/2025 02/24/2025

Continuing Calib Time: 23:46 Initial Calibration Time(s): 14:59 22:17

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW FROM TO		DIFF RT
Aroclor-1016-1	(1)	5.68	5.68	5.58	5.78	0.00
Aroclor-1016-2	(2)	5.70	5.70	5.60	5.80	0.00
Aroclor-1016-3	(3)	5.76	5.77	5.67	5.87	0.01
Aroclor-1016-4	(4)	5.86	5.86	5.76	5.96	0.00
Aroclor-1016-5	(5)	6.15	6.16	6.06	6.26	0.01
Aroclor-1260-1	(1)	7.27	7.28	7.18	7.38	0.01
Aroclor-1260-2	(2)	7.53	7.53	7.43	7.63	0.00
Aroclor-1260-3	(3)	7.89	7.89	7.79	7.99	0.00
Aroclor-1260-4	(4)	8.11	8.11	8.01	8.21	0.00
Aroclor-1260-5	(5)	8.43	8.43	8.33	8.53	0.00
Tetrachloro-m-xylene		4.53	4.53	4.43	4.63	0.00
Decachlorobiphenyl		10.26	10.26	10.16	10.36	0.00

CALIBRATION VERIFICATION SUMMARY

Contract: ATCE02

Lab Code: CHEM Case No.: Q1436 SAS No.: Q1436 SDG NO.: Q1436

Continuing Calib Date: 03/06/2025 Initial Calibration Date(s): 02/24/2025 02/24/2025

Continuing Calib Time: 23:46 Initial Calibration Time(s): 14:59 22:17

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW FROM TO		DIFF RT
Aroclor-1016-1	(1)	4.92	4.92	4.82	5.02	0.00
Aroclor-1016-2	(2)	4.94	4.94	4.84	5.04	0.00
Aroclor-1016-3	(3)	5.11	5.12	5.02	5.22	0.01
Aroclor-1016-4	(4)	5.16	5.16	5.06	5.26	0.00
Aroclor-1016-5	(5)	5.37	5.37	5.27	5.47	0.00
Aroclor-1260-1	(1)	6.41	6.41	6.31	6.51	0.00
Aroclor-1260-2	(2)	6.60	6.60	6.50	6.70	0.00
Aroclor-1260-3	(3)	6.75	6.75	6.65	6.85	0.00
Aroclor-1260-4	(4)	7.22	7.23	7.13	7.33	0.01
Aroclor-1260-5	(5)	7.46	7.47	7.37	7.57	0.01
Tetrachloro-m-xylene		3.83	3.83	3.73	3.93	0.00
Decachlorobiphenyl		8.88	8.89	8.79	8.99	0.01