

### CALIBRATION VERIFICATION SUMMARY

Contract: ATCE02

Lab Code: CHEM Case No.: Q1436 SAS No.: Q1436 SDG NO.: Q1436

Continuing Calib Date: 03/07/2025 Initial Calibration Date(s): 02/24/2025 02/24/2025

Continuing Calib Time: 16:34 Initial Calibration Time(s): 14:59 22:17

GC Column: ZB-MR1 ID: 0.32 (mm)

| COMPOUND             |     | CCAL<br>RT | AVG<br>RT | RT WINDOW |       | DIFF<br>RT |
|----------------------|-----|------------|-----------|-----------|-------|------------|
|                      |     |            |           | FROM      | TO    |            |
| Aroclor-1254-1       | (1) | 6.53       | 6.53      | 6.43      | 6.63  | 0.00       |
| Aroclor-1254-2       | (2) | 6.74       | 6.75      | 6.65      | 6.85  | 0.01       |
| Aroclor-1254-3       | (3) | 7.11       | 7.11      | 7.01      | 7.21  | 0.00       |
| Aroclor-1254-4       | (4) | 7.39       | 7.39      | 7.29      | 7.49  | 0.00       |
| Aroclor-1254-5       | (5) | 7.81       | 7.81      | 7.71      | 7.91  | 0.00       |
| Tetrachloro-m-xylene |     | 4.52       | 4.53      | 4.43      | 4.63  | 0.01       |
| Decachlorobiphenyl   |     | 10.25      | 10.26     | 10.16     | 10.36 | 0.01       |

### CALIBRATION VERIFICATION SUMMARY

Contract: ATCE02

Lab Code: CHEM Case No.: Q1436 SAS No.: Q1436 SDG NO.: Q1436

Continuing Calib Date: 03/07/2025 Initial Calibration Date(s): 02/24/2025 02/24/2025

Continuing Calib Time: 16:34 Initial Calibration Time(s): 14:59 22:17

GC Column: ZB-MR2 ID: 0.32 (mm)

| COMPOUND             |     | CCAL<br>RT | AVG<br>RT | RT WINDOW |      | DIFF<br>RT |
|----------------------|-----|------------|-----------|-----------|------|------------|
|                      |     |            |           | FROM      | TO   |            |
| Aroclor-1254-1       | (1) | 5.72       | 5.73      | 5.63      | 5.83 | 0.01       |
| Aroclor-1254-2       | (2) | 5.87       | 5.88      | 5.78      | 5.98 | 0.01       |
| Aroclor-1254-3       | (3) | 6.28       | 6.28      | 6.18      | 6.38 | 0.00       |
| Aroclor-1254-4       | (4) | 6.50       | 6.51      | 6.41      | 6.61 | 0.01       |
| Aroclor-1254-5       | (5) | 6.92       | 6.93      | 6.83      | 7.03 | 0.01       |
| Tetrachloro-m-xylene |     | 3.83       | 3.83      | 3.73      | 3.93 | 0.00       |
| Decachlorobiphenyl   |     | 8.88       | 8.89      | 8.79      | 8.99 | 0.01       |