

### CALIBRATION VERIFICATION SUMMARY

Contract: ATCE02

Lab Code: CHEM Case No.: Q1436 SAS No.: Q1436 SDG NO.: Q1436

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 02/20/2025 02/20/2025

Client Sample No.: CCAL04 Date Analyzed: 02/28/2025

Lab Sample No.: AR1660CCC500 Data File : PO109600.D Time Analyzed: 18:46

| COMPOUND             | RT    | RT WINDOW |       | CALC<br>AMOUNT(ng) | NOM<br>AMOUNT(ng) | %D   |
|----------------------|-------|-----------|-------|--------------------|-------------------|------|
|                      |       | FROM      | TO    |                    |                   |      |
| Aroclor-1016-1       | 4.790 | 4.691     | 4.891 | 536.470            | 500.000           | 7.3  |
| Aroclor-1016-2       | 4.810 | 4.711     | 4.911 | 541.680            | 500.000           | 8.3  |
| Aroclor-1016-3       | 4.866 | 4.767     | 4.967 | 537.630            | 500.000           | 7.5  |
| Aroclor-1016-4       | 4.987 | 4.888     | 5.088 | 539.760            | 500.000           | 8.0  |
| Aroclor-1016-5       | 5.245 | 5.145     | 5.345 | 539.990            | 500.000           | 8.0  |
| Aroclor-1260-1       | 6.287 | 6.188     | 6.388 | 515.020            | 500.000           | 3.0  |
| Aroclor-1260-2       | 6.475 | 6.377     | 6.577 | 506.140            | 500.000           | 1.2  |
| Aroclor-1260-3       | 6.844 | 6.746     | 6.946 | 497.740            | 500.000           | -0.5 |
| Aroclor-1260-4       | 7.104 | 7.005     | 7.205 | 496.350            | 500.000           | -0.7 |
| Aroclor-1260-5       | 7.346 | 7.247     | 7.447 | 504.120            | 500.000           | 0.8  |
| Decachlorobiphenyl   | 8.754 | 8.657     | 8.857 | 45.080             | 50.000            | -9.8 |
| Tetrachloro-m-xylene | 3.697 | 3.598     | 3.798 | 52.340             | 50.000            | 4.7  |

### CALIBRATION VERIFICATION SUMMARY

Contract: ATCE02

Lab Code: CHEM Case No.: Q1436 SAS No.: Q1436 SDG NO.: Q1436

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 02/20/2025 02/20/2025

Client Sample No.: CCAL04 Date Analyzed: 02/28/2025

Lab Sample No.: AR1660CCC500 Data File : PO109600.D Time Analyzed: 18:46

| COMPOUND             | RT    | RT WINDOW |       | CALC<br>AMOUNT(ng) | NOM<br>AMOUNT(ng) | %D   |
|----------------------|-------|-----------|-------|--------------------|-------------------|------|
|                      |       | FROM      | TO    |                    |                   |      |
| Aroclor-1016-1       | 4.777 | 4.678     | 4.878 | 562.880            | 500.000           | 12.6 |
| Aroclor-1016-2       | 4.796 | 4.697     | 4.897 | 578.530            | 500.000           | 15.7 |
| Aroclor-1016-3       | 4.972 | 4.872     | 5.072 | 571.850            | 500.000           | 14.4 |
| Aroclor-1016-4       | 5.014 | 4.914     | 5.114 | 536.120            | 500.000           | 7.2  |
| Aroclor-1016-5       | 5.228 | 5.127     | 5.327 | 564.540            | 500.000           | 12.9 |
| Aroclor-1260-1       | 6.261 | 6.161     | 6.361 | 542.060            | 500.000           | 8.4  |
| Aroclor-1260-2       | 6.448 | 6.348     | 6.548 | 548.120            | 500.000           | 9.6  |
| Aroclor-1260-3       | 6.601 | 6.502     | 6.702 | 534.630            | 500.000           | 6.9  |
| Aroclor-1260-4       | 7.073 | 6.974     | 7.174 | 528.860            | 500.000           | 5.8  |
| Aroclor-1260-5       | 7.313 | 7.214     | 7.414 | 547.960            | 500.000           | 9.6  |
| Decachlorobiphenyl   | 8.706 | 8.607     | 8.807 | 48.890             | 50.000            | -2.2 |
| Tetrachloro-m-xylene | 3.695 | 3.594     | 3.794 | 55.890             | 50.000            | 11.8 |