

CALIBRATION VERIFICATION SUMMARY

Contract: ATCE02

Lab Code: CHEM Case No.: Q1436 SAS No.: Q1436 SDG NO.: Q1436

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 02/20/2025 02/20/2025

Client Sample No.: CCAL05 Date Analyzed: 03/01/2025

Lab Sample No.: AR1660CCC500 Data File : PO109619.D Time Analyzed: 01:30

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	4.790	4.691	4.891	530.850	500.000	6.2
Aroclor-1016-2	4.810	4.711	4.911	536.210	500.000	7.2
Aroclor-1016-3	4.867	4.767	4.967	530.370	500.000	6.1
Aroclor-1016-4	4.987	4.888	5.088	533.410	500.000	6.7
Aroclor-1016-5	5.244	5.145	5.345	511.520	500.000	2.3
Aroclor-1260-1	6.287	6.188	6.388	522.890	500.000	4.6
Aroclor-1260-2	6.475	6.377	6.577	512.790	500.000	2.6
Aroclor-1260-3	6.844	6.746	6.946	505.270	500.000	1.1
Aroclor-1260-4	7.104	7.005	7.205	505.470	500.000	1.1
Aroclor-1260-5	7.346	7.247	7.447	509.860	500.000	2.0
Decachlorobiphenyl	8.754	8.657	8.857	46.440	50.000	-7.1
Tetrachloro-m-xylene	3.697	3.598	3.798	51.630	50.000	3.3

CALIBRATION VERIFICATION SUMMARY

Contract: ATCE02

Lab Code: CHEM Case No.: Q1436 SAS No.: Q1436 SDG NO.: Q1436

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 02/20/2025 02/20/2025

Client Sample No.: CCAL05 Date Analyzed: 03/01/2025

Lab Sample No.: AR1660CCC500 Data File : PO109619.D Time Analyzed: 01:30

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.778	4.678	4.878	557.620	500.000	11.5
Aroclor-1016-2	4.797	4.697	4.897	565.540	500.000	13.1
Aroclor-1016-3	4.973	4.872	5.072	562.850	500.000	12.6
Aroclor-1016-4	5.015	4.914	5.114	513.850	500.000	2.8
Aroclor-1016-5	5.228	5.127	5.327	535.890	500.000	7.2
Aroclor-1260-1	6.261	6.161	6.361	542.240	500.000	8.4
Aroclor-1260-2	6.449	6.348	6.548	547.770	500.000	9.6
Aroclor-1260-3	6.602	6.502	6.702	532.520	500.000	6.5
Aroclor-1260-4	7.074	6.974	7.174	528.040	500.000	5.6
Aroclor-1260-5	7.314	7.214	7.414	547.710	500.000	9.5
Decachlorobiphenyl	8.707	8.607	8.807	49.150	50.000	-1.7
Tetrachloro-m-xylene	3.695	3.594	3.794	54.710	50.000	9.4