

CALIBRATION VERIFICATION SUMMARY

Contract: ATCE02

Lab Code: CHEM Case No.: Q1436 SAS No.: Q1436 SDG NO.: Q1436

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 02/24/2025 02/24/2025

Client Sample No.: CCAL10 Date Analyzed: 03/07/2025

Lab Sample No.: AR1660CCC500 Data File : PP070338.D Time Analyzed: 08:57

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	5.679	5.581	5.781	485.660	500.000	-2.9
Aroclor-1016-2	5.701	5.602	5.802	495.030	500.000	-1.0
Aroclor-1016-3	5.763	5.665	5.865	483.780	500.000	-3.2
Aroclor-1016-4	5.860	5.762	5.962	491.160	500.000	-1.8
Aroclor-1016-5	6.154	6.056	6.256	494.980	500.000	-1.0
Aroclor-1260-1	7.273	7.175	7.375	501.550	500.000	0.3
Aroclor-1260-2	7.527	7.429	7.629	486.050	500.000	-2.8
Aroclor-1260-3	7.886	7.788	7.988	526.040	500.000	5.2
Aroclor-1260-4	8.111	8.012	8.212	511.390	500.000	2.3
Aroclor-1260-5	8.432	8.333	8.533	500.900	500.000	0.2
Decachlorobiphenyl	10.256	10.156	10.356	49.190	50.000	-1.6
Tetrachloro-m-xylene	4.527	4.427	4.627	54.290	50.000	8.6

CALIBRATION VERIFICATION SUMMARY

Contract: ATCE02

Lab Code: CHEM Case No.: Q1436 SAS No.: Q1436 SDG NO.: Q1436

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 02/24/2025 02/24/2025

Client Sample No.: CCAL10 Date Analyzed: 03/07/2025

Lab Sample No.: AR1660CCC500 Data File : PP070338.D Time Analyzed: 08:57

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.917	4.820	5.020	500.370	500.000	0.1
Aroclor-1016-2	4.936	4.839	5.039	505.120	500.000	1.0
Aroclor-1016-3	5.113	5.017	5.217	522.060	500.000	4.4
Aroclor-1016-4	5.155	5.058	5.258	522.480	500.000	4.5
Aroclor-1016-5	5.370	5.273	5.473	514.150	500.000	2.8
Aroclor-1260-1	6.407	6.312	6.512	479.610	500.000	-4.1
Aroclor-1260-2	6.595	6.500	6.700	487.620	500.000	-2.5
Aroclor-1260-3	6.749	6.654	6.854	454.350	500.000	-9.1
Aroclor-1260-4	7.221	7.126	7.326	526.710	500.000	5.3
Aroclor-1260-5	7.462	7.367	7.567	537.250	500.000	7.5
Decachlorobiphenyl	8.882	8.788	8.988	48.280	50.000	-3.4
Tetrachloro-m-xylene	3.829	3.730	3.930	51.020	50.000	2.0