

**CALIBRATION FACTOR OF INITIAL CALIBRATION**

**Contract:** ATCE02

**Lab Code:** CHEM      **Case No.:** Q1436      **SAS No.:** Q1436      **SDG NO.:** Q1436

**Instrument ID:** ECD\_P

**Calibration Date(s):** 02/24/2025      02/24/2025

**Calibration Times:** 14:59      22:17

**GC Column:** ZB-MR1      **ID:** 0.32 (mm)

LAB FILE ID:		CF 1000 =	PP069996.D	CF 750 =	PP069997.D	CF 500 =	PP069998.D	CF 250 =	PP069999.D	CF 050 =	PP070000.D		
COMPOUND		CF 1000	CF 750	CF 500	CF 250	CF 050	CF	% RSD					
Aroclor-1016-1	(1)	45557625	47855512	49968516	54800208	50948980	49826168	7					
Aroclor-1016-2	(2)	66706952	67781012	72991154	76640772	69769920	70777962	6					
Aroclor-1016-3	(3)	40794671	44579197	45146958	47844192	41247180	43922440	7					
Aroclor-1016-4	(4)	33978080	35043389	37559532	39668600	35057220	36261364	6					
Aroclor-1016-5	(5)	31514906	31680175	34068042	35364724	35064900	33538549	5					
Aroclor-1260-1	(1)	54823994	55486535	59388846	62695304	59405260	58359988	6					
Aroclor-1260-2	(2)	73383444	75987972	80712800	85961116	92581060	81725278	9					
Aroclor-1260-3	(3)	59489338	60560020	64011918	67792592	61951460	62761066	5					
Aroclor-1260-4	(4)	58704582	59596489	63141360	67201512	68361020	63400993	7					
Aroclor-1260-5	(5)	124764910	128178828	134805260	140592412	127581700	131184622	5					
Decachlorobiphenyl		1071742430	1139629213	1165555660	1227723520	1090515400	1139033245	5					
Tetrachloro-m-xylene		1433727050	1435859520	1533424440	1590932200	1344174400	1467623522	7					
Aroclor-1242-1	(1)	38359202	40766156	41566668	47963164	43620760	42455190	9					
Aroclor-1242-2	(2)	56757612	55666953	60328714	62664932	52294440	57542530	7					
Aroclor-1242-3	(3)	34692482	36785992	38528310	46539712	33080140	37925327	14					
Aroclor-1242-4	(4)	28738709	28291829	32018112	32372976	27652280	29814781	7					
Aroclor-1242-5	(5)	32588629	32785485	35429746	39255208	36117200	35235254	8					
Decachlorobiphenyl		1021050090	1049520227	1099648880	1125224680	979176800	1054924135	6					
Tetrachloro-m-xylene		1392047690	1350066600	1467421460	1456806320	1242599600	1381788334	7					
Aroclor-1248-1	(1)	29585124	31433301	32559076	36305496	33667780	32710155	8					
Aroclor-1248-2	(2)	39473771	42013749	42864276	46958912	39450360	42152214	7					
Aroclor-1248-3	(3)	43926772	44170904	46948106	50004072	40811360	45172243	8					
Aroclor-1248-4	(4)	53918087	55655911	58640506	66304332	60297700	58963307	8					
Aroclor-1248-5	(5)	51833391	52919351	56894804	67563228	56719440	57186043	11					
Decachlorobiphenyl		1044729320	1067835560	1119982260	1186291160	1031748600	1090117380	6					
Tetrachloro-m-xylene		1393266240	1426265507	1481945040	1527838200	1310175600	1427898117	6					
Aroclor-1254-1	(1)	51156577	57570420	60622162	67008468	47144000	56700325	14					
Aroclor-1254-2	(2)	78872042	81254612	86985854	93340172	97974740	87685484	9					
Aroclor-1254-3	(3)	79973406	82322503	87641586	92940192	102754200	89126377	10					
Aroclor-1254-4	(4)	66856800	68050653	73453394	76814612	87380160	74511124	11					
Aroclor-1254-5	(5)	65315189	68448207	69620926	72356948	60070200	67162294	7					
Decachlorobiphenyl		1075591380	1094765307	1221801800	1207489520	1097348200	1139399241	6					
Tetrachloro-m-xylene		1443804100	1475957520	1555264320	1605313440	1392239200	1494515716	6					
Aroclor-1268-1	(1)	188433469	184749517	193628348	206851508	186804180	192093404	5					

**CALIBRATION FACTOR OF INITIAL CALIBRATION**

Aroclor-1268-2	(2)	162697797	159627413	168307400	177293676	160000200	165585297	4
Aroclor-1268-3	(3)	141621356	137143849	145386650	153273656	140094780	143504058	4
Aroclor-1268-4	(4)	63392961	60284985	64357574	66970196	60240320	63049207	5
Aroclor-1268-5	(5)	414016384	401287087	418333378	444189832	401775920	415920520	4
Decachlorobiphenyl		1839528480	1828279133	1916502300	2031355480	1847859400	1892704959	4
Tetrachloro-m-xylene		1501238100	1376017987	1532711540	1621137720	1417123400	1489645749	7

**CALIBRATION FACTOR OF INITIAL CALIBRATION**

**Contract:** ATCE02

**Lab Code:** CHEM      **Case No.:** Q1436      **SAS No.:** Q1436      **SDG NO.:** Q1436

**Instrument ID:** ECD\_P

**Calibration Date(s):** 02/24/2025      02/24/2025

**Calibration Times:** 14:59      22:17

**GC Column:** ZB-MR2      **ID:** 0.32 (mm)

LAB FILE ID:		CF 1000 =	PP069996.D	CF 750 =	PP069997.D				
		CF 500 =	PP069998.D	CF 250 =	PP069999.D	CF 050 =	PP070000.D		
COMPOUND		CF 1000	CF 750	CF 500	CF 250	CF 050	CF	% RSD	
Aroclor-1016-1	(1)	29885830	30191223	33586918	36186236	37142680	33398577	10	
Aroclor-1016-2	(2)	42247324	42370143	46984138	50510676	50858960	46594248	9	
Aroclor-1016-3	(3)	23325129	22666977	25816242	27757500	25596220	25032414	8	
Aroclor-1016-4	(4)	18421261	17991491	20585360	22152124	21203180	20070683	9	
Aroclor-1016-5	(5)	23822938	23028484	26536406	28616852	27728040	25946544	9	
Aroclor-1260-1	(1)	44977992	45171227	49550340	56779968	51216800	49539265	10	
Aroclor-1260-2	(2)	59688698	59926561	64720792	71640216	71124240	65420101	9	
Aroclor-1260-3	(3)	52251359	51461445	57261004	62988144	77648420	60322074	18	
Aroclor-1260-4	(4)	44361395	45109507	48615116	51524344	54726600	48867392	9	
Aroclor-1260-5	(5)	111484150	113477796	119592622	125243744	126075940	119174850	6	
Decachlorobiphenyl		984571270	1029317907	1118262220	1135932120	1152437600	1084104223	7	
Tetrachloro-m-xylene		879784240	907640400	952717360	1058760600	981262000	956032920	7	
Aroclor-1242-1	(1)	25489035	24475017	27848416	27117444	28407820	26667546	6	
Aroclor-1242-2	(2)	35977820	35199769	38539928	39076044	37212540	37201220	4	
Aroclor-1242-3	(3)	19863858	18273261	20997152	20082888	19211500	19685732	5	
Aroclor-1242-4	(4)	18777771	17922656	19968364	21100536	20505160	19654897	7	
Aroclor-1242-5	(5)	24360224	22407967	26366090	27368000	23855560	24871568	8	
Decachlorobiphenyl		899284040	984619373	1017502000	1027377360	962336000	978223755	5	
Tetrachloro-m-xylene		920809800	854904253	959106780	929536920	894966400	911864831	4	
Aroclor-1248-1	(1)	19339318	20312839	22158310	22596092	21441940	21169700	6	
Aroclor-1248-2	(2)	25841628	26874321	28982946	30984640	28578600	28252427	7	
Aroclor-1248-3	(3)	26866525	28025000	30088164	32810700	30987640	29755606	8	
Aroclor-1248-4	(4)	31739244	33065895	35285822	36808472	38827540	35145395	8	
Aroclor-1248-5	(5)	32923317	34190145	36090288	38291344	39345240	36168067	7	
Decachlorobiphenyl		999826530	1002616133	1003875200	1054961080	1165261200	1045308029	7	
Tetrachloro-m-xylene		903111320	931293867	957405100	986258320	972075400	950028801	3	
Aroclor-1254-1	(1)	50279933	51640439	55752488	60013056	55160920	54569367	7	
Aroclor-1254-2	(2)	44271421	45567105	49578352	53549416	50761380	48745535	8	
Aroclor-1254-3	(3)	70979950	72598121	78881282	84287020	72336540	75816583	7	
Aroclor-1254-4	(4)	50771725	50149405	54249580	57298592	44830820	51460024	9	
Aroclor-1254-5	(5)	64408865	66088107	72696700	75630908	59021560	67569228	10	
Decachlorobiphenyl		963124600	1013018133	1145558760	1245512000	1148271200	1103096939	10	
Tetrachloro-m-xylene		957176240	952284093	991535260	1081903280	978837400	992347255	5	
Aroclor-1268-1	(1)	160284327	157759444	159569126	170516088	180615940	165748985	6	

**CALIBRATION FACTOR OF INITIAL CALIBRATION**

Aroclor-1268-2	(2)	139528496	137224495	137304354	146542344	164608100	145041558	8
Aroclor-1268-3	(3)	118669811	115934707	120470364	129761704	127485360	122464389	5
Aroclor-1268-4	(4)	50956768	50155827	52914114	57814536	54084040	53185057	6
Aroclor-1268-5	(5)	345492428	353062379	344150492	366619864	366772500	355219533	3
Decachlorobiphenyl		1574839450	1670989493	1686779340	1791292760	1882315600	1721243329	7
Tetrachloro-m-xylene		969499550	855552893	995885280	1076293640	1034726400	986391553	8