



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,  
Fax : 908 789 8922

**CALIBRATION VERIFICATION SUMMARY**

**Contract:** ATCE02

**Lab Code:** CHEM      **Case No.:** Q1436      **SAS No.:** Q1436      **SDG NO.:** Q1436

**Continuing Calib Date:** 03/06/2025      **Initial Calibration Date(s):** 02/20/2025      02/21/2025

**Continuing Calib Time:** 11:00      **Initial Calibration Time(s):** 16:46      01:02

**GC Column:** ZB-MR1      **ID:** 0.32      (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
Aroclor-1254-1 (1)	5.60	5.60	5.50	5.70	0.00
Aroclor-1254-2 (2)	5.75	5.75	5.65	5.85	0.00
Aroclor-1254-3 (3)	6.16	6.16	6.06	6.26	0.00
Aroclor-1254-4 (4)	6.39	6.39	6.29	6.49	0.00
Aroclor-1254-5 (5)	6.81	6.81	6.71	6.91	0.00
Tetrachloro-m-xylene	3.70	3.70	3.60	3.80	0.00
Decachlorobiphenyl	8.76	8.76	8.66	8.86	0.00



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,  
Fax : 908 789 8922

**CALIBRATION VERIFICATION SUMMARY**

**Contract:** ATCE02

**Lab Code:** CHEM      **Case No.:** Q1436      **SAS No.:** Q1436      **SDG NO.:** Q1436

**Continuing Calib Date:** 03/06/2025      **Initial Calibration Date(s):** 02/20/2025      02/21/2025

**Continuing Calib Time:** 11:00      **Initial Calibration Time(s):** 16:46      01:02

**GC Column:** ZB-MR2      **ID:** 0.32      (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
Aroclor-1254-1 (1)	5.58	5.58	5.48	5.68	0.00
Aroclor-1254-2 (2)	5.73	5.73	5.63	5.83	0.00
Aroclor-1254-3 (3)	6.13	6.13	6.03	6.23	0.00
Aroclor-1254-4 (4)	6.36	6.36	6.26	6.46	0.00
Aroclor-1254-5 (5)	6.78	6.77	6.67	6.87	-0.01
Tetrachloro-m-xylene	3.70	3.69	3.59	3.79	0.00
Decachlorobiphenyl	8.71	8.71	8.61	8.81	0.00