

CALIBRATION VERIFICATION SUMMARY

Contract: ATCE02

Lab Code: CHEM **Case No.:** Q1436 **SAS No.:** Q1436 **SDG NO.:** Q1436

Continuing Calib Date: 03/07/2025 **Initial Calibration Date(s):** 02/24/2025 02/24/2025

Continuing Calib Time: 10:58 **Initial Calibration Time(s):** 14:59 22:17

GC Column: ZB-MR1 **ID:** 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
Aroclor-1016-1 (1)	5.68	5.68	5.58	5.78	0.00
Aroclor-1016-2 (2)	5.70	5.70	5.60	5.80	0.00
Aroclor-1016-3 (3)	5.76	5.77	5.67	5.87	0.01
Aroclor-1016-4 (4)	5.86	5.86	5.76	5.96	0.00
Aroclor-1016-5 (5)	6.15	6.16	6.06	6.26	0.01
Aroclor-1260-1 (1)	7.27	7.28	7.18	7.38	0.01
Aroclor-1260-2 (2)	7.53	7.53	7.43	7.63	0.00
Aroclor-1260-3 (3)	7.88	7.89	7.79	7.99	0.01
Aroclor-1260-4 (4)	8.11	8.11	8.01	8.21	0.00
Aroclor-1260-5 (5)	8.43	8.43	8.33	8.53	0.00
Tetrachloro-m-xylene	4.53	4.53	4.43	4.63	0.00
Decachlorobiphenyl	10.25	10.26	10.16	10.36	0.01

CALIBRATION VERIFICATION SUMMARY

Contract: ATCE02

Lab Code: CHEM **Case No.:** Q1436 **SAS No.:** Q1436 **SDG NO.:** Q1436

Continuing Calib Date: 03/07/2025 **Initial Calibration Date(s):** 02/24/2025 02/24/2025

Continuing Calib Time: 10:58 **Initial Calibration Time(s):** 14:59 22:17

GC Column: ZB-MR2 **ID:** 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
Aroclor-1016-1 (1)	4.92	4.92	4.82	5.02	0.00
Aroclor-1016-2 (2)	4.94	4.94	4.84	5.04	0.01
Aroclor-1016-3 (3)	5.11	5.12	5.02	5.22	0.01
Aroclor-1016-4 (4)	5.15	5.16	5.06	5.26	0.01
Aroclor-1016-5 (5)	5.37	5.37	5.27	5.47	0.00
Aroclor-1260-1 (1)	6.41	6.41	6.31	6.51	0.00
Aroclor-1260-2 (2)	6.59	6.60	6.50	6.70	0.01
Aroclor-1260-3 (3)	6.75	6.75	6.65	6.85	0.00
Aroclor-1260-4 (4)	7.22	7.23	7.13	7.33	0.01
Aroclor-1260-5 (5)	7.46	7.47	7.37	7.57	0.01
Tetrachloro-m-xylene	3.83	3.83	3.73	3.93	0.00
Decachlorobiphenyl	8.88	8.89	8.79	8.99	0.01