

CALIBRATION VERIFICATION SUMMARY

Contract: ATCE02

Lab Code: CHEM **Case No.:** Q1436 **SAS No.:** Q1436 **SDG NO.:** Q1436

GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 02/20/2025 02/20/2025

Client Sample No.: CCAL03 **Date Analyzed:** 02/28/2025

Lab Sample No.: AR1660CCC500 **Data File :** PO109573.D **Time Analyzed:** 08:19

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.790	4.691	4.891	529.190	500.000	5.8
Aroclor-1016-2	4.810	4.711	4.911	531.530	500.000	6.3
Aroclor-1016-3	4.866	4.767	4.967	525.270	500.000	5.1
Aroclor-1016-4	4.987	4.888	5.088	527.920	500.000	5.6
Aroclor-1016-5	5.245	5.145	5.345	524.260	500.000	4.9
Aroclor-1260-1	6.287	6.188	6.388	490.460	500.000	-1.9
Aroclor-1260-2	6.475	6.377	6.577	480.290	500.000	-3.9
Aroclor-1260-3	6.843	6.746	6.946	459.530	500.000	-8.1
Aroclor-1260-4	7.104	7.005	7.205	451.610	500.000	-9.7
Aroclor-1260-5	7.346	7.247	7.447	455.290	500.000	-8.9
Decachlorobiphenyl	8.753	8.657	8.857	40.920	50.000	-18.2
Tetrachloro-m-xylene	3.697	3.598	3.798	51.950	50.000	3.9

CALIBRATION VERIFICATION SUMMARY

Contract: ATCE02

Lab Code: CHEM Case No.: Q1436 SAS No.: Q1436 SDG NO.: Q1436

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 02/20/2025 02/20/2025

Client Sample No.: CCAL03 Date Analyzed: 02/28/2025

Lab Sample No.: AR1660CCC500 Data File : PO109573.D Time Analyzed: 08:19

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.778	4.678	4.878	577.940	500.000	15.6
Aroclor-1016-2	4.797	4.697	4.897	583.390	500.000	16.7
Aroclor-1016-3	4.973	4.872	5.072	581.840	500.000	16.4
Aroclor-1016-4	5.015	4.914	5.114	541.010	500.000	8.2
Aroclor-1016-5	5.228	5.127	5.327	573.250	500.000	14.7
Aroclor-1260-1	6.261	6.161	6.361	527.680	500.000	5.5
Aroclor-1260-2	6.448	6.348	6.548	528.180	500.000	5.6
Aroclor-1260-3	6.602	6.502	6.702	518.000	500.000	3.6
Aroclor-1260-4	7.073	6.974	7.174	506.710	500.000	1.3
Aroclor-1260-5	7.314	7.214	7.414	519.610	500.000	3.9
Decachlorobiphenyl	8.706	8.607	8.807	46.420	50.000	-7.2
Tetrachloro-m-xylene	3.695	3.594	3.794	56.940	50.000	13.9