

CALIBRATION VERIFICATION SUMMARY

Contract: ATCE02

Lab Code: CHEM Case No.: Q1436 SAS No.: Q1436 SDG NO.: Q1436

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 02/24/2025 02/24/2025

Client Sample No.: CCAL08 Date Analyzed: 03/06/2025

Lab Sample No.: AR1660CCC500 Data File : PP070309.D Time Analyzed: 15:31

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.683	5.581	5.781	435.300	500.000	-12.9
Aroclor-1016-2	5.704	5.602	5.802	450.600	500.000	-9.9
Aroclor-1016-3	5.766	5.665	5.865	429.610	500.000	-14.1
Aroclor-1016-4	5.864	5.762	5.962	417.130	500.000	-16.6
Aroclor-1016-5	6.157	6.056	6.256	416.920	500.000	-16.6
Aroclor-1260-1	7.277	7.175	7.375	418.140	500.000	-16.4
Aroclor-1260-2	7.530	7.429	7.629	398.500	500.000	-20.3
Aroclor-1260-3	7.889	7.788	7.988	421.350	500.000	-15.7
Aroclor-1260-4	8.114	8.012	8.212	413.840	500.000	-17.2
Aroclor-1260-5	8.435	8.333	8.533	433.500	500.000	-13.3
Decachlorobiphenyl	10.263	10.156	10.356	42.190	50.000	-15.6
Tetrachloro-m-xylene	4.529	4.427	4.627	49.550	50.000	-0.9

CALIBRATION VERIFICATION SUMMARY

Contract: ATCE02

Lab Code: CHEM Case No.: Q1436 SAS No.: Q1436 SDG NO.: Q1436

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 02/24/2025 02/24/2025

Client Sample No.: CCAL08 Date Analyzed: 03/06/2025

Lab Sample No.: AR1660CCC500 Data File : PP070309.D Time Analyzed: 15:31

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.919	4.820	5.020	486.150	500.000	-2.8
Aroclor-1016-2	4.938	4.839	5.039	490.170	500.000	-2.0
Aroclor-1016-3	5.115	5.017	5.217	499.410	500.000	-0.1
Aroclor-1016-4	5.157	5.058	5.258	480.910	500.000	-3.8
Aroclor-1016-5	5.372	5.273	5.473	529.400	500.000	5.9
Aroclor-1260-1	6.410	6.312	6.512	421.120	500.000	-15.8
Aroclor-1260-2	6.598	6.500	6.700	442.360	500.000	-11.5
Aroclor-1260-3	6.753	6.654	6.854	418.030	500.000	-16.4
Aroclor-1260-4	7.224	7.126	7.326	426.610	500.000	-14.7
Aroclor-1260-5	7.466	7.367	7.567	466.250	500.000	-6.8
Decachlorobiphenyl	8.886	8.788	8.988	42.280	50.000	-15.4
Tetrachloro-m-xylene	3.830	3.730	3.930	49.120	50.000	-1.8