

CALIBRATION VERIFICATION SUMMARY

Contract: ATCE02

Lab Code: CHEM Case No.: Q1436 SAS No.: Q1436 SDG NO.: Q1436

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 02/24/2025 02/24/2025

Client Sample No.: CCAL09 Date Analyzed: 03/06/2025

Lab Sample No.: AR1660CCC500 Data File : PP070324.D Time Analyzed: 20:20

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.680	5.581	5.781	467.790	500.000	-6.4
Aroclor-1016-2	5.702	5.602	5.802	471.280	500.000	-5.7
Aroclor-1016-3	5.764	5.665	5.865	462.590	500.000	-7.5
Aroclor-1016-4	5.861	5.762	5.962	469.590	500.000	-6.1
Aroclor-1016-5	6.154	6.056	6.256	481.150	500.000	-3.8
Aroclor-1260-1	7.274	7.175	7.375	484.470	500.000	-3.1
Aroclor-1260-2	7.528	7.429	7.629	438.510	500.000	-12.3
Aroclor-1260-3	7.887	7.788	7.988	474.540	500.000	-5.1
Aroclor-1260-4	8.111	8.012	8.212	460.280	500.000	-7.9
Aroclor-1260-5	8.433	8.333	8.533	472.900	500.000	-5.4
Decachlorobiphenyl	10.256	10.156	10.356	45.870	50.000	-8.3
Tetrachloro-m-xylene	4.526	4.427	4.627	51.810	50.000	3.6

CALIBRATION VERIFICATION SUMMARY

Contract: ATCE02

Lab Code: CHEM Case No.: Q1436 SAS No.: Q1436 SDG NO.: Q1436

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 02/24/2025 02/24/2025

Client Sample No.: CCAL09 Date Analyzed: 03/06/2025

Lab Sample No.: AR1660CCC500 Data File : PP070324.D Time Analyzed: 20:20

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.917	4.820	5.020	505.450	500.000	1.1
Aroclor-1016-2	4.935	4.839	5.039	499.910	500.000	0.0
Aroclor-1016-3	5.113	5.017	5.217	521.090	500.000	4.2
Aroclor-1016-4	5.155	5.058	5.258	506.100	500.000	1.2
Aroclor-1016-5	5.370	5.273	5.473	559.510	500.000	11.9
Aroclor-1260-1	6.408	6.312	6.512	486.720	500.000	-2.7
Aroclor-1260-2	6.596	6.500	6.700	484.630	500.000	-3.1
Aroclor-1260-3	6.750	6.654	6.854	450.760	500.000	-9.8
Aroclor-1260-4	7.223	7.126	7.326	509.360	500.000	1.9
Aroclor-1260-5	7.463	7.367	7.567	503.880	500.000	0.8
Decachlorobiphenyl	8.883	8.788	8.988	45.030	50.000	-9.9
Tetrachloro-m-xylene	3.828	3.730	3.930	49.540	50.000	-0.9