

Instrument ID: ECD_O

Daily Analysis Runlog For Sequence/QC Batch ID # PO022025

Review By	yogesh	Review On	2/21/2025 8:09:53 AM
Supervise By	Ankita	Supervise On	2/21/2025 9:30:37 AM
SubDirectory	PO022025	HP Acquire Method	HP Processing Method PO022025

STD. NAME	STD REF.#
Tune/Reschk	
Initial Calibration Stds	PP23735,PP23736,PP23737,PP23738,PP23739,PP23740,PP23741,PP23742,PP23743,PP23744,PP23745,PP23747,PP23748,PP23749,PP23750,PP23751,PP23752,PP23753,PP23754,PP23755,PP23756,PP23757,PP23758,PP23759,PP23760,PP23761,PP23762,PP23763,PP23764,PP23765,PP23766,PP23767,PP23768,PP23769,PP23770,PP23771,PP23772,PP23773,PP23774,PP23775
CCC	PP23737,PP23742,PP23749,PP23754,PP23758,PP23763,PP23768,PP23773
Internal Standard/PEM	
ICV/I.BLK	PP23778,PP23780,PP23783,PP23784,PP23786,PP23788,PP23790,PP23947
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleId	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	HEXANE	HEXANE	PO109424.D	20 Feb 2025 16:09		YP/AJ	Ok
2	I.BLK	I.BLK	PO109425.D	20 Feb 2025 16:28		YP/AJ	Ok
3	AR1660ICC1000	AR1660ICC1000	PO109426.D	20 Feb 2025 16:46		YP/AJ	Ok
4	AR1660ICC750	AR1660ICC750	PO109427.D	20 Feb 2025 17:04		YP/AJ	Ok
5	AR1660ICC500	AR1660ICC500	PO109428.D	20 Feb 2025 17:23		YP/AJ	Ok
6	AR1660ICC250	AR1660ICC250	PO109429.D	20 Feb 2025 17:41		YP/AJ	Ok
7	AR1660ICC050	AR1660ICC050	PO109430.D	20 Feb 2025 17:59		YP/AJ	Ok
8	AR1221ICC500	AR1221ICC500	PO109431.D	20 Feb 2025 18:18		YP/AJ	Ok
9	AR1232ICC500	AR1232ICC500	PO109432.D	20 Feb 2025 18:36		YP/AJ	Ok
10	AR1242ICC1000	AR1242ICC1000	PO109433.D	20 Feb 2025 18:55		YP/AJ	Ok
11	AR1242ICC750	AR1242ICC750	PO109434.D	20 Feb 2025 19:13		YP/AJ	Ok
12	AR1242ICC500	AR1242ICC500	PO109435.D	20 Feb 2025 19:31		YP/AJ	Ok
13	AR1242ICC250	AR1242ICC250	PO109436.D	20 Feb 2025 19:50		YP/AJ	Ok
14	AR1242ICC050	AR1242ICC050	PO109437.D	20 Feb 2025 20:08		YP/AJ	Ok,M
15	AR1248ICC1000	AR1248ICC1000	PO109438.D	20 Feb 2025 20:26		YP/AJ	Ok
16	AR1248ICC750	AR1248ICC750	PO109439.D	20 Feb 2025 20:45		YP/AJ	Ok
17	AR1248ICC500	AR1248ICC500	PO109440.D	20 Feb 2025 21:03		YP/AJ	Ok
18	AR1248ICC250	AR1248ICC250	PO109441.D	20 Feb 2025 21:21		YP/AJ	Ok

Instrument ID: ECD_O

Daily Analysis Runlog For Sequence/QC Batch ID # PO022025

Review By	yogesh	Review On	2/21/2025 8:09:53 AM
Supervise By	Ankita	Supervise On	2/21/2025 9:30:37 AM
SubDirectory	PO022025	HP Acquire Method	HP Processing Method PO022025
STD. NAME	STD REF.#		
Tune/Reschk			
Initial Calibration Stds	PP23735,PP23736,PP23737,PP23738,PP23739,PP23740,PP23741,PP23742,PP23743,PP23744,PP23745,PP23747,PP23748,PP23749,PP23750,P P23751,PP23752,PP23753,PP23754,PP23755,PP23756,PP23757,PP23758,PP23759,PP23760,PP23761,PP23762,PP23763,PP23764,PP23765,PP 23766,PP23767,PP23768,PP23769,PP23770,PP23771,PP23772,PP23773,PP23774,PP23775		
CCC	PP23737,PP23742,PP23749,PP23754,PP23758,PP23763,PP23768,PP23773		
Internal Standard/PEM			
ICV/I.BLK	PP23778,PP23780,PP23783,PP23784,PP23786,PP23788,PP23790,PP23947		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

19	AR1248ICC050	AR1248ICC050	PO109442.D	20 Feb 2025 21:40		YP/AJ	Ok,M
20	AR1254ICC1000	AR1254ICC1000	PO109443.D	20 Feb 2025 21:58		YP/AJ	Ok
21	AR1254ICC750	AR1254ICC750	PO109444.D	20 Feb 2025 22:17		YP/AJ	Ok
22	AR1254ICC500	AR1254ICC500	PO109445.D	20 Feb 2025 22:35		YP/AJ	Ok
23	AR1254ICC250	AR1254ICC250	PO109446.D	20 Feb 2025 22:53		YP/AJ	Ok
24	AR1254ICC050	AR1254ICC050	PO109447.D	20 Feb 2025 23:12		YP/AJ	Ok,M
25	AR1262ICC500	AR1262ICC500	PO109448.D	20 Feb 2025 23:30		YP/AJ	Ok
26	AR1268ICC1000	AR1268ICC1000	PO109449.D	20 Feb 2025 23:48		YP/AJ	Ok
27	AR1268ICC750	AR1268ICC750	PO109450.D	21 Feb 2025 00:07		YP/AJ	Ok
28	AR1268ICC500	AR1268ICC500	PO109451.D	21 Feb 2025 00:25		YP/AJ	Ok
29	AR1268ICC250	AR1268ICC250	PO109452.D	21 Feb 2025 00:43		YP/AJ	Ok
30	AR1268ICC050	AR1268ICC050	PO109453.D	21 Feb 2025 01:02		YP/AJ	Ok
31	PO022025ICV500	ICVPO022025	PO109454.D	21 Feb 2025 01:20		YP/AJ	Ok
32	AR1242ICV500	ICVPO022025AR1242	PO109455.D	21 Feb 2025 01:38		YP/AJ	Ok
33	AR1248ICV500	ICVPO022025AR1248	PO109456.D	21 Feb 2025 01:57		YP/AJ	Ok
34	AR1254ICV500	ICVPO022025AR1254	PO109457.D	21 Feb 2025 02:15		YP/AJ	Ok
35	AR1268ICV500	ICVPO022025AR1268	PO109458.D	21 Feb 2025 02:34		YP/AJ	Ok

M : Manual Integration

Instrument ID: ECD_O

Daily Analysis Runlog For Sequence/QC Batch ID # PO022725

Review By	yogesh	Review On	2/28/2025 12:43:01 PM
Supervise By	Ankita	Supervise On	2/28/2025 12:44:27 PM
SubDirectory	PO022725	HP Acquire Method	HP Processing Method PO022025

STD. NAME	STD REF.#
Tune/Reschk	
Initial Calibration Stds	PP23735,PP23736,PP23737,PP23738,PP23739,PP23740,PP23741,PP23742,PP23743,PP23744,PP23745,PP23747,PP23748,PP23749,PP23750,P P23751,PP23752,PP23753,PP23754,PP23755,PP23756,PP23757,PP23758,PP23759,PP23760,PP23761,PP23762,PP23763,PP23764,PP23765,PP 23766,PP23767,PP23768,PP23769,PP23770,PP23771,PP23772,PP23773,PP23774,PP23775
CCC	PP23737,PP23742,PP23749,PP23754,PP23758,PP23763,PP23768,PP23773
Internal Standard/PEM	
ICV/I.BLK	PP23778,PP23780,PP23783,PP23784,PP23786,PP23788,PP23790,PP23947
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleId	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	HEXANE	HEXANE	PO109517.D	27 Feb 2025 10:13		YP/AJ	Ok
2	AR1660CCC500	AR1660CCC500	PO109518.D	27 Feb 2025 10:32		YP/AJ	Ok
3	AR1242CCC500	AR1242CCC500	PO109519.D	27 Feb 2025 10:50		YP/AJ	Ok
4	AR1248CCC500	AR1248CCC500	PO109520.D	27 Feb 2025 11:08		YP/AJ	Ok
5	AR1254CCC500	AR1254CCC500	PO109521.D	27 Feb 2025 11:27		YP/AJ	Ok
6	I.BLK	I.BLK	PO109522.D	27 Feb 2025 11:45		YP/AJ	Ok
7	PP24209	PP24209	PO109523.D	27 Feb 2025 12:03		YP/AJ	Ok
8	PB166892BL	PB166892BL	PO109524.D	27 Feb 2025 13:16		YP/AJ	Ok
9	PB166892BS	PB166892BS	PO109525.D	27 Feb 2025 13:35		YP/AJ	Ok
10	Q1438-02	K084-15B	PO109526.D	27 Feb 2025 13:53	DCB High in both column, AR1254 + AR1268 Hit, Need Dilution	YP/AJ	Dilution
11	Q1438-04	K084-16B	PO109527.D	27 Feb 2025 14:11	DCB High in both column, AR1254 + AR1268 Hit, Need Dilution	YP/AJ	Dilution
12	Q1438-06	K084-DUP1	PO109528.D	27 Feb 2025 14:30	DCB High in both column, AR1254 + AR1268 Hit, Need Dilution	YP/AJ	Dilution
13	Q1438-07	K084-DUP2	PO109529.D	27 Feb 2025 14:48	AR1254 + AR1268 Hit, Need Dilution	YP/AJ	Dilution
14	Q1438-08	K084-DUP3	PO109530.D	27 Feb 2025 15:06	DCB High in both column, AR1254 + AR1268 Hit, Need Dilution	YP/AJ	Dilution

Instrument ID: ECD_O

Daily Analysis Runlog For Sequence/QC Batch ID # PO022725

Review By	yogesh	Review On	2/28/2025 12:43:01 PM
Supervise By	Ankita	Supervise On	2/28/2025 12:44:27 PM
SubDirectory	PO022725	HP Acquire Method	HP Processing Method PO022025

STD. NAME	STD REF.#
Tune/Reschk	
Initial Calibration Stds	PP23735,PP23736,PP23737,PP23738,PP23739,PP23740,PP23741,PP23742,PP23743,PP23744,PP23745,PP23747,PP23748,PP23749,PP23750,P P23751,PP23752,PP23753,PP23754,PP23755,PP23756,PP23757,PP23758,PP23759,PP23760,PP23761,PP23762,PP23763,PP23764,PP23765,PP 23766,PP23767,PP23768,PP23769,PP23770,PP23771,PP23772,PP23773,PP23774,PP23775
CCC	PP23737,PP23742,PP23749,PP23754,PP23758,PP23763,PP23768,PP23773
Internal Standard/PEM	
ICV/I.BLK	PP23778,PP23780,PP23783,PP23784,PP23786,PP23788,PP23790,PP23947
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

15	Q1438-10	K084-DUP5	PO109531.D	27 Feb 2025 15:25	DCB High in both column, AR1254 + AR1268 Hit, Need Dilution	YP/AJ	Dilution
16	Q1440-01	OR-02-022625	PO109532.D	27 Feb 2025 15:43		YP/AJ	Ok
17	AR1660CCC500	AR1660CCC500	PO109533.D	27 Feb 2025 17:49		YP/AJ	Ok
18	AR1242CCC500	AR1242CCC500	PO109534.D	27 Feb 2025 18:07		YP/AJ	Ok
19	AR1248CCC500	AR1248CCC500	PO109535.D	27 Feb 2025 18:26		YP/AJ	Ok
20	AR1254CCC500	AR1254CCC500	PO109536.D	27 Feb 2025 19:53		YP/AJ	Ok
21	I.BLK	I.BLK	PO109537.D	27 Feb 2025 20:12		YP/AJ	Ok
22	Q1441-01	HD-02-022625	PO109538.D	27 Feb 2025 20:30		YP/AJ	Ok
23	Q1442-01	280	PO109539.D	27 Feb 2025 20:49		YP/AJ	Ok
24	Q1442-02	281	PO109540.D	27 Feb 2025 21:07	AR1260 Hit	YP/AJ	Ok
25	Q1442-03	351	PO109541.D	27 Feb 2025 21:26		YP/AJ	Ok
26	Q1442-03MS	351MS	PO109542.D	27 Feb 2025 21:44		YP/AJ	Ok
27	Q1442-03MSD	351MSD	PO109543.D	27 Feb 2025 22:02		YP/AJ	Ok,M
28	Q1443-01	3733	PO109544.D	27 Feb 2025 22:21		YP/AJ	Ok
29	Q1447-01	WC1	PO109545.D	27 Feb 2025 22:39	AR1260 Hit	YP/AJ	Ok,M
30	PB166889BL	PB166889BL	PO109546.D	27 Feb 2025 22:57		YP/AJ	Ok
31	PB166889BS	PB166889BS	PO109547.D	27 Feb 2025 23:16		YP/AJ	Ok
32	AR1660CCC500	AR1660CCC500	PO109548.D	28 Feb 2025 00:02		YP/AJ	Ok
33	AR1242CCC500	AR1242CCC500	PO109549.D	28 Feb 2025 00:21		YP/AJ	Ok

Instrument ID: ECD_O

Daily Analysis Runlog For Sequence/QC Batch ID # PO022725

Review By	yogesh	Review On	2/28/2025 12:43:01 PM
Supervise By	Ankita	Supervise On	2/28/2025 12:44:27 PM
SubDirectory	PO022725	HP Acquire Method	HP Processing Method PO022025
STD. NAME	STD REF.#		
Tune/Reschk			
Initial Calibration Stds	PP23735,PP23736,PP23737,PP23738,PP23739,PP23740,PP23741,PP23742,PP23743,PP23744,PP23745,PP23747,PP23748,PP23749,PP23750,P P23751,PP23752,PP23753,PP23754,PP23755,PP23756,PP23757,PP23758,PP23759,PP23760,PP23761,PP23762,PP23763,PP23764,PP23765,PP 23766,PP23767,PP23768,PP23769,PP23770,PP23771,PP23772,PP23773,PP23774,PP23775		
CCC	PP23737,PP23742,PP23749,PP23754,PP23758,PP23763,PP23768,PP23773		
Internal Standard/PEM			
ICV/I.BLK	PP23778,PP23780,PP23783,PP23784,PP23786,PP23788,PP23790,PP23947		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

34	AR1248CCC500	AR1248CCC500	PO109550.D	28 Feb 2025 00:38		YP/AJ	Ok
35	AR1254CCC500	AR1254CCC500	PO109551.D	28 Feb 2025 00:57		YP/AJ	Ok
36	I.BLK	I.BLK	PO109552.D	28 Feb 2025 01:14		YP/AJ	Ok
37	Q1433-01	K084-1A	PO109553.D	28 Feb 2025 01:32	DCB High in both column, AR1254 + AR1268 Hit, Need Dilution	YP/AJ	Dilution
38	Q1433-02	K084-1B	PO109554.D	28 Feb 2025 01:50	will be reanalyzed	YP/AJ	Not Ok
39	Q1433-04	K084-2A	PO109555.D	28 Feb 2025 02:08	DCB High in both column, AR1254 + AR1268 Hit, Need Dilution	YP/AJ	Dilution
40	Q1433-05	K084-2B	PO109556.D	28 Feb 2025 02:26	DCB High in both column, AR1254 + AR1268 Hit, Need Dilution	YP/AJ	Dilution
41	Q1433-07	K084-3A	PO109557.D	28 Feb 2025 02:45	DCB High in both column, AR1254 + AR1268 Hit, Need Dilution	YP/AJ	Dilution
42	Q1433-08	K084-3B	PO109558.D	28 Feb 2025 03:03	DCB High in both column, AR1254 + AR1268 Hit, Need Dilution	YP/AJ	Dilution
43	Q1433-10	K084-4A	PO109559.D	28 Feb 2025 03:21	DCB High in both column, AR1254 + AR1268 Hit, Need Dilution	YP/AJ	Dilution
44	Q1434-01	K084-4B	PO109560.D	28 Feb 2025 03:39	AR1254,AR1268 hit Need Dilution	YP/AJ	Dilution
45	Q1434-03	K084-5A	PO109561.D	28 Feb 2025 03:57	AR1254, AR1268 hit Need Dilution	YP/AJ	Dilution
46	Q1434-03MS	K084-5AMS	PO109562.D	28 Feb 2025 04:16	DCB High in both column, Recovery Fail	YP/AJ	Ok
47	Q1434-03MSD	K084-5AMSD	PO109563.D	28 Feb 2025 04:34	DCB High in both column, Recovery Fail	YP/AJ	Ok

Instrument ID: ECD_O

Daily Analysis Runlog For Sequence/QC Batch ID # PO022725

Review By	yogesh	Review On	2/28/2025 12:43:01 PM
Supervise By	Ankita	Supervise On	2/28/2025 12:44:27 PM
SubDirectory	PO022725	HP Acquire Method	HP Processing Method PO022025
STD. NAME	STD REF.#		
Tune/Reschk			
Initial Calibration Stds	PP23735,PP23736,PP23737,PP23738,PP23739,PP23740,PP23741,PP23742,PP23743,PP23744,PP23745,PP23747,PP23748,PP23749,PP23750,P P23751,PP23752,PP23753,PP23754,PP23755,PP23756,PP23757,PP23758,PP23759,PP23760,PP23761,PP23762,PP23763,PP23764,PP23765,PP 23766,PP23767,PP23768,PP23769,PP23770,PP23771,PP23772,PP23773,PP23774,PP23775		
CCC	PP23737,PP23742,PP23749,PP23754,PP23758,PP23763,PP23768,PP23773		
Internal Standard/PEM			
ICV/I.BLK	PP23778,PP23780,PP23783,PP23784,PP23786,PP23788,PP23790,PP23947		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

48	Q1434-04	K084-5B	PO109564.D	28 Feb 2025 04:52	AR1254, AR1268 hit Need Dilution	YP/AJ	Dilution
49	Q1434-06	K084-6B	PO109565.D	28 Feb 2025 05:11	DCB High in both column, AR1254 + AR1268 Hit, Need Dilution	YP/AJ	Dilution
50	Q1434-08	K084-7A	PO109566.D	28 Feb 2025 05:29	DCB High in both column, AR1254 + AR1268 Hit , Need Dilution	YP/AJ	Dilution
51	Q1434-09	K084-7B	PO109567.D	28 Feb 2025 05:46	DCB High in both column, AR1254 + AR1268 Hit , Need Dilution	YP/AJ	Dilution
52	Q1436-01	K084-8A	PO109568.D	28 Feb 2025 06:05	DCB High in both column, AR1254 + AR1268 Hit , Need Dilution	YP/AJ	Dilution
53	Q1436-04	K084-9A	PO109569.D	28 Feb 2025 06:23	DCB High in both column, AR1254 , AR1268 hit Need Dilution	YP/AJ	Dilution
54	Q1436-07	K084-10A	PO109570.D	28 Feb 2025 06:42	DCB High in both column, AR1254, AR1268 hit Need Dilution	YP/AJ	Dilution
55	Q1436-10	K084-11A	PO109571.D	28 Feb 2025 07:00	DCB High in both column, AR1254 + AR1268 Hit , Need Dilution	YP/AJ	Dilution
56	Q1437-03	K084-12A	PO109572.D	28 Feb 2025 07:18	DCB High in both column, AR1254 + AR1268 Hit , Need Dilution	YP/AJ	Dilution
57	AR1660CCC500	AR1660CCC500	PO109573.D	28 Feb 2025 08:19		YP/AJ	Ok
58	I.BLK	I.BLK	PO109574.D	28 Feb 2025 08:38		YP/AJ	Ok,M
59	Q1437-06	K084-13A	PO109575.D	28 Feb 2025 08:55	AR1254 + AR1268 Hit , Need Dilution	YP/AJ	Dilution
60	Q1437-09	K084-14A	PO109576.D	28 Feb 2025 09:14	DCB High in 2nd column,AR1254 + AR1268 Hit , Need Dilution	YP/AJ	Dilution

Instrument ID: ECD_O

Daily Analysis Runlog For Sequence/QC Batch ID # PO022725

Review By	yogesh	Review On	2/28/2025 12:43:01 PM
Supervise By	Ankita	Supervise On	2/28/2025 12:44:27 PM
SubDirectory	PO022725	HP Acquire Method	HP Processing Method PO022025

STD. NAME	STD REF.#
Tune/Reschk	
Initial Calibration Stds	PP23735,PP23736,PP23737,PP23738,PP23739,PP23740,PP23741,PP23742,PP23743,PP23744,PP23745,PP23747,PP23748,PP23749,PP23750,P P23751,PP23752,PP23753,PP23754,PP23755,PP23756,PP23757,PP23758,PP23759,PP23760,PP23761,PP23762,PP23763,PP23764,PP23765,PP 23766,PP23767,PP23768,PP23769,PP23770,PP23771,PP23772,PP23773,PP23774,PP23775
CCC	PP23737,PP23742,PP23749,PP23754,PP23758,PP23763,PP23768,PP23773
Internal Standard/PEM	
ICV/I.BLK	PP23778,PP23780,PP23783,PP23784,PP23786,PP23788,PP23790,PP23947
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

61	AR1660CC500	AR1660CC500	PO109577.D	28 Feb 2025 10:00		YP/AJ	Ok
62	I.BLK	I.BLK	PO109578.D	28 Feb 2025 10:19		YP/AJ	Ok

M : Manual Integration

Instrument ID: ECD_O

Daily Analysis Runlog For Sequence/QC Batch ID # PO022825

Review By	yogesh	Review On	3/3/2025 9:32:13 AM
Supervise By	Ankita	Supervise On	3/4/2025 12:55:41 PM
SubDirectory	PO022825	HP Acquire Method	HP Processing Method PO022025

STD. NAME	STD REF.#
Tune/Reschk	
Initial Calibration Stds	PP23735,PP23736,PP23737,PP23738,PP23739,PP23740,PP23741,PP23742,PP23743,PP23744,PP23745,PP23747,PP23748,PP23749,PP23750,PP23751,PP23752,PP23753,PP23754,PP23755,PP23756,PP23757,PP23758,PP23759,PP23760,PP23761,PP23762,PP23763,PP23764,PP23765,PP23766,PP23767,PP23768,PP23769,PP23770,PP23771,PP23772,PP23773,PP23774,PP23775
CCC	PP23737,PP23742,PP23749,PP23754,PP23758,PP23763,PP23768,PP23773
Internal Standard/PEM	
ICV/I.BLK	PP23778,PP23780,PP23783,PP23784,PP23786,PP23788,PP23790,PP23947
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	AR1660CCC500	AR1660CCC500	PO109579.D	28 Feb 2025 10:57		YP/AJ	Ok
2	I.BLK	I.BLK	PO109580.D	28 Feb 2025 11:16		YP/AJ	Ok
3	Q1438-02DL	K084-15BDL	PO109581.D	28 Feb 2025 11:34	AR1254 + AR1268 Hit, DCB High in both column	YP/AJ	Ok
4	Q1438-04DL	K084-16BDL	PO109582.D	28 Feb 2025 11:52	AR1254 + AR1268 Hit, DCB High in both column	YP/AJ	Ok
5	Q1438-06DL	K084-DUP1DL	PO109583.D	28 Feb 2025 12:10	AR1254 + AR1268 Hit, Need Dilution, all surrogate fail	YP/AJ	Dilution
6	Q1438-06DL2	K084-DUP1DL2	PO109584.D	28 Feb 2025 12:28	AR1254 + AR1268 Hit, all surrogate fail	YP/AJ	Ok,M
7	Q1438-07DL	K084-DUP2DL	PO109585.D	28 Feb 2025 12:47	AR1254 + AR1268 Hit, DCB High in 2nd column, TCMX High in both column	YP/AJ	Ok,M
8	Q1438-08DL	K084-DUP3DL	PO109586.D	28 Feb 2025 13:04	AR1254 + AR1268 Hit, Need Dilution, DCB High in both column	YP/AJ	Dilution
9	Q1438-08DL2	K084-DUP3DL2	PO109587.D	28 Feb 2025 13:22	AR1254 + AR1268 Hit,DCB High in both column	YP/AJ	Ok
10	Q1438-10DL	K084-DUP5DL	PO109588.D	28 Feb 2025 13:41	AR1254 + AR1268 Hit, DCB High in both column, TCMX High in 2nd column	YP/AJ	Ok
11	Q1433-02	K084-1B	PO109589.D	28 Feb 2025 14:13	AR1254 + AR1268 Hit	YP/AJ	Ok
12	Q1433-01DL	K084-1ADL	PO109590.D	28 Feb 2025 14:32	AR1254 + AR1268 Hit, DCB High in both column	YP/AJ	Ok,M
13	Q1433-04DL	K084-2ADL	PO109591.D	28 Feb 2025 15:04	AR1254 + AR1268 Hit, Need Dilution, DCB High in both column	YP/AJ	Dilution
14	Q1433-04DL2	K084-2ADL2	PO109592.D	28 Feb 2025 15:23	AR1254 + AR1268 Hit, DCB High in both column	YP/AJ	Ok

Instrument ID: ECD_O

Daily Analysis Runlog For Sequence/QC Batch ID # PO022825

Review By	yogesh	Review On	3/3/2025 9:32:13 AM
Supervise By	Ankita	Supervise On	3/4/2025 12:55:41 PM
SubDirectory	PO022825	HP Acquire Method	HP Processing Method PO022025
STD. NAME	STD REF.#		
Tune/Reschk			
Initial Calibration Stds	PP23735,PP23736,PP23737,PP23738,PP23739,PP23740,PP23741,PP23742,PP23743,PP23744,PP23745,PP23747,PP23748,PP23749,PP23750,P P23751,PP23752,PP23753,PP23754,PP23755,PP23756,PP23757,PP23758,PP23759,PP23760,PP23761,PP23762,PP23763,PP23764,PP23765,PP 23766,PP23767,PP23768,PP23769,PP23770,PP23771,PP23772,PP23773,PP23774,PP23775		
CCC	PP23737,PP23742,PP23749,PP23754,PP23758,PP23763,PP23768,PP23773		
Internal Standard/PEM			
ICV/I.BLK	PP23778,PP23780,PP23783,PP23784,PP23786,PP23788,PP23790,PP23947		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

15	Q1433-05DL	K084-2BDL	PO109593.D	28 Feb 2025 15:41	AR1254 + AR1268 Hit, Need Dilution	YP/AJ	Dilution
16	Q1433-05DL2	K084-2BDL2	PO109594.D	28 Feb 2025 15:59	AR1254 + AR1268 Hit, DCB High in both column	YP/AJ	Ok,M
17	Q1433-07DL	K084-3ADL	PO109595.D	28 Feb 2025 16:18	AR1254 + AR1268 Hit, Need Dilution	YP/AJ	Dilution
18	Q1433-07DL2	K084-3ADL2	PO109596.D	28 Feb 2025 16:36	AR1254 + AR1268 Hit, All surrogate High,	YP/AJ	Ok
19	Q1433-08DL	K084-3BDL	PO109597.D	28 Feb 2025 16:54	AR1254 + AR1268 Hit, DCB High in both column	YP/AJ	Ok
20	Q1433-10DL	K084-4ADL	PO109598.D	28 Feb 2025 17:13	AR1254 + AR1268 Hit, Need Dilution,DCB High in both column	YP/AJ	Dilution
21	Q1433-10DL2	K084-4ADL2	PO109599.D	28 Feb 2025 17:31	AR1254 + AR1268 Hit,	YP/AJ	Dilution
22	AR1660CCC500	AR1660CCC500	PO109600.D	28 Feb 2025 18:46		YP/AJ	Ok
23	I.BLK	I.BLK	PO109601.D	28 Feb 2025 19:04		YP/AJ	Ok
24	Q1434-01DL	K084-4BDL	PO109602.D	28 Feb 2025 19:22	AR1254,AR1268 hit , DCB High in both column, not match	YP/AJ	Ok
25	Q1434-03DL	K084-5ADL	PO109603.D	28 Feb 2025 19:40	AR1254, AR1268 hit, DCB High in both column	YP/AJ	Ok
26	Q1434-04DL	K084-5BDL	PO109604.D	28 Feb 2025 19:59	AR1254, AR1268 hit	YP/AJ	Ok
27	Q1434-06DL	K084-6BDL	PO109605.D	28 Feb 2025 20:17	DCB High in both column, AR1254 + AR1268 Hit,	YP/AJ	Ok
28	Q1434-08DL	K084-7ADL	PO109606.D	28 Feb 2025 20:35	AR1254 + AR1268 Hit , Need Dilution,Need to confirm Dilution factor, all surrogate fail	YP/AJ	Not Ok
29	Q1434-09DL	K084-7BDL	PO109607.D	28 Feb 2025 20:54	AR1254 + AR1268 Hit , DCB High in both column	YP/AJ	Ok,M
30	Q1436-01DL	K084-8ADL	PO109608.D	28 Feb 2025 21:12	AR1254 + AR1268 Hit , Need Dilution,	YP/AJ	Dilution

Instrument ID: ECD_O

Daily Analysis Runlog For Sequence/QC Batch ID # PO022825

Review By	yogesh	Review On	3/3/2025 9:32:13 AM
Supervise By	Ankita	Supervise On	3/4/2025 12:55:41 PM
SubDirectory	PO022825	HP Acquire Method	HP Processing Method PO022025
STD. NAME	STD REF.#		
Tune/Reschk	PP23735,PP23736,PP23737,PP23738,PP23739,PP23740,PP23741,PP23742,PP23743,PP23744,PP23745,PP23747,PP23748,PP23749,PP23750,P P23751,PP23752,PP23753,PP23754,PP23755,PP23756,PP23757,PP23758,PP23759,PP23760,PP23761,PP23762,PP23763,PP23764,PP23765,PP 23766,PP23767,PP23768,PP23769,PP23770,PP23771,PP23772,PP23773,PP23774,PP23775 PP23737,PP23742,PP23749,PP23754,PP23758,PP23763,PP23768,PP23773 PP23778,PP23780,PP23783,PP23784,PP23786,PP23788,PP23790,PP23947		
Initial Calibration Stds			
CCC			
Internal Standard/PEM			
ICV/I.BLK			
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

31	Q1436-01DL2	K084-8ADL2	PO109609.D	28 Feb 2025 21:30	AR1254 + AR1268 Hit,	YP/AJ	Ok
32	Q1436-04DL	K084-9ADL	PO109610.D	28 Feb 2025 21:49	AR1254 , AR1268 hit, DCB High in both column, not match with original	YP/AJ	Ok
33	Q1436-07DL	K084-10ADL	PO109611.D	28 Feb 2025 22:07	AR1254, AR1268 hit Need Dilution, DCB High in both column	YP/AJ	Dilution
34	Q1436-07DL2	K084-10ADL2	PO109612.D	28 Feb 2025 22:26	AR1254, AR1268 hit,	YP/AJ	Ok
35	Q1436-10DL	K084-11ADL	PO109613.D	28 Feb 2025 22:44	AR1254 + AR1268 Hit , Need Dilution	YP/AJ	Dilution
36	Q1436-10DL2	K084-11ADL2	PO109614.D	28 Feb 2025 23:02	AR1254 + AR1268 Hit	YP/AJ	Ok
37	Q1437-03DL	K084-12ADL	PO109615.D	28 Feb 2025 23:21	AR1254 + AR1268 Hit , Need Dilution, DCB High in both column	YP/AJ	Dilution
38	Q1437-03DL2	K084-12ADL2	PO109616.D	28 Feb 2025 23:39	AR1254 + AR1268 Hit, DCB High in both column	YP/AJ	Ok,M
39	Q1437-06DL	K084-13ADL	PO109617.D	28 Feb 2025 23:58	AR1254 + AR1268 Hit , DCB High in both column	YP/AJ	Ok
40	Q1437-09DL	K084-14ADL	PO109618.D	01 Mar 2025 00:15	DCB High in both column,AR1254 + AR1268 Hit, DCB not match with original, pls confirm	YP/AJ	Ok
41	AR1660CCC500	AR1660CCC500	PO109619.D	01 Mar 2025 01:30		YP/AJ	Ok,M
42	I.BLK	I.BLK	PO109620.D	01 Mar 2025 01:49		YP/AJ	Ok

M : Manual Integration

Instrument ID: ECD_O

Daily Analysis Runlog For Sequence/QC Batch ID # PO030625

Review By	yogesh	Review On	3/6/2025 10:53:50 AM
Supervise By	mohammad	Supervise On	3/8/2025 4:31:13 AM
SubDirectory	PO030625	HP Acquire Method	HP Processing Method PO022025

STD. NAME	STD REF.#
Tune/Reschk	
Initial Calibration Stds	PP23735,PP23736,PP23737,PP23738,PP23739,PP23740,PP23741,PP23742,PP23743,PP23744,PP23745,PP23747,PP23748,PP23749,PP23750,PP23751,PP23752,PP23753,PP23754,PP23755,PP23756,PP23757,PP23758,PP23759,PP23760,PP23761,PP23762,PP23763,PP23764,PP23765,PP23766,PP23767,PP23768,PP23769,PP23770,PP23771,PP23772,PP23773,PP23774,PP23775
CCC	PP23737,PP23742,PP23749,PP23754,PP23758,PP23763,PP23768,PP23773
Internal Standard/PEM	
ICV/I.BLK	PP23778,PP23780,PP23783,PP23784,PP23786,PP23788,PP23790,PP23947
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	HEXANE	HEXANE	PO109647.D	06 Mar 2025 09:28		YP/AJ	Ok
2	AR1660CCC500	AR1660CCC500	PO109648.D	06 Mar 2025 10:05		YP/AJ	Ok,M
3	AR1242CCC500	AR1242CCC500	PO109649.D	06 Mar 2025 10:24		YP/AJ	Ok
4	AR1248CCC500	AR1248CCC500	PO109650.D	06 Mar 2025 10:42		YP/AJ	Ok
5	AR1254CCC500	AR1254CCC500	PO109651.D	06 Mar 2025 11:00		YP/AJ	Ok
6	I.BLK	I.BLK	PO109652.D	06 Mar 2025 11:20		YP/AJ	Ok
7	PB167003BL	PB167003BL	PO109653.D	06 Mar 2025 12:29		YP/AJ	Ok
8	PB167003BS	PB167003BS	PO109654.D	06 Mar 2025 12:47		YP/AJ	Ok
9	Q1498-01	4A-4B-4C-1928 BLD-W	PO109655.D	06 Mar 2025 13:06		YP/AJ	Ok,M
10	Q1498-02	5A-5B-5C-1985 BLDG	PO109656.D	06 Mar 2025 13:24		YP/AJ	Ok,M
11	Q1498-03	6A-6B-6C-1985	PO109657.D	06 Mar 2025 13:43		YP/AJ	Ok
12	Q1498-04	7A-7B-7C-ROOF-5	PO109658.D	06 Mar 2025 14:01		YP/AJ	Ok,M
13	Q1498-05	8A-8B-8C-ROOF-5	PO109659.D	06 Mar 2025 14:20		YP/AJ	Ok
14	Q1498-06	9A-9B-9C-ROOF-6	PO109660.D	06 Mar 2025 14:37		YP/AJ	Ok
15	Q1498-07	10A-10B-10C-ROOF-6	PO109661.D	06 Mar 2025 14:55		YP/AJ	Ok,M
16	AR1660CCC500	AR1660CCC500	PO109662.D	06 Mar 2025 16:11		YP/AJ	Ok,M
17	AR1242CCC500	AR1242CCC500	PO109663.D	06 Mar 2025 16:47		YP/AJ	Ok
18	AR1248CCC500	AR1248CCC500	PO109664.D	06 Mar 2025 17:06		YP/AJ	Ok

Instrument ID: ECD_O

Daily Analysis Runlog For Sequence/QC Batch ID # PO030625

Review By	yogesh	Review On	3/6/2025 10:53:50 AM
Supervise By	mohammad	Supervise On	3/8/2025 4:31:13 AM
SubDirectory	PO030625	HP Acquire Method	HP Processing Method PO022025
STD. NAME	STD REF.#		
Tune/Reschk			
Initial Calibration Stds	PP23735,PP23736,PP23737,PP23738,PP23739,PP23740,PP23741,PP23742,PP23743,PP23744,PP23745,PP23747,PP23748,PP23749,PP23750,P P23751,PP23752,PP23753,PP23754,PP23755,PP23756,PP23757,PP23758,PP23759,PP23760,PP23761,PP23762,PP23763,PP23764,PP23765,PP 23766,PP23767,PP23768,PP23769,PP23770,PP23771,PP23772,PP23773,PP23774,PP23775		
CCC	PP23737,PP23742,PP23749,PP23754,PP23758,PP23763,PP23768,PP23773		
Internal Standard/PEM			
ICV/I.BLK	PP23778,PP23780,PP23783,PP23784,PP23786,PP23788,PP23790,PP23947		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

19	AR1254CCC500	AR1254CCC500	PO109665.D	06 Mar 2025 17:24		YP/AJ	Ok
20	I.BLK	I.BLK	PO109666.D	06 Mar 2025 18:01		YP/AJ	Ok
21	PB167009BL	PB167009BL	PO109667.D	06 Mar 2025 18:19		YP/AJ	Ok
22	PB167009BS	PB167009BS	PO109668.D	06 Mar 2025 18:38		YP/AJ	Ok
23	PB167009BSD	PB167009BSD	PO109669.D	06 Mar 2025 18:56		YP/AJ	Ok
24	Q1488-13	ENV-102-GW01	PO109670.D	06 Mar 2025 19:15		YP/AJ	Ok
25	Q1488-14	ENV-104-GW01	PO109671.D	06 Mar 2025 19:33		YP/AJ	Ok
26	Q1494-01	PURGE-WATER	PO109672.D	06 Mar 2025 19:51		YP/AJ	Ok
27	Q1437-07	K084-13B	PO109673.D	06 Mar 2025 20:10	AR1254 + AR1268 Hit, Need 5x Dilution	YP/AJ	Dilution
28	Q1437-10	K084-14B	PO109674.D	06 Mar 2025 20:28	AR1254 + AR1268 Hit	YP/AJ	Ok,M
29	Q1438-03	K084-15C	PO109675.D	06 Mar 2025 20:47	AR1254 + AR1268 Hit, Need 20x Dilution	YP/AJ	Dilution
30	Q1438-05	K084-16C	PO109676.D	06 Mar 2025 21:04	AR1254 + AR1268 Hit, Need 10x Dilution	YP/AJ	Dilution
31	AR1660CCC500	AR1660CCC500	PO109677.D	06 Mar 2025 22:19		YP/AJ	Ok,M
32	AR1242CCC500	AR1242CCC500	PO109678.D	06 Mar 2025 22:56		YP/AJ	Ok
33	AR1248CCC500	AR1248CCC500	PO109679.D	06 Mar 2025 23:14		YP/AJ	Ok
34	AR1254CCC500	AR1254CCC500	PO109680.D	06 Mar 2025 23:33		YP/AJ	Ok
35	I.BLK	I.BLK	PO109681.D	07 Mar 2025 00:09		YP/AJ	Ok
36	PB167022BL	PB167022BL	PO109682.D	07 Mar 2025 00:28		YP/AJ	Ok

Instrument ID: ECD_O

Daily Analysis Runlog For Sequence/QC Batch ID # PO030625

Review By	yogesh	Review On	3/6/2025 10:53:50 AM
Supervise By	mohammad	Supervise On	3/8/2025 4:31:13 AM
SubDirectory	PO030625	HP Acquire Method	HP Processing Method PO022025
STD. NAME	STD REF.#		
Tune/Reschk	PP23735,PP23736,PP23737,PP23738,PP23739,PP23740,PP23741,PP23742,PP23743,PP23744,PP23745,PP23747,PP23748,PP23749,PP23750,P P23751,PP23752,PP23753,PP23754,PP23755,PP23756,PP23757,PP23758,PP23759,PP23760,PP23761,PP23762,PP23763,PP23764,PP23765,PP 23766,PP23767,PP23768,PP23769,PP23770,PP23771,PP23772,PP23773,PP23774,PP23775 PP23737,PP23742,PP23749,PP23754,PP23758,PP23763,PP23768,PP23773 PP23778,PP23780,PP23783,PP23784,PP23786,PP23788,PP23790,PP23947		
Initial Calibration Stds			
CCC			
Internal Standard/PEM			
ICV/I.BLK			
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

37	PB167022BS	PB167022BS	PO109683.D	07 Mar 2025 00:46		YP/AJ	Ok
38	Q1433-03	K084-1C	PO109684.D	07 Mar 2025 01:05	AR1254 + AR1268 Hit	YP/AJ	Ok,M
39	Q1433-06	K084-2C	PO109685.D	07 Mar 2025 01:23	AR1254 + AR1268 Hit, Need 2x Dilution	YP/AJ	Dilution
40	Q1433-09	K084-3C	PO109686.D	07 Mar 2025 01:41	AR1254 + AR1268 Hit, Need 4x dilution	YP/AJ	Dilution
41	Q1434-02	K084-4C	PO109687.D	07 Mar 2025 02:00	AR1254 + AR1268 Hit, Need 5x dilution	YP/AJ	Dilution
42	Q1434-02MS	K084-4CMS	PO109688.D	07 Mar 2025 02:18	AR1660 Recovery Fai	YP/AJ	Ok,M
43	Q1434-02MSD	K084-4CMSD	PO109689.D	07 Mar 2025 02:37		YP/AJ	Ok,M
44	Q1434-05	K084-5C	PO109690.D	07 Mar 2025 02:55	AR1254 + AR1268 Hit, Need 5x dilution	YP/AJ	Dilution
45	Q1434-07	K084-6C	PO109691.D	07 Mar 2025 03:13	AR1254 + AR1268 Hit, need 4x dilution, DCB High in both column	YP/AJ	Dilution
46	AR1660CCC500	AR1660CCC500	PO109692.D	07 Mar 2025 04:29		YP/AJ	Ok,M
47	AR1242CCC500	AR1242CCC500	PO109693.D	07 Mar 2025 05:05		YP/AJ	Ok
48	AR1248CCC500	AR1248CCC500	PO109694.D	07 Mar 2025 05:24		YP/AJ	Ok
49	AR1254CCC500	AR1254CCC500	PO109695.D	07 Mar 2025 05:42		YP/AJ	Ok,M
50	I.BLK	I.BLK	PO109696.D	07 Mar 2025 06:19		YP/AJ	Ok

M : Manual Integration

Instrument ID: ECD_P

Daily Analysis Runlog For Sequence/QC Batch ID # PP022425

Review By	yogesh	Review On	2/25/2025 9:00:36 AM
Supervise By	Ankita	Supervise On	2/25/2025 9:48:29 AM
SubDirectory	PP022425	HP Acquire Method	HP Processing Method PP022425

STD. NAME	STD REF.#
Tune/Reschk	
Initial Calibration Stds	PP23735,PP23736,PP23737,PP23738,PP23739,PP23740,PP23741,PP23742,PP23743,PP23744,PP23745,PP23747,PP23748,PP23749,PP23750,PP23751,PP23752,PP23753,PP23754,PP23755,PP23756,PP23757,PP23758,PP23759,PP23760,PP23761,PP23762,PP23763,PP23764,PP23765,PP23766,PP23767,PP23768,PP23769,PP23770,PP23771,PP23772,PP23773,PP23774,PP23775
CCC	PP23737,PP23742,PP23749,PP23754,PP23758,PP23763,PP23768,PP23773
Internal Standard/PEM	
ICV/I.BLK	PP23778,PP23780,PP23783,PP23784,PP23786,PP23788,PP23790,PP23947
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	HEXANE	HEXANE	PP069994.D	24 Feb 2025 14:27		YP\AJ	Ok
2	I.BLK	I.BLK	PP069995.D	24 Feb 2025 14:43		YP\AJ	Ok
3	AR1660ICC1000	AR1660ICC1000	PP069996.D	24 Feb 2025 14:59		YP\AJ	Ok
4	AR1660ICC750	AR1660ICC750	PP069997.D	24 Feb 2025 15:15		YP\AJ	Ok
5	AR1660ICC500	AR1660ICC500	PP069998.D	24 Feb 2025 15:32		YP\AJ	Ok
6	AR1660ICC250	AR1660ICC250	PP069999.D	24 Feb 2025 15:48		YP\AJ	Ok
7	AR1660ICC050	AR1660ICC050	PP070000.D	24 Feb 2025 16:04		YP\AJ	Ok,M
8	AR1221ICC500	AR1221ICC500	PP070001.D	24 Feb 2025 16:20		YP\AJ	Ok,M
9	AR1232ICC500	AR1232ICC500	PP070002.D	24 Feb 2025 16:37		YP\AJ	Ok
10	AR1242ICC1000	AR1242ICC1000	PP070003.D	24 Feb 2025 16:53		YP\AJ	Ok
11	AR1242ICC750	AR1242ICC750	PP070004.D	24 Feb 2025 17:09		YP\AJ	Ok,M
12	AR1242ICC500	AR1242ICC500	PP070005.D	24 Feb 2025 17:25		YP\AJ	Ok
13	AR1242ICC250	AR1242ICC250	PP070006.D	24 Feb 2025 17:42		YP\AJ	Ok,M
14	AR1242ICC050	AR1242ICC050	PP070007.D	24 Feb 2025 17:58		YP\AJ	Ok,M
15	AR1248ICC1000	AR1248ICC1000	PP070008.D	24 Feb 2025 18:14		YP\AJ	Ok
16	AR1248ICC750	AR1248ICC750	PP070009.D	24 Feb 2025 18:30		YP\AJ	Ok
17	AR1248ICC500	AR1248ICC500	PP070010.D	24 Feb 2025 18:46		YP\AJ	Ok
18	AR1248ICC250	AR1248ICC250	PP070011.D	24 Feb 2025 19:03		YP\AJ	Ok

Instrument ID: ECD_P

Daily Analysis Runlog For Sequence/QC Batch ID # PP022425

Review By	yogesh	Review On	2/25/2025 9:00:36 AM
Supervise By	Ankita	Supervise On	2/25/2025 9:48:29 AM
SubDirectory	PP022425	HP Acquire Method	HP Processing Method PP022425
STD. NAME	STD REF.#		
Tune/Reschk			
Initial Calibration Stds	PP23735,PP23736,PP23737,PP23738,PP23739,PP23740,PP23741,PP23742,PP23743,PP23744,PP23745,PP23747,PP23748,PP23749,PP23750,P P23751,PP23752,PP23753,PP23754,PP23755,PP23756,PP23757,PP23758,PP23759,PP23760,PP23761,PP23762,PP23763,PP23764,PP23765,PP 23766,PP23767,PP23768,PP23769,PP23770,PP23771,PP23772,PP23773,PP23774,PP23775		
CCC	PP23737,PP23742,PP23749,PP23754,PP23758,PP23763,PP23768,PP23773		
Internal Standard/PEM			
ICV/I.BLK	PP23778,PP23780,PP23783,PP23784,PP23786,PP23788,PP23790,PP23947		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

19	AR1248ICC050	AR1248ICC050	PP070012.D	24 Feb 2025 19:19		YPIAJ	Ok,M
20	AR1254ICC1000	AR1254ICC1000	PP070013.D	24 Feb 2025 19:35		YPIAJ	Ok,M
21	AR1254ICC750	AR1254ICC750	PP070014.D	24 Feb 2025 19:51		YPIAJ	Ok,M
22	AR1254ICC500	AR1254ICC500	PP070015.D	24 Feb 2025 20:08		YPIAJ	Ok
23	AR1254ICC250	AR1254ICC250	PP070016.D	24 Feb 2025 20:24		YPIAJ	Ok
24	AR1254ICC050	AR1254ICC050	PP070017.D	24 Feb 2025 20:40		YPIAJ	Ok,M
25	AR1262ICC500	AR1262ICC500	PP070018.D	24 Feb 2025 20:56		YPIAJ	Ok
26	AR1268ICC1000	AR1268ICC1000	PP070019.D	24 Feb 2025 21:12		YPIAJ	Ok
27	AR1268ICC750	AR1268ICC750	PP070020.D	24 Feb 2025 21:29		YPIAJ	Ok,M
28	AR1268ICC500	AR1268ICC500	PP070021.D	24 Feb 2025 21:45		YPIAJ	Ok
29	AR1268ICC250	AR1268ICC250	PP070022.D	24 Feb 2025 22:01		YPIAJ	Ok,M
30	AR1268ICC050	AR1268ICC050	PP070023.D	24 Feb 2025 22:17		YPIAJ	Ok,M
31	PP022425ICV500	ICVPP022425	PP070024.D	24 Feb 2025 22:34		YPIAJ	Ok
32	AR1242ICV500	ICVPP022425AR1242	PP070025.D	24 Feb 2025 22:50		YPIAJ	Ok
33	AR1248ICV500	ICVPP022425AR1248	PP070026.D	24 Feb 2025 23:06		YPIAJ	Ok
34	AR1254ICV500	ICVPP022425AR1254	PP070027.D	24 Feb 2025 23:22		YPIAJ	Ok
35	AR1268ICV500	ICVPP022425AR1268	PP070028.D	24 Feb 2025 23:38		YPIAJ	Ok

M : Manual Integration

Instrument ID: ECD_P

Daily Analysis Runlog For Sequence/QC Batch ID # PP030625

Review By	yogesh	Review On	3/6/2025 10:54:36 AM
Supervise By	mohammad	Supervise On	3/8/2025 4:33:30 AM
SubDirectory	PP030625	HP Acquire Method	HP Processing Method PP022425
STD. NAME	STD REF.#		
Tune/Reschk			
Initial Calibration Stds	PP23735,PP23736,PP23737,PP23738,PP23739,PP23740,PP23741,PP23742,PP23743,PP23744,PP23745,PP23747,PP23748,PP23749,PP23750,PP23751,PP23752,PP23753,PP23754,PP23755,PP23756,PP23757,PP23758,PP23759,PP23760,PP23761,PP23762,PP23763,PP23764,PP23765,PP23766,PP23767,PP23768,PP23769,PP23770,PP23771,PP23772,PP23773,PP23774,PP23775		
CCC	PP23737,PP23742,PP23749,PP23754,PP23758,PP23763,PP23768,PP23773		
Internal Standard/PEM			
ICV/I.BLK	PP23778,PP23780,PP23783,PP23784,PP23786,PP23788,PP23790,PP23947		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

Sr#	SampleId	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	HEXANE	HEXANE	PP070294.D	06 Mar 2025 09:16		YP\AJ	Ok
2	AR1660CCC500	AR1660CCC500	PP070295.D	06 Mar 2025 09:51		YP\AJ	Ok
3	AR1242CCC500	AR1242CCC500	PP070296.D	06 Mar 2025 10:09		YP\AJ	Ok,M
4	AR1248CCC500	AR1248CCC500	PP070297.D	06 Mar 2025 10:25		YP\AJ	Ok
5	AR1254CCC500	AR1254CCC500	PP070298.D	06 Mar 2025 10:41		YP\AJ	Ok,M
6	I.BLK	I.BLK	PP070299.D	06 Mar 2025 11:02		YP\AJ	Ok
7	PB167010BL	PB167010BL	PP070300.D	06 Mar 2025 12:21		YP\AJ	Ok
8	PB167010BS	PB167010BS	PP070301.D	06 Mar 2025 12:37		YP\AJ	Ok
9	Q1490-01	CLAY-SLUDGE-DRUM	PP070302.D	06 Mar 2025 12:53	AR1260 Hit	YP\AJ	Ok,M
10	Q1494-02	ASPHALT	PP070303.D	06 Mar 2025 13:10		YP\AJ	Ok,M
11	Q1494-03	ASPHALT-SOIL	PP070304.D	06 Mar 2025 13:26		YP\AJ	Ok
12	Q1494-05	SOIL	PP070305.D	06 Mar 2025 13:42		YP\AJ	Ok
13	Q1494-07	SOIL-COMP	PP070306.D	06 Mar 2025 13:59		YP\AJ	Ok
14	Q1497-07MS	Q1497-07MS	PP070307.D	06 Mar 2025 14:15		YP\AJ	Ok,M
15	Q1494-07MSD	SOIL-COMPMSD	PP070308.D	06 Mar 2025 14:31		YP\AJ	Ok,M
16	AR1660CCC500	AR1660CCC500	PP070309.D	06 Mar 2025 15:31		YP\AJ	Ok
17	AR1242CCC500	AR1242CCC500	PP070310.D	06 Mar 2025 15:47		YP\AJ	Ok,M
18	AR1248CCC500	AR1248CCC500	PP070311.D	06 Mar 2025 16:06		YP\AJ	Ok,M

Instrument ID: ECD_P

Daily Analysis Runlog For Sequence/QC Batch ID # PP030625

Review By	yogesh	Review On	3/6/2025 10:54:36 AM
Supervise By	mohammad	Supervise On	3/8/2025 4:33:30 AM
SubDirectory	PP030625	HP Acquire Method	HP Processing Method PP022425
STD. NAME	STD REF.#		
Tune/Reschk			
Initial Calibration Stds	PP23735,PP23736,PP23737,PP23738,PP23739,PP23740,PP23741,PP23742,PP23743,PP23744,PP23745,PP23747,PP23748,PP23749,PP23750,P P23751,PP23752,PP23753,PP23754,PP23755,PP23756,PP23757,PP23758,PP23759,PP23760,PP23761,PP23762,PP23763,PP23764,PP23765,PP 23766,PP23767,PP23768,PP23769,PP23770,PP23771,PP23772,PP23773,PP23774,PP23775		
CCC	PP23737,PP23742,PP23749,PP23754,PP23758,PP23763,PP23768,PP23773		
Internal Standard/PEM			
ICV/I.BLK	PP23778,PP23780,PP23783,PP23784,PP23786,PP23788,PP23790,PP23947		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

19	AR1254CCC500	AR1254CCC500	PP070312.D	06 Mar 2025 16:22		YPIAJ	Ok,M
20	I.BLK	I.BLK	PP070313.D	06 Mar 2025 16:38		YPIAJ	Ok
21	PB167005BL	PB167005BL	PP070314.D	06 Mar 2025 16:55		YPIAJ	Ok
22	PB167005BS	PB167005BS	PP070315.D	06 Mar 2025 17:11		YPIAJ	Ok,M
23	Q1472-01	40308	PP070316.D	06 Mar 2025 17:27	TCMX High in both column, NEED clean up	YPIAJ	Not Ok
24	PB167022BL	PB167022BL	PP070317.D	06 Mar 2025 17:43		YPIAJ	Ok
25	Q1434-10	K084-7C	PP070318.D	06 Mar 2025 18:00	AR1254 + AR1268 Hit, Need dilution 2X	YPIAJ	Dilution
26	Q1436-02	K084-8B	PP070319.D	06 Mar 2025 18:16	AR1254 + AR1268 Hit, Need dilution 2X	YPIAJ	Dilution
27	Q1436-03	K084-8C	PP070320.D	06 Mar 2025 18:32	AR1254 + AR1268 Hit	YPIAJ	Ok,M
28	Q1436-05	K084-9B	PP070321.D	06 Mar 2025 18:48	AR1254 + AR1268 Hit,	YPIAJ	Ok,M
29	Q1436-06	K084-9C	PP070322.D	06 Mar 2025 19:05	AR1254 + AR1268 Hit,	YPIAJ	Ok,M
30	Q1436-08	K084-10B	PP070323.D	06 Mar 2025 19:21	AR1254 + AR1268 Hit, Need dilution , 2X,10X DCB High in both column	YPIAJ	Dilution
31	AR1660CCC500	AR1660CCC500	PP070324.D	06 Mar 2025 20:20		YPIAJ	Ok
32	AR1242CCC500	AR1242CCC500	PP070325.D	06 Mar 2025 20:53		YPIAJ	Ok,M
33	AR1248CCC500	AR1248CCC500	PP070326.D	06 Mar 2025 21:09		YPIAJ	Ok,M
34	AR1254CCC500	AR1254CCC500	PP070327.D	06 Mar 2025 21:25		YPIAJ	Ok,M
35	I.BLK	I.BLK	PP070328.D	06 Mar 2025 21:58		YPIAJ	Ok,M
36	Q1436-09	K084-10C	PP070329.D	06 Mar 2025 22:14	AR1254 + AR1268 Hit,	YPIAJ	Ok,M

Instrument ID: ECD_P

Daily Analysis Runlog For Sequence/QC Batch ID # PP030625

Review By	yogesh	Review On	3/6/2025 10:54:36 AM
Supervise By	mohammad	Supervise On	3/8/2025 4:33:30 AM
SubDirectory	PP030625	HP Acquire Method	HP Processing Method PP022425
STD. NAME	STD REF.#		
Tune/Reschk	PP23735,PP23736,PP23737,PP23738,PP23739,PP23740,PP23741,PP23742,PP23743,PP23744,PP23745,PP23747,PP23748,PP23749,PP23750,P P23751,PP23752,PP23753,PP23754,PP23755,PP23756,PP23757,PP23758,PP23759,PP23760,PP23761,PP23762,PP23763,PP23764,PP23765,PP 23766,PP23767,PP23768,PP23769,PP23770,PP23771,PP23772,PP23773,PP23774,PP23775 PP23737,PP23742,PP23749,PP23754,PP23758,PP23763,PP23768,PP23773 PP23778,PP23780,PP23783,PP23784,PP23786,PP23788,PP23790,PP23947		
Initial Calibration Stds			
CCC			
Internal Standard/PEM			
ICV/I.BLK			
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

37	Q1437-01	K084-11B	PP070330.D	06 Mar 2025 22:30	AR1254 + AR1268 Hit, Need dilution 20X	YP\AJ	Dilution
38	Q1437-04	K084-12B	PP070331.D	06 Mar 2025 22:47	AR1254 + AR1268 Hit, Need dilution, 2X, 20X DCB High in both column	YP\AJ	Dilution
39	AR1660CCC500	AR1660CCC500	PP070332.D	06 Mar 2025 23:46		YP\AJ	Ok
40	AR1242CCC500	AR1242CCC500	PP070333.D	07 Mar 2025 00:18		YP\AJ	Ok,M
41	AR1248CCC500	AR1248CCC500	PP070334.D	07 Mar 2025 00:34		YP\AJ	Ok,M
42	AR1254CCC500	AR1254CCC500	PP070335.D	07 Mar 2025 00:51		YP\AJ	Ok,M
43	I.BLK	I.BLK	PP070336.D	07 Mar 2025 01:23		YP\AJ	Ok

M : Manual Integration

Instrument ID: ECD_P

Daily Analysis Runlog For Sequence/QC Batch ID # PP030725

Review By	yogesh	Review On	3/7/2025 1:29:40 PM
Supervise By	Ankita	Supervise On	3/10/2025 11:00:00 AM
SubDirectory	PP030725	HP Acquire Method	HP Processing Method PP022425

STD. NAME	STD REF.#
Tune/Reschk	
Initial Calibration Stds	PP23735,PP23736,PP23737,PP23738,PP23739,PP23740,PP23741,PP23742,PP23743,PP23744,PP23745,PP23747,PP23748,PP23749,PP23750,PP23751,PP23752,PP23753,PP23754,PP23755,PP23756,PP23757,PP23758,PP23759,PP23760,PP23761,PP23762,PP23763,PP23764,PP23765,PP23766,PP23767,PP23768,PP23769,PP23770,PP23771,PP23772,PP23773,PP23774,PP23775
CCC	PP23737,PP23742,PP23749,PP23754,PP23758,PP23763,PP23768,PP23773
Internal Standard/PEM	
ICV/I.BLK	PP23778,PP23780,PP23783,PP23784,PP23786,PP23788,PP23790,PP23947
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	HEXANE	HEXANE	PP070337.D	07 Mar 2025 08:41		YPIAJ	Ok
2	AR1660CCC500	AR1660CCC500	PP070338.D	07 Mar 2025 08:57		YPIAJ	Ok
3	I.BLK	I.BLK	PP070339.D	07 Mar 2025 09:15		YPIAJ	Ok
4	Q1434-10DL	K084-7CDL	PP070340.D	07 Mar 2025 09:32	AR1254 + AR1268 Hit	YPIAJ	Ok,M
5	Q1436-02DL	K084-8BDL	PP070341.D	07 Mar 2025 09:48	AR1254 + AR1268 Hit	YPIAJ	Ok,M
6	Q1436-08DL	K084-10BDL	PP070342.D	07 Mar 2025 10:04	AR1254 + AR1268 Hit, Need dilution	YPIAJ	Dilution
7	Q1436-08DL2	K084-10BDL2	PP070343.D	07 Mar 2025 10:20	AR1254 + AR1268 Hit	YPIAJ	Ok,M
8	AR1660CCC500	AR1660CCC500	PP070344.D	07 Mar 2025 10:58		YPIAJ	Ok
9	AR1242CCC500	AR1242CCC500	PP070345.D	07 Mar 2025 11:14		YPIAJ	Ok
10	AR1248CCC500	AR1248CCC500	PP070346.D	07 Mar 2025 11:31		YPIAJ	Ok
11	AR1254CCC500	AR1254CCC500	PP070347.D	07 Mar 2025 11:47		YPIAJ	Ok,M
12	I.BLK	I.BLK	PP070348.D	07 Mar 2025 12:03		YPIAJ	Ok
13	PB167005BL	PB167005BL	PP070349.D	07 Mar 2025 12:19		YPIAJ	Ok
14	Q1472-01	40308	PP070350.D	07 Mar 2025 12:36	TCMX High 2nd column	YPIAJ	Ok,M
15	Q1437-01DL	K084-11BDL	PP070351.D	07 Mar 2025 12:52	AR1254 + AR1268 Hit, DCB High in both column, TCMX high in 2nd column	YPIAJ	Ok,M
16	Q1437-04DL	K084-12BDL	PP070352.D	07 Mar 2025 13:08	AR1254 + AR1268 Hit, Need dilution 20x	YPIAJ	Dilution

Instrument ID: ECD_P

Daily Analysis Runlog For Sequence/QC Batch ID # PP030725

Review By	yogesh	Review On	3/7/2025 1:29:40 PM
Supervise By	Ankita	Supervise On	3/10/2025 11:00:00 AM
SubDirectory	PP030725	HP Acquire Method	HP Processing Method PP022425
STD. NAME	STD REF.#		
Tune/Reschk	PP23735,PP23736,PP23737,PP23738,PP23739,PP23740,PP23741,PP23742,PP23743,PP23744,PP23745,PP23747,PP23748,PP23749,PP23750,P P23751,PP23752,PP23753,PP23754,PP23755,PP23756,PP23757,PP23758,PP23759,PP23760,PP23761,PP23762,PP23763,PP23764,PP23765,PP 23766,PP23767,PP23768,PP23769,PP23770,PP23771,PP23772,PP23773,PP23774,PP23775 PP23737,PP23742,PP23749,PP23754,PP23758,PP23763,PP23768,PP23773 PP23778,PP23780,PP23783,PP23784,PP23786,PP23788,PP23790,PP23947		
Initial Calibration Stds			
CCC			
Internal Standard/PEM			
ICV/I.BLK			
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

17	Q1437-04DL2	K084-12BDL2	PP070353.D	07 Mar 2025 13:24	AR1254 + AR1268 Hit, All Surrogate High	YPIAJ	Ok,M
18	PB167029BL	PB167029BL	PP070354.D	07 Mar 2025 13:41		YPIAJ	Ok
19	PB167029BS	PB167029BS	PP070355.D	07 Mar 2025 13:57		YPIAJ	Ok
20	Q1437-02	K084-11C	PP070356.D	07 Mar 2025 14:13	AR1254 + AR1268 Hit,	YPIAJ	Ok,M
21	Q1437-05	K084-12C	PP070357.D	07 Mar 2025 14:29	AR1254 + AR1268 Hit	YPIAJ	Ok
22	Q1437-08	K084-13C	PP070358.D	07 Mar 2025 14:46	AR1254 + AR1268 Hit	YPIAJ	Ok,M
23	AR1660CCC500	AR1660CCC500	PP070359.D	07 Mar 2025 15:45		YPIAJ	Ok
24	AR1242CCC500	AR1242CCC500	PP070360.D	07 Mar 2025 16:01		YPIAJ	Ok,M
25	AR1248CCC500	AR1248CCC500	PP070361.D	07 Mar 2025 16:18		YPIAJ	Ok,M
26	AR1254CCC500	AR1254CCC500	PP070362.D	07 Mar 2025 16:34		YPIAJ	Ok,M
27	I.BLK	I.BLK	PP070363.D	07 Mar 2025 16:50		YPIAJ	Ok
28	Q1438-01	K084-14C	PP070364.D	07 Mar 2025 17:06	AR1254 + AR1268 Hit	YPIAJ	Ok,M
29	Q1438-09	K084-DUP4	PP070365.D	07 Mar 2025 17:23	AR1254 Hit	YPIAJ	Ok,M
30	Q1507-01	50-MIDDLESEX-AVE	PP070366.D	07 Mar 2025 17:39	AR1254 Hit	YPIAJ	Ok,M
31	Q1508-01	RBR251372	PP070367.D	07 Mar 2025 17:55	AR1254 Hit	YPIAJ	Ok,M
32	Q1508-01MS	RBR251372MS	PP070368.D	07 Mar 2025 18:11		YPIAJ	Ok,M
33	Q1508-01MSD	RBR251372MSD	PP070369.D	07 Mar 2025 18:27		YPIAJ	Ok,M
34	Q1514-01	ENV-105-SB01	PP070370.D	07 Mar 2025 18:44		YPIAJ	Ok
35	Q1514-03	ENV-105-SB02	PP070371.D	07 Mar 2025 19:00		YPIAJ	Ok,M

Instrument ID: ECD_P

Daily Analysis Runlog For Sequence/QC Batch ID # PP030725

Review By	yogesh	Review On	3/7/2025 1:29:40 PM
Supervise By	Ankita	Supervise On	3/10/2025 11:00:00 AM
SubDirectory	PP030725	HP Acquire Method	HP Processing Method PP022425
STD. NAME	STD REF.#		
Tune/Reschk			
Initial Calibration Stds	PP23735,PP23736,PP23737,PP23738,PP23739,PP23740,PP23741,PP23742,PP23743,PP23744,PP23745,PP23747,PP23748,PP23749,PP23750,P P23751,PP23752,PP23753,PP23754,PP23755,PP23756,PP23757,PP23758,PP23759,PP23760,PP23761,PP23762,PP23763,PP23764,PP23765,PP 23766,PP23767,PP23768,PP23769,PP23770,PP23771,PP23772,PP23773,PP23774,PP23775		
CCC	PP23737,PP23742,PP23749,PP23754,PP23758,PP23763,PP23768,PP23773		
Internal Standard/PEM			
ICV/I.BLK	PP23778,PP23780,PP23783,PP23784,PP23786,PP23788,PP23790,PP23947		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

36	Q1514-05	ENV-103-SB01	PP070372.D	07 Mar 2025 19:16	AR1254 + AR1260 Hit	YPIAJ	Ok,M
37	Q1515-01	AU-06-030625	PP070373.D	07 Mar 2025 19:32		YPIAJ	Ok,M
38	AR1660CCC500	AR1660CCC500	PP070374.D	07 Mar 2025 20:32		YPIAJ	Ok,M
39	AR1242CCC500	AR1242CCC500	PP070375.D	07 Mar 2025 20:48		YPIAJ	Ok,M
40	AR1248CCC500	AR1248CCC500	PP070376.D	07 Mar 2025 21:04		YPIAJ	Ok,M
41	AR1254CCC500	AR1254CCC500	PP070377.D	07 Mar 2025 21:21		YPIAJ	Ok,M
42	I.BLK	I.BLK	PP070378.D	07 Mar 2025 21:37		YPIAJ	Ok,M
43	Q1516-01	GAS-AUD-25-0004	PP070379.D	07 Mar 2025 21:53	AR1254 Hit	YPIAJ	Ok,M
44	Q1516-02	GAS-AUD-25-0025	PP070380.D	07 Mar 2025 22:09		YPIAJ	Ok,M
45	Q1518-01	OILY-DEBRIS-COMP	PP070381.D	07 Mar 2025 22:26		YPIAJ	Ok
46	Q1510-01	FMC-25-0001-0005	PP070382.D	07 Mar 2025 22:42	DCB High in 2nd column	YPIAJ	Ok,M
47	Q1510-02	FMC-25-0006	PP070383.D	07 Mar 2025 22:58	DCB High in 2nd column,	YPIAJ	Ok,M
48	AR1660CCC500	AR1660CCC500	PP070384.D	07 Mar 2025 23:58		YPIAJ	Ok
49	AR1242CCC500	AR1242CCC500	PP070385.D	08 Mar 2025 00:14		YPIAJ	Ok,M
50	AR1248CCC500	AR1248CCC500	PP070386.D	08 Mar 2025 00:30		YPIAJ	Ok,M
51	AR1254CCC500	AR1254CCC500	PP070387.D	08 Mar 2025 00:46		YPIAJ	Ok
52	I.BLK	I.BLK	PP070388.D	08 Mar 2025 01:03		YPIAJ	Ok

M : Manual Integration