

CALIBRATION VERIFICATION SUMMARY

Contract: ATCE02

Lab Code: CHEM Case No.: Q1437 SAS No.: Q1437 SDG NO.: Q1437

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 02/20/2025 02/20/2025

Client Sample No.: CCAL11 Date Analyzed: 03/07/2025

Lab Sample No.: AR1660CCC500 Data File : PO109708.D Time Analyzed: 12:30

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.790	4.691	4.891	535.420	500.000	7.1
Aroclor-1016-2	4.809	4.711	4.911	540.100	500.000	8.0
Aroclor-1016-3	4.865	4.767	4.967	535.310	500.000	7.1
Aroclor-1016-4	4.986	4.888	5.088	538.450	500.000	7.7
Aroclor-1016-5	5.244	5.145	5.345	542.310	500.000	8.5
Aroclor-1260-1	6.285	6.188	6.388	522.080	500.000	4.4
Aroclor-1260-2	6.475	6.377	6.577	517.570	500.000	3.5
Aroclor-1260-3	6.843	6.746	6.946	508.990	500.000	1.8
Aroclor-1260-4	7.104	7.005	7.205	502.090	500.000	0.4
Aroclor-1260-5	7.346	7.247	7.447	502.420	500.000	0.5
Decachlorobiphenyl	8.753	8.657	8.857	45.700	50.000	-8.6
Tetrachloro-m-xylene	3.696	3.598	3.798	52.770	50.000	5.5

CALIBRATION VERIFICATION SUMMARY

Contract: ATCE02

Lab Code: CHEM Case No.: Q1437 SAS No.: Q1437 SDG NO.: Q1437

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 02/20/2025 02/20/2025

Client Sample No.: CCAL11 Date Analyzed: 03/07/2025

Lab Sample No.: AR1660CCC500 Data File : PO109708.D Time Analyzed: 12:30

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.776	4.678	4.878	564.710	500.000	12.9
Aroclor-1016-2	4.796	4.697	4.897	569.610	500.000	13.9
Aroclor-1016-3	4.971	4.872	5.072	567.450	500.000	13.5
Aroclor-1016-4	5.013	4.914	5.114	533.490	500.000	6.7
Aroclor-1016-5	5.227	5.127	5.327	564.430	500.000	12.9
Aroclor-1260-1	6.260	6.161	6.361	539.900	500.000	8.0
Aroclor-1260-2	6.448	6.348	6.548	544.050	500.000	8.8
Aroclor-1260-3	6.601	6.502	6.702	533.960	500.000	6.8
Aroclor-1260-4	7.073	6.974	7.174	526.130	500.000	5.2
Aroclor-1260-5	7.313	7.214	7.414	539.230	500.000	7.8
Decachlorobiphenyl	8.706	8.607	8.807	48.170	50.000	-3.7
Tetrachloro-m-xylene	3.694	3.594	3.794	55.950	50.000	11.9