

CALIBRATION VERIFICATION SUMMARY

Contract: ATCE02

Lab Code: CHEM Case No.: Q1437 SAS No.: Q1437 SDG NO.: Q1437

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 02/24/2025 02/24/2025

Client Sample No.: CCAL14 Date Analyzed: 03/07/2025

Lab Sample No.: AR1660CCC500 Data File : PP070344.D Time Analyzed: 10:58

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.677	5.581	5.781	480.930	500.000	-3.8
Aroclor-1016-2	5.699	5.602	5.802	500.630	500.000	0.1
Aroclor-1016-3	5.762	5.665	5.865	483.140	500.000	-3.4
Aroclor-1016-4	5.858	5.762	5.962	491.090	500.000	-1.8
Aroclor-1016-5	6.152	6.056	6.256	494.120	500.000	-1.2
Aroclor-1260-1	7.272	7.175	7.375	497.790	500.000	-0.4
Aroclor-1260-2	7.525	7.429	7.629	482.220	500.000	-3.6
Aroclor-1260-3	7.884	7.788	7.988	522.690	500.000	4.5
Aroclor-1260-4	8.108	8.012	8.212	516.070	500.000	3.2
Aroclor-1260-5	8.429	8.333	8.533	506.350	500.000	1.3
Decachlorobiphenyl	10.251	10.156	10.356	49.190	50.000	-1.6
Tetrachloro-m-xylene	4.525	4.427	4.627	53.970	50.000	7.9

CALIBRATION VERIFICATION SUMMARY

Contract: ATCE02

Lab Code: CHEM Case No.: Q1437 SAS No.: Q1437 SDG NO.: Q1437

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 02/24/2025 02/24/2025

Client Sample No.: CCAL14 Date Analyzed: 03/07/2025

Lab Sample No.: AR1660CCC500 Data File : PP070344.D Time Analyzed: 10:58

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	4.916	4.820	5.020	505.660	500.000	1.1
Aroclor-1016-2	4.935	4.839	5.039	503.030	500.000	0.6
Aroclor-1016-3	5.112	5.017	5.217	522.440	500.000	4.5
Aroclor-1016-4	5.154	5.058	5.258	515.150	500.000	3.0
Aroclor-1016-5	5.369	5.273	5.473	528.830	500.000	5.8
Aroclor-1260-1	6.406	6.312	6.512	487.230	500.000	-2.6
Aroclor-1260-2	6.594	6.500	6.700	492.820	500.000	-1.4
Aroclor-1260-3	6.748	6.654	6.854	461.080	500.000	-7.8
Aroclor-1260-4	7.221	7.126	7.326	526.310	500.000	5.3
Aroclor-1260-5	7.462	7.367	7.567	540.800	500.000	8.2
Decachlorobiphenyl	8.881	8.788	8.988	46.780	50.000	-6.4
Tetrachloro-m-xylene	3.828	3.730	3.930	50.560	50.000	1.1