

## Report of Analysis

Client:	ATC Group Services LLC		Date Collected:	02/25/25	
Project:	K084-SCA PCBs NYC - 2022SCA421		Date Received:	02/26/25	
Client Sample ID:	K084-4CMS		SDG No.:	Q1437	
Lab Sample ID:	Q1434-02MS		Matrix:	SOIL	
Analytical Method:	SW8082A		% Solid:	72.3	Decanted:
Sample Wt/Vol:	30.04	Units: g	Final Vol:	10000	uL
Soil Aliquot Vol:		uL	Test:	PCB Group1	
Extraction Type:			Injection Volume :		
GPC Factor :	1.0	PH :			
Prep Method :	SW3541B				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PO109688.D	1	03/06/25 11:40	03/07/25 02:18	PB167022

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
<b>TARGETS</b>						
12674-11-2	Aroclor-1016	257		4.70	23.5	ug/kg
11104-28-2	Aroclor-1221	8.90	U	8.90	23.5	ug/kg
11141-16-5	Aroclor-1232	4.70	U	4.70	23.5	ug/kg
53469-21-9	Aroclor-1242	4.70	U	4.70	23.5	ug/kg
12672-29-6	Aroclor-1248	10.9	U	10.9	23.5	ug/kg
11097-69-1	Aroclor-1254	1700	E	3.80	23.5	ug/kg
37324-23-5	Aroclor-1262	6.30	U	6.30	23.5	ug/kg
11100-14-4	Aroclor-1268	131		4.70	23.5	ug/kg
11096-82-5	Aroclor-1260	711	EP	4.00	23.5	ug/kg
Total PCBs	Total PCBs	2800		17.2	23.5	ug/kg
<b>SURROGATES</b>						
877-09-8	Tetrachloro-m-xylene	21.8		32 - 144	109%	SPK: 20
2051-24-3	Decachlorobiphenyl	26.0		32 - 175	130%	SPK: 20

### Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

\* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit