



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q1437

Client: ATC Group Services LLC

Analytical Method: 8082A Datafile : PO109547.D

Lab Sample ID	Parameter	Spike	Result	Units	Rec	RPD	Qual	RPD		Limits	
								Qual	Low	High	RPD
PB166889BS	AR1016	166.5	142	ug/kg	85				71	120	
	AR1260	166.5	130	ug/kg	78				65	130	
	Total PCBs	0	272	ug/kg					0	0	