

CALIBRATION VERIFICATION SUMMARY

Contract: ATCE02

Lab Code: CHEM Case No.: Q1437 SAS No.: Q1437 SDG NO.: Q1437

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 02/20/2025 02/20/2025

Client Sample No.: CCAL04 Date Analyzed: 02/28/2025

Lab Sample No.: AR1660CC500 Data File : PO109577.D Time Analyzed: 10:00

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.790	4.691	4.891	538.300	500.000	7.7
Aroclor-1016-2	4.810	4.711	4.911	544.480	500.000	8.9
Aroclor-1016-3	4.866	4.767	4.967	537.090	500.000	7.4
Aroclor-1016-4	4.986	4.888	5.088	543.410	500.000	8.7
Aroclor-1016-5	5.244	5.145	5.345	546.860	500.000	9.4
Aroclor-1260-1	6.286	6.188	6.388	506.170	500.000	1.2
Aroclor-1260-2	6.475	6.377	6.577	493.750	500.000	-1.3
Aroclor-1260-3	6.843	6.746	6.946	478.310	500.000	-4.3
Aroclor-1260-4	7.104	7.005	7.205	473.550	500.000	-5.3
Aroclor-1260-5	7.346	7.247	7.447	484.160	500.000	-3.2
Decachlorobiphenyl	8.753	8.657	8.857	43.040	50.000	-13.9
Tetrachloro-m-xylene	3.696	3.598	3.798	52.980	50.000	6.0

CALIBRATION VERIFICATION SUMMARY

Contract: ATCE02

Lab Code: CHEM Case No.: Q1437 SAS No.: Q1437 SDG NO.: Q1437

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 02/20/2025 02/20/2025

Client Sample No.: CCAL04 Date Analyzed: 02/28/2025

Lab Sample No.: AR1660CC500 Data File : PO109577.D Time Analyzed: 10:00

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.778	4.678	4.878	583.040	500.000	16.6
Aroclor-1016-2	4.797	4.697	4.897	587.910	500.000	17.6
Aroclor-1016-3	4.972	4.872	5.072	583.900	500.000	16.8
Aroclor-1016-4	5.015	4.914	5.114	538.110	500.000	7.6
Aroclor-1016-5	5.227	5.127	5.327	586.400	500.000	17.3
Aroclor-1260-1	6.261	6.161	6.361	552.380	500.000	10.5
Aroclor-1260-2	6.448	6.348	6.548	556.350	500.000	11.3
Aroclor-1260-3	6.601	6.502	6.702	537.920	500.000	7.6
Aroclor-1260-4	7.074	6.974	7.174	527.160	500.000	5.4
Aroclor-1260-5	7.314	7.214	7.414	543.250	500.000	8.7
Decachlorobiphenyl	8.705	8.607	8.807	47.720	50.000	-4.6
Tetrachloro-m-xylene	3.694	3.594	3.794	57.460	50.000	14.9