

CALIBRATION VERIFICATION SUMMARY

Contract: ATCE02

Lab Code: CHEM Case No.: Q1437 SAS No.: Q1437 SDG NO.: Q1437

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 02/20/2025 02/20/2025

Client Sample No.: CCAL07 Date Analyzed: 03/06/2025

Lab Sample No.: AR1660CCC500 Data File : PO109662.D Time Analyzed: 16:11

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	4.791	4.691	4.891	484.120	500.000	-3.2
Aroclor-1016-2	4.810	4.711	4.911	486.830	500.000	-2.6
Aroclor-1016-3	4.867	4.767	4.967	484.640	500.000	-3.1
Aroclor-1016-4	4.987	4.888	5.088	484.750	500.000	-3.1
Aroclor-1016-5	5.245	5.145	5.345	467.760	500.000	-6.4
Aroclor-1260-1	6.287	6.188	6.388	480.610	500.000	-3.9
Aroclor-1260-2	6.476	6.377	6.577	468.300	500.000	-6.3
Aroclor-1260-3	6.845	6.746	6.946	444.240	500.000	-11.2
Aroclor-1260-4	7.105	7.005	7.205	434.880	500.000	-13.0
Aroclor-1260-5	7.347	7.247	7.447	430.260	500.000	-13.9
Decachlorobiphenyl	8.756	8.657	8.857	39.220	50.000	-21.6
Tetrachloro-m-xylene	3.696	3.598	3.798	47.820	50.000	-4.4

CALIBRATION VERIFICATION SUMMARY

Contract: ATCE02

Lab Code: CHEM Case No.: Q1437 SAS No.: Q1437 SDG NO.: Q1437

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 02/20/2025 02/20/2025

Client Sample No.: CCAL07 Date Analyzed: 03/06/2025

Lab Sample No.: AR1660CCC500 Data File : PO109662.D Time Analyzed: 16:11

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.778	4.678	4.878	536.050	500.000	7.2
Aroclor-1016-2	4.797	4.697	4.897	543.930	500.000	8.8
Aroclor-1016-3	4.973	4.872	5.072	540.440	500.000	8.1
Aroclor-1016-4	5.015	4.914	5.114	451.810	500.000	-9.6
Aroclor-1016-5	5.228	5.127	5.327	521.700	500.000	4.3
Aroclor-1260-1	6.261	6.161	6.361	512.490	500.000	2.5
Aroclor-1260-2	6.449	6.348	6.548	510.220	500.000	2.0
Aroclor-1260-3	6.603	6.502	6.702	498.850	500.000	-0.2
Aroclor-1260-4	7.074	6.974	7.174	483.020	500.000	-3.4
Aroclor-1260-5	7.315	7.214	7.414	492.120	500.000	-1.6
Decachlorobiphenyl	8.708	8.607	8.807	43.320	50.000	-13.4
Tetrachloro-m-xylene	3.695	3.594	3.794	53.270	50.000	6.5