

CALIBRATION VERIFICATION SUMMARY

Contract: ATCE02

Lab Code: CHEM Case No.: Q1437 SAS No.: Q1437 SDG NO.: Q1437

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 02/20/2025 02/20/2025

Client Sample No.: CCAL08 Date Analyzed: 03/06/2025

Lab Sample No.: AR1660CCC500 Data File : PO109677.D Time Analyzed: 22:19

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	4.790	4.691	4.891	516.410	500.000	3.3
Aroclor-1016-2	4.810	4.711	4.911	521.040	500.000	4.2
Aroclor-1016-3	4.866	4.767	4.967	516.800	500.000	3.4
Aroclor-1016-4	4.987	4.888	5.088	517.710	500.000	3.5
Aroclor-1016-5	5.244	5.145	5.345	489.290	500.000	-2.1
Aroclor-1260-1	6.286	6.188	6.388	524.710	500.000	4.9
Aroclor-1260-2	6.475	6.377	6.577	507.570	500.000	1.5
Aroclor-1260-3	6.843	6.746	6.946	485.990	500.000	-2.8
Aroclor-1260-4	7.104	7.005	7.205	479.350	500.000	-4.1
Aroclor-1260-5	7.345	7.247	7.447	481.100	500.000	-3.8
Decachlorobiphenyl	8.753	8.657	8.857	43.340	50.000	-13.3
Tetrachloro-m-xylene	3.696	3.598	3.798	50.660	50.000	1.3

CALIBRATION VERIFICATION SUMMARY

Contract: ATCE02

Lab Code: CHEM Case No.: Q1437 SAS No.: Q1437 SDG NO.: Q1437

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 02/20/2025 02/20/2025

Client Sample No.: CCAL08 Date Analyzed: 03/06/2025

Lab Sample No.: AR1660CCC500 Data File : PO109677.D Time Analyzed: 22:19

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	4.777	4.678	4.878	553.190	500.000	10.6
Aroclor-1016-2	4.796	4.697	4.897	570.410	500.000	14.1
Aroclor-1016-3	4.972	4.872	5.072	563.990	500.000	12.8
Aroclor-1016-4	5.014	4.914	5.114	456.810	500.000	-8.6
Aroclor-1016-5	5.227	5.127	5.327	534.200	500.000	6.8
Aroclor-1260-1	6.260	6.161	6.361	551.440	500.000	10.3
Aroclor-1260-2	6.448	6.348	6.548	550.920	500.000	10.2
Aroclor-1260-3	6.601	6.502	6.702	533.730	500.000	6.7
Aroclor-1260-4	7.073	6.974	7.174	518.870	500.000	3.8
Aroclor-1260-5	7.314	7.214	7.414	533.180	500.000	6.6
Decachlorobiphenyl	8.707	8.607	8.807	46.160	50.000	-7.7
Tetrachloro-m-xylene	3.694	3.594	3.794	55.410	50.000	10.8