

CALIBRATION VERIFICATION SUMMARY

Contract: ATCE02

Lab Code: CHEM Case No.: Q1437 SAS No.: Q1437 SDG NO.: Q1437

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 02/24/2025 02/24/2025

Client Sample No.: CCAL13 Date Analyzed: 03/06/2025

Lab Sample No.: AR1660CCC500 Data File : PP070332.D Time Analyzed: 23:46

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.679	5.581	5.781	439.060	500.000	-12.2
Aroclor-1016-2	5.701	5.602	5.802	468.000	500.000	-6.4
Aroclor-1016-3	5.763	5.665	5.865	497.050	500.000	-0.6
Aroclor-1016-4	5.860	5.762	5.962	435.100	500.000	-13.0
Aroclor-1016-5	6.154	6.056	6.256	462.800	500.000	-7.4
Aroclor-1260-1	7.274	7.175	7.375	483.900	500.000	-3.2
Aroclor-1260-2	7.528	7.429	7.629	460.220	500.000	-8.0
Aroclor-1260-3	7.887	7.788	7.988	497.090	500.000	-0.6
Aroclor-1260-4	8.111	8.012	8.212	481.180	500.000	-3.8
Aroclor-1260-5	8.431	8.333	8.533	473.930	500.000	-5.2
Decachlorobiphenyl	10.256	10.156	10.356	45.010	50.000	-10.0
Tetrachloro-m-xylene	4.526	4.427	4.627	50.280	50.000	0.6

CALIBRATION VERIFICATION SUMMARY

Contract: ATCE02

Lab Code: CHEM Case No.: Q1437 SAS No.: Q1437 SDG NO.: Q1437

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 02/24/2025 02/24/2025

Client Sample No.: CCAL13 Date Analyzed: 03/06/2025

Lab Sample No.: AR1660CCC500 Data File : PP070332.D Time Analyzed: 23:46

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.917	4.820	5.020	463.950	500.000	-7.2
Aroclor-1016-2	4.936	4.839	5.039	470.730	500.000	-5.9
Aroclor-1016-3	5.114	5.017	5.217	461.920	500.000	-7.6
Aroclor-1016-4	5.156	5.058	5.258	425.260	500.000	-14.9
Aroclor-1016-5	5.371	5.273	5.473	523.410	500.000	4.7
Aroclor-1260-1	6.408	6.312	6.512	468.790	500.000	-6.2
Aroclor-1260-2	6.596	6.500	6.700	473.290	500.000	-5.3
Aroclor-1260-3	6.750	6.654	6.854	417.150	500.000	-16.6
Aroclor-1260-4	7.223	7.126	7.326	493.650	500.000	-1.3
Aroclor-1260-5	7.463	7.367	7.567	522.620	500.000	4.5
Decachlorobiphenyl	8.884	8.788	8.988	44.030	50.000	-11.9
Tetrachloro-m-xylene	3.828	3.730	3.930	48.220	50.000	-3.6