



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q1437

Client: ATC Group Services LLC

Analytical Method: 8082A Datafile : PO109683.D

Lab Sample ID	Parameter	Spike	Result	Units	Rec	RPD	Qual	RPD		Limits	
								Qual	Low	High	RPD
PB167022BS	AR1016	166.6	168	ug/kg	101				71	120	
	AR1260	166.6	164	ug/kg	98				65	130	