

CALIBRATION VERIFICATION SUMMARY

Contract: ATCE02

Lab Code: CHEM Case No.: Q1438 SAS No.: Q1438 SDG NO.: Q1438

Continuing Calib Date: 02/28/2025 Initial Calibration Date(s): 02/20/2025 02/21/2025

Continuing Calib Time: 00:38 Initial Calibration Time(s): 16:46 01:02

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1248-1	(1)	4.79	4.79	4.69	4.89	0.00
Aroclor-1248-2	(2)	5.03	5.03	4.93	5.13	0.00
Aroclor-1248-3	(3)	5.24	5.25	5.15	5.35	0.01
Aroclor-1248-4	(4)	5.60	5.60	5.50	5.70	0.00
Aroclor-1248-5	(5)	5.64	5.64	5.54	5.74	0.00
Tetrachloro-m-xylene		3.70	3.70	3.60	3.80	0.01
Decachlorobiphenyl		8.75	8.76	8.66	8.86	0.01

CALIBRATION VERIFICATION SUMMARY

Contract: ATCE02

Lab Code: CHEM Case No.: Q1438 SAS No.: Q1438 SDG NO.: Q1438

Continuing Calib Date: 02/28/2025 Initial Calibration Date(s): 02/20/2025 02/21/2025

Continuing Calib Time: 00:38 Initial Calibration Time(s): 16:46 01:02

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1248-1	(1)	4.78	4.78	4.68	4.88	0.00
Aroclor-1248-2	(2)	5.01	5.01	4.91	5.11	0.00
Aroclor-1248-3	(3)	5.06	5.06	4.96	5.16	0.00
Aroclor-1248-4	(4)	5.23	5.23	5.13	5.33	0.00
Aroclor-1248-5	(5)	5.62	5.62	5.52	5.72	0.00
Tetrachloro-m-xylene		3.70	3.69	3.59	3.79	0.00
Decachlorobiphenyl		8.71	8.71	8.61	8.81	0.00