

### CALIBRATION VERIFICATION SUMMARY

Contract: ATCE02

Lab Code: CHEM Case No.: Q1438 SAS No.: Q1438 SDG NO.: Q1438

Continuing Calib Date: 03/07/2025 Initial Calibration Date(s): 02/20/2025 02/21/2025

Continuing Calib Time: 05:42 Initial Calibration Time(s): 16:46 01:02

GC Column: ZB-MR1 ID: 0.32 (mm)

| COMPOUND             |     | CCAL<br>RT | AVG<br>RT | RT WINDOW |      | DIFF<br>RT |
|----------------------|-----|------------|-----------|-----------|------|------------|
|                      |     |            |           | FROM      | TO   |            |
| Aroclor-1254-1       | (1) | 5.60       | 5.60      | 5.50      | 5.70 | 0.00       |
| Aroclor-1254-2       | (2) | 5.75       | 5.75      | 5.65      | 5.85 | 0.00       |
| Aroclor-1254-3       | (3) | 6.16       | 6.16      | 6.06      | 6.26 | 0.00       |
| Aroclor-1254-4       | (4) | 6.38       | 6.39      | 6.29      | 6.49 | 0.01       |
| Aroclor-1254-5       | (5) | 6.80       | 6.81      | 6.71      | 6.91 | 0.01       |
| Tetrachloro-m-xylene |     | 3.70       | 3.70      | 3.60      | 3.80 | 0.00       |
| Decachlorobiphenyl   |     | 8.76       | 8.76      | 8.66      | 8.86 | 0.00       |

### CALIBRATION VERIFICATION SUMMARY

Contract: ATCE02

Lab Code: CHEM Case No.: Q1438 SAS No.: Q1438 SDG NO.: Q1438

Continuing Calib Date: 03/07/2025 Initial Calibration Date(s): 02/20/2025 02/21/2025

Continuing Calib Time: 05:42 Initial Calibration Time(s): 16:46 01:02

GC Column: ZB-MR2 ID: 0.32 (mm)

| COMPOUND             | CCAL<br>RT | AVG<br>RT | RT WINDOW |      | DIFF<br>RT |
|----------------------|------------|-----------|-----------|------|------------|
|                      |            |           | FROM      | TO   |            |
| Aroclor-1254-1 (1)   | 5.58       | 5.58      | 5.48      | 5.68 | 0.00       |
| Aroclor-1254-2 (2)   | 5.73       | 5.73      | 5.63      | 5.83 | 0.00       |
| Aroclor-1254-3 (3)   | 6.13       | 6.13      | 6.03      | 6.23 | 0.00       |
| Aroclor-1254-4 (4)   | 6.36       | 6.36      | 6.26      | 6.46 | 0.00       |
| Aroclor-1254-5 (5)   | 6.78       | 6.77      | 6.67      | 6.87 | -0.01      |
| Tetrachloro-m-xylene | 3.69       | 3.69      | 3.59      | 3.79 | 0.00       |
| Decachlorobiphenyl   | 8.71       | 8.71      | 8.61      | 8.81 | 0.00       |