

CALIBRATION VERIFICATION SUMMARY

Contract: ATCE02

Lab Code: CHEM Case No.: Q1438 SAS No.: Q1438 SDG NO.: Q1438

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 02/20/2025 02/20/2025

Client Sample No.: CCAL01 Date Analyzed: 02/27/2025

Lab Sample No.: AR1660CCC500 Data File : PO109518.D Time Analyzed: 10:32

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	4.791	4.691	4.891	493.830	500.000	-1.2
Aroclor-1016-2	4.810	4.711	4.911	495.020	500.000	-1.0
Aroclor-1016-3	4.866	4.767	4.967	496.720	500.000	-0.7
Aroclor-1016-4	4.987	4.888	5.088	495.780	500.000	-0.8
Aroclor-1016-5	5.245	5.145	5.345	482.940	500.000	-3.4
Aroclor-1260-1	6.287	6.188	6.388	479.690	500.000	-4.1
Aroclor-1260-2	6.475	6.377	6.577	482.740	500.000	-3.5
Aroclor-1260-3	6.844	6.746	6.946	484.020	500.000	-3.2
Aroclor-1260-4	7.104	7.005	7.205	487.270	500.000	-2.5
Aroclor-1260-5	7.345	7.247	7.447	493.150	500.000	-1.4
Decachlorobiphenyl	8.755	8.657	8.857	44.730	50.000	-10.5
Tetrachloro-m-xylene	3.696	3.598	3.798	48.740	50.000	-2.5

CALIBRATION VERIFICATION SUMMARY

Contract: ATCE02

Lab Code: CHEM Case No.: Q1438 SAS No.: Q1438 SDG NO.: Q1438

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 02/20/2025 02/20/2025

Client Sample No.: CCAL01 Date Analyzed: 02/27/2025

Lab Sample No.: AR1660CCC500 Data File : PO109518.D Time Analyzed: 10:32

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	4.776	4.678	4.878	527.740	500.000	5.5
Aroclor-1016-2	4.796	4.697	4.897	528.720	500.000	5.7
Aroclor-1016-3	4.971	4.872	5.072	531.390	500.000	6.3
Aroclor-1016-4	5.013	4.914	5.114	518.370	500.000	3.7
Aroclor-1016-5	5.226	5.127	5.327	512.200	500.000	2.4
Aroclor-1260-1	6.260	6.161	6.361	505.600	500.000	1.1
Aroclor-1260-2	6.447	6.348	6.548	515.360	500.000	3.1
Aroclor-1260-3	6.601	6.502	6.702	512.610	500.000	2.5
Aroclor-1260-4	7.072	6.974	7.174	513.470	500.000	2.7
Aroclor-1260-5	7.313	7.214	7.414	528.450	500.000	5.7
Decachlorobiphenyl	8.705	8.607	8.807	48.990	50.000	-2.0
Tetrachloro-m-xylene	3.693	3.594	3.794	52.130	50.000	4.3