

CALIBRATION VERIFICATION SUMMARY

Contract: ATCE02

Lab Code: CHEM Case No.: Q1438 SAS No.: Q1438 SDG NO.: Q1438

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 02/20/2025 02/20/2025

Client Sample No.: CCAL02 Date Analyzed: 02/27/2025

Lab Sample No.: AR1660CCC500 Data File : PO109533.D Time Analyzed: 17:49

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	4.791	4.691	4.891	506.740	500.000	1.3
Aroclor-1016-2	4.811	4.711	4.911	506.890	500.000	1.4
Aroclor-1016-3	4.867	4.767	4.967	504.160	500.000	0.8
Aroclor-1016-4	4.988	4.888	5.088	502.480	500.000	0.5
Aroclor-1016-5	5.245	5.145	5.345	480.720	500.000	-3.9
Aroclor-1260-1	6.287	6.188	6.388	476.100	500.000	-4.8
Aroclor-1260-2	6.476	6.377	6.577	466.060	500.000	-6.8
Aroclor-1260-3	6.845	6.746	6.946	439.850	500.000	-12.0
Aroclor-1260-4	7.105	7.005	7.205	434.850	500.000	-13.0
Aroclor-1260-5	7.347	7.247	7.447	428.000	500.000	-14.4
Decachlorobiphenyl	8.756	8.657	8.857	39.490	50.000	-21.0
Tetrachloro-m-xylene	3.697	3.598	3.798	49.320	50.000	-1.4

CALIBRATION VERIFICATION SUMMARY

Contract: ATCE02

Lab Code: CHEM Case No.: Q1438 SAS No.: Q1438 SDG NO.: Q1438

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 02/20/2025 02/20/2025

Client Sample No.: CCAL02 Date Analyzed: 02/27/2025

Lab Sample No.: AR1660CCC500 Data File : PO109533.D Time Analyzed: 17:49

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.779	4.678	4.878	553.210	500.000	10.6
Aroclor-1016-2	4.799	4.697	4.897	559.460	500.000	11.9
Aroclor-1016-3	4.974	4.872	5.072	559.210	500.000	11.8
Aroclor-1016-4	5.016	4.914	5.114	537.920	500.000	7.6
Aroclor-1016-5	5.229	5.127	5.327	534.530	500.000	6.9
Aroclor-1260-1	6.263	6.161	6.361	500.020	500.000	0.0
Aroclor-1260-2	6.450	6.348	6.548	504.790	500.000	1.0
Aroclor-1260-3	6.603	6.502	6.702	495.560	500.000	-0.9
Aroclor-1260-4	7.075	6.974	7.174	480.720	500.000	-3.9
Aroclor-1260-5	7.316	7.214	7.414	501.850	500.000	0.4
Decachlorobiphenyl	8.708	8.607	8.807	46.250	50.000	-7.5
Tetrachloro-m-xylene	3.696	3.594	3.794	54.640	50.000	9.3