

CALIBRATION VERIFICATION SUMMARY

Contract: ATCE02

Lab Code: CHEM Case No.: Q1438 SAS No.: Q1438 SDG NO.: Q1438

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 02/20/2025 02/20/2025

Client Sample No.: CCAL03 Date Analyzed: 02/28/2025

Lab Sample No.: AR1660CCC500 Data File : PO109548.D Time Analyzed: 00:02

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.792	4.691	4.891	495.600	500.000	-0.9
Aroclor-1016-2	4.811	4.711	4.911	499.510	500.000	-0.1
Aroclor-1016-3	4.867	4.767	4.967	491.200	500.000	-1.8
Aroclor-1016-4	4.988	4.888	5.088	487.140	500.000	-2.6
Aroclor-1016-5	5.245	5.145	5.345	478.990	500.000	-4.2
Aroclor-1260-1	6.288	6.188	6.388	473.590	500.000	-5.3
Aroclor-1260-2	6.476	6.377	6.577	444.830	500.000	-11.0
Aroclor-1260-3	6.845	6.746	6.946	422.900	500.000	-15.4
Aroclor-1260-4	7.105	7.005	7.205	421.070	500.000	-15.8
Aroclor-1260-5	7.347	7.247	7.447	428.290	500.000	-14.3
Decachlorobiphenyl	8.755	8.657	8.857	39.430	50.000	-21.1
Tetrachloro-m-xylene	3.698	3.598	3.798	49.160	50.000	-1.7

CALIBRATION VERIFICATION SUMMARY

Contract: ATCE02

Lab Code: CHEM Case No.: Q1438 SAS No.: Q1438 SDG NO.: Q1438

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 02/20/2025 02/20/2025

Client Sample No.: CCAL03 Date Analyzed: 02/28/2025

Lab Sample No.: AR1660CCC500 Data File : PO109548.D Time Analyzed: 00:02

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	4.779	4.678	4.878	550.950	500.000	10.2
Aroclor-1016-2	4.798	4.697	4.897	554.860	500.000	11.0
Aroclor-1016-3	4.974	4.872	5.072	550.190	500.000	10.0
Aroclor-1016-4	5.015	4.914	5.114	513.690	500.000	2.7
Aroclor-1016-5	5.229	5.127	5.327	539.180	500.000	7.8
Aroclor-1260-1	6.262	6.161	6.361	504.870	500.000	1.0
Aroclor-1260-2	6.449	6.348	6.548	505.070	500.000	1.0
Aroclor-1260-3	6.603	6.502	6.702	493.490	500.000	-1.3
Aroclor-1260-4	7.075	6.974	7.174	488.450	500.000	-2.3
Aroclor-1260-5	7.315	7.214	7.414	496.860	500.000	-0.6
Decachlorobiphenyl	8.708	8.607	8.807	45.170	50.000	-9.7
Tetrachloro-m-xylene	3.695	3.594	3.794	54.270	50.000	8.5