

CALIBRATION VERIFICATION SUMMARY

Contract: ATCE02

Lab Code: CHEM Case No.: Q1438 SAS No.: Q1438 SDG NO.: Q1438

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 02/20/2025 02/20/2025

Client Sample No.: CCAL04 Date Analyzed: 02/28/2025

Lab Sample No.: AR1660CCC500 Data File : PO109579.D Time Analyzed: 10:57

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.792	4.691	4.891	505.520	500.000	1.1
Aroclor-1016-2	4.811	4.711	4.911	510.010	500.000	2.0
Aroclor-1016-3	4.867	4.767	4.967	505.900	500.000	1.2
Aroclor-1016-4	4.989	4.888	5.088	509.050	500.000	1.8
Aroclor-1016-5	5.246	5.145	5.345	489.670	500.000	-2.1
Aroclor-1260-1	6.288	6.188	6.388	476.600	500.000	-4.7
Aroclor-1260-2	6.476	6.377	6.577	468.270	500.000	-6.3
Aroclor-1260-3	6.846	6.746	6.946	451.200	500.000	-9.8
Aroclor-1260-4	7.105	7.005	7.205	450.390	500.000	-9.9
Aroclor-1260-5	7.347	7.247	7.447	455.830	500.000	-8.8
Decachlorobiphenyl	8.757	8.657	8.857	41.780	50.000	-16.4
Tetrachloro-m-xylene	3.698	3.598	3.798	49.920	50.000	-0.2

CALIBRATION VERIFICATION SUMMARY

Contract: ATCE02

Lab Code: CHEM Case No.: Q1438 SAS No.: Q1438 SDG NO.: Q1438

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 02/20/2025 02/20/2025

Client Sample No.: CCAL04 Date Analyzed: 02/28/2025

Lab Sample No.: AR1660CCC500 Data File : PO109579.D Time Analyzed: 10:57

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	4.778	4.678	4.878	543.570	500.000	8.7
Aroclor-1016-2	4.797	4.697	4.897	551.220	500.000	10.2
Aroclor-1016-3	4.972	4.872	5.072	548.270	500.000	9.7
Aroclor-1016-4	5.014	4.914	5.114	526.960	500.000	5.4
Aroclor-1016-5	5.228	5.127	5.327	522.400	500.000	4.5
Aroclor-1260-1	6.261	6.161	6.361	513.020	500.000	2.6
Aroclor-1260-2	6.448	6.348	6.548	519.260	500.000	3.9
Aroclor-1260-3	6.601	6.502	6.702	513.170	500.000	2.6
Aroclor-1260-4	7.074	6.974	7.174	505.050	500.000	1.0
Aroclor-1260-5	7.314	7.214	7.414	519.320	500.000	3.9
Decachlorobiphenyl	8.707	8.607	8.807	46.110	50.000	-7.8
Tetrachloro-m-xylene	3.694	3.594	3.794	53.530	50.000	7.1