

CALIBRATION VERIFICATION SUMMARY

Contract: ATCE02

Lab Code: CHEM Case No.: Q1438 SAS No.: Q1438 SDG NO.: Q1438

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 02/20/2025 02/20/2025

Client Sample No.: CCAL08 Date Analyzed: 03/07/2025

Lab Sample No.: AR1660CCC500 Data File : PO109692.D Time Analyzed: 04:29

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	4.791	4.691	4.891	515.140	500.000	3.0
Aroclor-1016-2	4.810	4.711	4.911	517.080	500.000	3.4
Aroclor-1016-3	4.866	4.767	4.967	512.890	500.000	2.6
Aroclor-1016-4	4.987	4.888	5.088	513.090	500.000	2.6
Aroclor-1016-5	5.244	5.145	5.345	480.970	500.000	-3.8
Aroclor-1260-1	6.286	6.188	6.388	511.810	500.000	2.4
Aroclor-1260-2	6.475	6.377	6.577	495.100	500.000	-1.0
Aroclor-1260-3	6.844	6.746	6.946	470.290	500.000	-5.9
Aroclor-1260-4	7.104	7.005	7.205	458.790	500.000	-8.2
Aroclor-1260-5	7.346	7.247	7.447	456.010	500.000	-8.8
Decachlorobiphenyl	8.754	8.657	8.857	41.410	50.000	-17.2
Tetrachloro-m-xylene	3.697	3.598	3.798	51.040	50.000	2.1

CALIBRATION VERIFICATION SUMMARY

Contract: ATCE02

Lab Code: CHEM Case No.: Q1438 SAS No.: Q1438 SDG NO.: Q1438

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 02/20/2025 02/20/2025

Client Sample No.: CCAL08 Date Analyzed: 03/07/2025

Lab Sample No.: AR1660CCC500 Data File : PO109692.D Time Analyzed: 04:29

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	4.778	4.678	4.878	560.290	500.000	12.1
Aroclor-1016-2	4.797	4.697	4.897	563.430	500.000	12.7
Aroclor-1016-3	4.973	4.872	5.072	559.930	500.000	12.0
Aroclor-1016-4	5.014	4.914	5.114	446.180	500.000	-10.8
Aroclor-1016-5	5.227	5.127	5.327	522.720	500.000	4.5
Aroclor-1260-1	6.261	6.161	6.361	541.540	500.000	8.3
Aroclor-1260-2	6.448	6.348	6.548	543.690	500.000	8.7
Aroclor-1260-3	6.602	6.502	6.702	521.390	500.000	4.3
Aroclor-1260-4	7.074	6.974	7.174	497.030	500.000	-0.6
Aroclor-1260-5	7.314	7.214	7.414	514.420	500.000	2.9
Decachlorobiphenyl	8.707	8.607	8.807	45.540	50.000	-8.9
Tetrachloro-m-xylene	3.695	3.594	3.794	55.390	50.000	10.8