

CALIBRATION VERIFICATION SUMMARY

Contract: ATCE02

Lab Code: CHEM Case No.: Q1438 SAS No.: Q1438 SDG NO.: Q1438

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 02/20/2025 02/20/2025

Client Sample No.: CCAL09 Date Analyzed: 03/07/2025

Lab Sample No.: AR1660CCC500 Data File : PO109698.D Time Analyzed: 08:58

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	4.789	4.691	4.891	537.160	500.000	7.4
Aroclor-1016-2	4.808	4.711	4.911	545.990	500.000	9.2
Aroclor-1016-3	4.865	4.767	4.967	538.720	500.000	7.7
Aroclor-1016-4	4.985	4.888	5.088	541.630	500.000	8.3
Aroclor-1016-5	5.243	5.145	5.345	525.230	500.000	5.0
Aroclor-1260-1	6.285	6.188	6.388	514.630	500.000	2.9
Aroclor-1260-2	6.473	6.377	6.577	508.900	500.000	1.8
Aroclor-1260-3	6.842	6.746	6.946	493.980	500.000	-1.2
Aroclor-1260-4	7.103	7.005	7.205	501.800	500.000	0.4
Aroclor-1260-5	7.344	7.247	7.447	515.340	500.000	3.1
Decachlorobiphenyl	8.753	8.657	8.857	45.820	50.000	-8.4
Tetrachloro-m-xylene	3.695	3.598	3.798	53.490	50.000	7.0

CALIBRATION VERIFICATION SUMMARY

Contract: ATCE02

Lab Code: CHEM Case No.: Q1438 SAS No.: Q1438 SDG NO.: Q1438

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 02/20/2025 02/20/2025

Client Sample No.: CCAL09 Date Analyzed: 03/07/2025

Lab Sample No.: AR1660CCC500 Data File : PO109698.D Time Analyzed: 08:58

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	4.775	4.678	4.878	558.160	500.000	11.6
Aroclor-1016-2	4.795	4.697	4.897	566.930	500.000	13.4
Aroclor-1016-3	4.969	4.872	5.072	562.100	500.000	12.4
Aroclor-1016-4	5.012	4.914	5.114	542.530	500.000	8.5
Aroclor-1016-5	5.225	5.127	5.327	541.170	500.000	8.2
Aroclor-1260-1	6.258	6.161	6.361	530.650	500.000	6.1
Aroclor-1260-2	6.446	6.348	6.548	537.600	500.000	7.5
Aroclor-1260-3	6.599	6.502	6.702	530.820	500.000	6.2
Aroclor-1260-4	7.071	6.974	7.174	519.270	500.000	3.9
Aroclor-1260-5	7.311	7.214	7.414	532.270	500.000	6.5
Decachlorobiphenyl	8.703	8.607	8.807	47.670	50.000	-4.7
Tetrachloro-m-xylene	3.693	3.594	3.794	55.440	50.000	10.9