

CALIBRATION VERIFICATION SUMMARY

Contract: ATCE02

Lab Code: CHEM Case No.: Q1438 SAS No.: Q1438 SDG NO.: Q1438

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 02/24/2025 02/24/2025

Client Sample No.: CCAL12 Date Analyzed: 03/07/2025

Lab Sample No.: AR1660CCC500 Data File : PP070359.D Time Analyzed: 15:45

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	5.678	5.581	5.781	485.280	500.000	-2.9
Aroclor-1016-2	5.700	5.602	5.802	488.960	500.000	-2.2
Aroclor-1016-3	5.762	5.665	5.865	477.110	500.000	-4.6
Aroclor-1016-4	5.860	5.762	5.962	483.790	500.000	-3.2
Aroclor-1016-5	6.153	6.056	6.256	506.220	500.000	1.2
Aroclor-1260-1	7.272	7.175	7.375	498.040	500.000	-0.4
Aroclor-1260-2	7.526	7.429	7.629	480.170	500.000	-4.0
Aroclor-1260-3	7.885	7.788	7.988	511.910	500.000	2.4
Aroclor-1260-4	8.110	8.012	8.212	500.740	500.000	0.1
Aroclor-1260-5	8.430	8.333	8.533	507.150	500.000	1.4
Decachlorobiphenyl	10.254	10.156	10.356	49.030	50.000	-1.9
Tetrachloro-m-xylene	4.525	4.427	4.627	53.530	50.000	7.1

CALIBRATION VERIFICATION SUMMARY

Contract: ATCE02

Lab Code: CHEM Case No.: Q1438 SAS No.: Q1438 SDG NO.: Q1438

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 02/24/2025 02/24/2025

Client Sample No.: CCAL12 Date Analyzed: 03/07/2025

Lab Sample No.: AR1660CCC500 Data File : PP070359.D Time Analyzed: 15:45

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.917	4.820	5.020	495.440	500.000	-0.9
Aroclor-1016-2	4.936	4.839	5.039	500.480	500.000	0.1
Aroclor-1016-3	5.113	5.017	5.217	521.100	500.000	4.2
Aroclor-1016-4	5.155	5.058	5.258	517.520	500.000	3.5
Aroclor-1016-5	5.370	5.273	5.473	523.670	500.000	4.7
Aroclor-1260-1	6.408	6.312	6.512	483.400	500.000	-3.3
Aroclor-1260-2	6.596	6.500	6.700	486.610	500.000	-2.7
Aroclor-1260-3	6.750	6.654	6.854	461.360	500.000	-7.7
Aroclor-1260-4	7.223	7.126	7.326	534.160	500.000	6.8
Aroclor-1260-5	7.463	7.367	7.567	538.180	500.000	7.6
Decachlorobiphenyl	8.883	8.788	8.988	49.460	50.000	-1.1
Tetrachloro-m-xylene	3.829	3.730	3.930	49.650	50.000	-0.7