



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Initial and Continuing Calibration Verification

Client: Parkway Generation Operating LLC

SDG No.: Q1439

Project: Linden Gen - NJDEP WW Discharge Permit

RunNo.: LB134836

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV1 Ammonia as N	mg/L	0.98	1	98	90-110	02/27/2025
Sample ID: CCV1 Ammonia as N	mg/L	0.97	1	97	90-110	02/27/2025
Sample ID: CCV2 Ammonia as N	mg/L	0.98	1	98	90-110	02/27/2025

Initial and Continuing Calibration Verification

Client: Parkway Generation Operating LLC

SDG No.: Q1439

Project: Linden Gen - NJDEP WW Discharge Permit

RunNo.: LB134838

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV1						
Bromide	mg/L	10.2	10	102	90-110	02/21/2025
Chloride	mg/L	3	3	100	90-110	02/21/2025
Fluoride	mg/L	2	2	100	90-110	02/21/2025
Nitrite	mg/L	3	3	100	90-110	02/21/2025
Nitrate	mg/L	2.5	2.5	100	90-110	02/21/2025
Sulfate	mg/L	15.1	15	101	90-110	02/21/2025
Orthophosphate as P	mg/L	5.1	5	102	90-110	02/21/2025
Sample ID: CCV1						
Bromide	mg/L	10	10	100	90-110	02/27/2025
Chloride	mg/L	3	3	100	90-110	02/27/2025
Fluoride	mg/L	2	2	100	90-110	02/27/2025
Nitrite	mg/L	3	3	100	90-110	02/27/2025
Nitrate	mg/L	2.5	2.5	100	90-110	02/27/2025
Sulfate	mg/L	14.9	15	99	90-110	02/27/2025
Orthophosphate as P	mg/L	5	5	100	90-110	02/27/2025
Sample ID: CCV2						
Bromide	mg/L	10	10	100	90-110	02/27/2025
Chloride	mg/L	3	3	100	90-110	02/27/2025
Fluoride	mg/L	2	2	100	90-110	02/27/2025
Nitrite	mg/L	3	3	100	90-110	02/27/2025
Nitrate	mg/L	2.5	2.5	100	90-110	02/27/2025
Sulfate	mg/L	14.9	15	99	90-110	02/27/2025
Orthophosphate as P	mg/L	4.7	5	94	90-110	02/27/2025

Initial and Continuing Calibration Verification

Client: Parkway Generation Operating LLC

SDG No.: Q1439

Project: Linden Gen - NJDEP WW Discharge Permit

RunNo.: LB134856

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV COD	mg/L	49.329	50	99	95-105	01/22/2025
Sample ID: CCV1 COD	mg/L	50.319	50	101	95-105	02/28/2025
Sample ID: CCV2 COD	mg/L	50.319	50	101	95-105	02/28/2025
Sample ID: CCV3 COD	mg/L	50.319	50	101	95-105	02/28/2025

Initial and Continuing Calibration Verification

Client: Parkway Generation Operating LLC

SDG No.: Q1439

Project: Linden Gen - NJDEP WW Discharge Permit

RunNo.: LB134856

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
---------	-------	--------	------------	---------------	---------------------------	------------------