

Laboratory Control Sample Summary

Client:	Parkway Generation Operating LLC	SDG No.:	Q1439
Project:	Linden Gen - NJDEP WW Discharge Permit	Run No.:	LB134810

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB134810BS							
TDS	mg/L	100	95.0		95	1	90-110	02/26/2025

Laboratory Control Sample Summary

Client:	Parkway Generation Operating LLC	SDG No.:	Q1439
Project:	Linden Gen - NJDEP WW Discharge Permit	Run No.:	LB134816

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB134816BS							
BOD5	mg/L	198	178		90	1	84.6-115.4	02/26/2025

Laboratory Control Sample Summary

Client:	Parkway Generation Operating LLC	SDG No.:	Q1439
Project:	Linden Gen - NJDEP WW Discharge Permit	Run No.:	LB134838

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB134838BSW							
Bromide	mg/L	10	10.0		100	1	90-110	02/27/2025
Chloride	mg/L	3	3.00		100	1	90-110	02/27/2025
Fluoride	mg/L	2	2.00		100	1	90-110	02/27/2025
Nitrite	mg/L	3	3.00		100	1	90-110	02/27/2025
Nitrate	mg/L	2.5	2.50		100	1	90-110	02/27/2025
Sulfate	mg/L	15	14.7		98	1	90-110	02/27/2025
Orthophosphate as P	mg/L	5	5.10		102	1	90-110	02/27/2025

Laboratory Control Sample Summary

Client:	Parkway Generation Operating LLC	SDG No.:	Q1439
Project:	Linden Gen - NJDEP WW Discharge Permit	Run No.:	LB134851

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB134851BS							
TSS	mg/L	550	533		97	1	90-110	02/28/2025

Laboratory Control Sample Summary

Client:	Parkway Generation Operating LLC	SDG No.:	Q1439
Project:	Linden Gen - NJDEP WW Discharge Permit	Run No.:	LB134856

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB134856BS							
COD	mg/L	50	49.3		99	1	90-110	02/28/2025

Laboratory Control Sample Summary

Client:	Parkway Generation Operating LLC	SDG No.:	Q1439
Project:	Linden Gen - NJDEP WW Discharge Permit	Run No.:	LB134836

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	PB166902BS							
Ammonia as N	mg/L	1	0.97		97	1	90-110	02/27/2025