



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

## **CASE NARRATIVE**

### **G Environmental**

**Project Name: Nelson**

**Project # N/A**

**Chemtech Project # Q1447**

**Test Name: Metals ICP-TAL,Mercury**

### **A. Number of Samples and Date of Receipt:**

2 Solid samples were received on 02/26/2025.

### **B. Parameters:**

According to the Chain of Custody document, the following analyses were requested: Corrosivity, EPH, Ignitability, Mercury, Metals ICP-TAL, METALS-TAL, Paint Filter, PCB, pH, RCRA CHARACTERISTICS, Reactive Cyanide, Reactive Sulfide, SVOC-TCL BNA -20, TCLP Extraction, TCLP ICP Metals, TCLP Mercury, TCLP METALS and VOC-TCLVOA-10. This data package contains results for Metals ICP-TAL,Mercury.

### **C. Analytical Techniques:**

The analysis of Metals ICP-TAL was based on method 6010D, digestion based on method 3050 (soils). The analysis and digestion of Mercury was based on method 7471B.

### **D. QA/ QC Samples:**

The Holding Times were met for all analysis.

The Blank Spike met requirements for all samples.

The Duplicate (WC1DUP) analysis met criteria for all samples except for Calcium, Iron, Nickel, Zinc due to matrix interference.

The Duplicate (WC1MSD) analysis met criteria for all samples except for Calcium, Iron, Lead, Manganese, Zinc due to matrix interference.

The Matrix Spike (WC1MS) analysis met criteria for all samples except for Antimony, Zinc due to matrix interference.

The Matrix Spike Duplicate (WC1MSD) analysis met criteria for all samples except for Antimony, Lead due to matrix interference.

The Blank analysis did not indicate the presence of lab contamination.

The Calibration met the requirements.

The Serial Dilution (WC1L) met criteria for all samples except for Iron, Lead, Manganese due to unknown interference.

### **E. Additional Comments:**

---

I certify that the data package is in compliance with the terms and conditions of the contract, both technically and for completeness, for other than the conditions detailed



above. The laboratory manager or his designee, as verified by the following signature has authorized release of the data contained in this hard copy data package.

Signature\_\_\_\_\_