



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

CASE NARRATIVE

G Environmental

Project Name: Nelson

Project # N/A

Chemtech Project # Q1447

Test Name: TCLP Mercury, TCLP ICP Metals

A. Number of Samples and Date of Receipt:

2 Solid samples were received on 02/26/2025.

B. Parameters:

According to the Chain of Custody document, the following analyses were requested: Corrosivity, EPH, Ignitability, Mercury, Metals ICP-TAL, METALS-TAL, Paint Filter, PCB, pH, RCRA CHARACTERISTICS, Reactive Cyanide, Reactive Sulfide, SVOC-TCL BNA -20, TCLP Extraction, TCLP ICP Metals, TCLP Mercury, TCLP METALS and VOC-TCLVOA-10. This data package contains results for TCLP Mercury, TCLP ICP Metals.

C. Analytical Techniques:

The analysis of TCLP ICP Metals was based on method 6010D, digestion based on method 3010 (waters). The analysis and digestion of TCLP Mercury was based on method 7470A and TCLP extraction method was 1311.

D. QA/ QC Samples:

The Holding Times were met for all analysis.

The Blank Spike met requirements for all samples.

The Duplicate (WC1DUP) analysis met criteria for all samples except for Barium, Lead due to matrix interference.

The Matrix Spike (WC1MS) analysis met criteria for all samples except for Barium due to matrix interference.

The Matrix Spike Duplicate (WC1MSD) analysis met criteria for all samples except for Barium due to matrix interference.

The Blank analysis did not indicate the presence of lab contamination.

The Calibration met the requirements.

The Serial Dilution met the acceptable requirements.

E. Additional Comments:

I certify that the data package is in compliance with the terms and conditions of the contract, both technically and for completeness, for other than the conditions detailed



above. The laboratory manager or his designee, as verified by the following signature has authorized release of the data contained in this hard copy data package.

Signature_____