

Initial and Continuing Calibration Verification

Client: G Environmental

SDG No.: Q1447

Project: Nelson

RunNo.: LB134827

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV pH	pH	7.00	7	100	90-110	02/27/2025
Sample ID: CCV1 pH	pH	2.01	2.00	101	90-110	02/27/2025
Sample ID: CCV2 pH	pH	12.02	12.00	100	90-110	02/27/2025

Initial and Continuing Calibration Verification

Client: G Environmental

SDG No.: Q1447

Project: Nelson

RunNo.: LB134831

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV1 Reactive Cyanide	mg/L	0.096	0.099	97	85-115	02/27/2025
Sample ID: CCV1 Reactive Cyanide	mg/L	0.24	0.25	96	90-110	02/27/2025
Sample ID: CCV2 Reactive Cyanide	mg/L	0.24	0.25	96	90-110	02/27/2025
Sample ID: CCV3 Reactive Cyanide	mg/L	0.24	0.25	96	90-110	02/27/2025