



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 2b -

CRDL STANDARD FOR AA & ICP

Client: G Environmental SDG No.: Q1447
Contract: GENV01 Lab Code: CHEM Case No.: Q1447 SAS No.: Q1447
Initial Calibration Source: _____
Continuing Calibration Source: _____

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CRA	Mercury	0.18	0.2	91	40 - 160	CV	02/27/2025	10:04	LB134824
CRI01	Aluminum	102	100	102	40 - 160	P	03/05/2025	11:00	LB134904
	Antimony	51.7	50.0	103	40 - 160	P	03/05/2025	11:00	LB134904
	Arsenic	20.2	20.0	101	40 - 160	P	03/05/2025	11:00	LB134904
	Barium	91.8	100	92	40 - 160	P	03/05/2025	11:00	LB134904
	Beryllium	5.92	6.0	99	40 - 160	P	03/05/2025	11:00	LB134904
	Cadmium	5.85	6.0	98	40 - 160	P	03/05/2025	11:00	LB134904
	Calcium	1970	2000	98	40 - 160	P	03/05/2025	11:00	LB134904
	Chromium	10.5	10.0	105	40 - 160	P	03/05/2025	11:00	LB134904
	Cobalt	29.4	30.0	98	40 - 160	P	03/05/2025	11:00	LB134904
	Copper	22.5	20.0	112	40 - 160	P	03/05/2025	11:00	LB134904
	Iron	93.2	100	93	40 - 160	P	03/05/2025	11:00	LB134904
	Lead	12.3	12.0	103	40 - 160	P	03/05/2025	11:00	LB134904
	Magnesium	2080	2000	104	40 - 160	P	03/05/2025	11:00	LB134904
	Manganese	19.8	20.0	99	40 - 160	P	03/05/2025	11:00	LB134904
	Nickel	39.1	40.0	98	40 - 160	P	03/05/2025	11:00	LB134904
	Potassium	1920	2000	96	40 - 160	P	03/05/2025	11:00	LB134904
	Selenium	18.9	20.0	95	40 - 160	P	03/05/2025	11:00	LB134904
	Silver	10.5	10.0	105	40 - 160	P	03/05/2025	11:00	LB134904
	Sodium	1830	2000	92	40 - 160	P	03/05/2025	11:00	LB134904
	Thallium	41.3	40.0	103	40 - 160	P	03/05/2025	11:00	LB134904
	Vanadium	40.1	40.0	100	40 - 160	P	03/05/2025	11:00	LB134904
	Zinc	42.9	40.0	107	40 - 160	P	03/05/2025	11:00	LB134904