


LB134848

7470A

INSTRUMENT ID: CV1



Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	
0.0	0.000	-0.042	-0.042	26	0.000	26	0				0/00
0.2	0.200	0.179	-0.021	980	0.0 %	980					-11
2.5	2.500	2.456	-0.044	10827	0.0 %	10827					-2
5.0	5.000	5.254	0.254	22926	0.0 %	22926					5
7.5	7.500	7.409	-0.091	32245	0.0 %	32245					-1
10.0	10.000	9.945	-0.055	43212	0.0 %	43212					-1

LB134848 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 28 Feb 2025 09:52:37

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	26	-	0.0000	7470A	PPB	28 Feb 2025 10:21:16	S	Std
0.2 - 1	50.2	980	-	0.2000	7470A	PPB	28 Feb 2025 10:23:32	S	Std
2.5 - 1	52.5	10827	-	2.5000	7470A	PPB	28 Feb 2025 10:25:49	S	Std
5.0 - 1	55	22926	-	5.0000	7470A	PPB	28 Feb 2025 10:28:07	S	Std
7.5 - 1	57.5	32245	-	7.5000	7470A	PPB	28 Feb 2025 10:30:24	S	Std
10.0 - 1	510	43212	-	10.0000	7470A	PPB	28 Feb 2025 10:32:45	S	Std
ICV22 - 1	ICV22	17432	3.9832	-	7470A	PPB	28 Feb 2025 10:35:35	U	SMPL
ICB22 - 1	ICB22	-25	-0.0538	-	7470A	PPB	28 Feb 2025 10:40:25	U	SMPL
CCV35 - 1	CCV35	23495	5.3853	-	7470A	PPB	28 Feb 2025 10:42:39	U	SMPL
CCB35 - 1	CCB35	44	-0.0378	-	7470A	PPB	28 Feb 2025 10:47:32	U	SMPL
CRA - 1	CRA	1042	0.1930	-	7470A	PPB	28 Feb 2025 10:49:46	U	SMPL
HighStd - 1	HighStd	44319	10.2010	-	7470A	PPB	28 Feb 2025 10:52:02	U	SMPL
ChkStd - 1	ChkStd	29757	6.8335	-	7470A	PPB	28 Feb 2025 10:54:20	U	SMPL
PB166936BL - 1	PBW	-214	-0.0975	-	7470A	PPB	28 Feb 2025 10:56:38	U	SMPL
PB166936BS - 1	LCSW	17914	4.0947	-	7470A	PPB	28 Feb 2025 10:58:57	U	SMPL
Q1427-02 - 1	VNJ-227	-18	-0.0522	-	7470A	PPB	28 Feb 2025 11:01:13	U	SMPL
Q1442-01 - 1	280	773	0.1308	-	7470A	PPB	28 Feb 2025 11:03:32	U	SMPL
Q1442-02 - 1	281	241	0.0077	-	7470A	PPB	28 Feb 2025 11:05:49	U	SMPL
Q1443-01 - 1	3733	336	0.0297	-	7470A	PPB	28 Feb 2025 11:08:09	U	SMPL
Q1447-02 - 1	WC1	326	0.0274	-	7470A	PPB	28 Feb 2025 11:10:24	U	SMPL
CCV36 - 1	CCV36	20854	4.7746	-	7470A	PPB	28 Feb 2025 11:16:31	U	SMPL
CCB36 - 1	CCB36	-5	-0.0492	-	7470A	PPB	28 Feb 2025 11:18:46	U	SMPL
Q1447-02DUP - 1	WC1DUP	224	0.0038	-	7470A	PPB	28 Feb 2025 11:21:04	U	SMPL
Q1447-02MS - 1	WC1MS	16346	3.7321	-	7470A	PPB	28 Feb 2025 11:23:19	U	SMPL
Q1447-02MSD - 1	WC1MSD	16170	3.6914	-	7470A	PPB	28 Feb 2025 11:25:35	U	SMPL
PB166937BL - 1	PBW	2	-0.0475	-	7470A	PPB	28 Feb 2025 11:27:53	U	SMPL
PB166937BS - 1	LCSW	16019	3.6565	-	7470A	PPB	28 Feb 2025 11:30:10	U	SMPL
Q1435-01 - 1	286107	-15	-0.0515	-	7470A	PPB	28 Feb 2025 11:32:26	U	SMPL
Q1439-02 - 1	LRSA-MOD	116	-0.0212	-	7470A	PPB	28 Feb 2025 11:34:45	U	SMPL
Q1439-02DUP - 1	LRSA-MODDUP	147	-0.0140	-	7470A	PPB	28 Feb 2025 11:37:02	U	SMPL
Q1439-02MS - 1	LRSA-MODMS	15345	3.5006	-	7470A	PPB	28 Feb 2025 11:39:19	U	SMPL
Q1439-02MSD - 1	LRSA-MODMSD	15342	3.4999	-	7470A	PPB	28 Feb 2025 11:41:34	U	SMPL
CCV37 - 1	CCV37	20724	4.7445	-	7470A	PPB	28 Feb 2025 11:43:51	U	SMPL
CCB37 - 1	CCB37	213	0.0013	-	7470A	PPB	28 Feb 2025 11:48:51	U	SMPL
Q1442-05 - 1	278	181	-0.0061	-	7470A	PPB	28 Feb 2025 11:51:06	U	SMPL
PB166879TB - 1	PB166879TB	119	-0.0205	-	7470A	PPB	28 Feb 2025 11:53:22	U	SMPL
Q1447-02LX5 - 1		64	-0.0332	-	7470A	PPB	28 Feb 2025 11:55:38	U	SMPL
Q1447-02A - 1		14408	3.2839	-	7470A	PPB	28 Feb 2025 11:57:53	U	SMPL
Q1439-02LX5 - 1		-190	-0.0919	-	7470A	PPB	28 Feb 2025 12:00:09	U	SMPL
Q1439-02A - 1		15606	3.5610	-	7470A	PPB	28 Feb 2025 12:02:28	U	SMPL
CCV38 - 1	CCV38	20161	4.6143	-	7470A	PPB	28 Feb 2025 12:06:58	U	SMPL
CCB38 - 1	CCB38	-161	-0.0852	-	7470A	PPB	28 Feb 2025 12:09:15	U	SMPL