



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 2b -

CRDL STANDARD FOR AA & ICP

Client: G Environmental **SDG No.:** Q1447
Contract: GENV01 **Lab Code:** CHEM **Case No.:** Q1447 **SAS No.:** Q1447
Initial Calibration Source: _____
Continuing Calibration Source: _____

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CRA	Mercury	0.19	0.2	96	40 - 160	CV	02/28/2025	10:49	LB134848
CRI01	Arsenic	18.7	20.0	94	40 - 160	P	03/03/2025	19:30	LB134881
	Barium	87.6	100	88	40 - 160	P	03/03/2025	19:30	LB134881
	Cadmium	5.64	6.0	94	40 - 160	P	03/03/2025	19:30	LB134881
	Chromium	9.59	10.0	96	40 - 160	P	03/03/2025	19:30	LB134881
	Lead	11.5	12.0	96	40 - 160	P	03/03/2025	19:30	LB134881
	Selenium	18.0	20.0	90	40 - 160	P	03/03/2025	19:30	LB134881
	Silver	9.81	10.0	98	40 - 160	P	03/03/2025	19:30	LB134881
CRI01	Arsenic	20.2	20.0	101	40 - 160	P	03/05/2025	11:00	LB134904
	Barium	91.8	100	92	40 - 160	P	03/05/2025	11:00	LB134904
	Cadmium	5.85	6.0	98	40 - 160	P	03/05/2025	11:00	LB134904
	Chromium	10.5	10.0	105	40 - 160	P	03/05/2025	11:00	LB134904
	Lead	12.3	12.0	103	40 - 160	P	03/05/2025	11:00	LB134904
	Selenium	18.9	20.0	95	40 - 160	P	03/05/2025	11:00	LB134904
	Silver	10.5	10.0	105	40 - 160	P	03/05/2025	11:00	LB134904