



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client:	G Environmental				SDG No.:	Q1447				
Contract:	GENV01	Lab Code:	CHEM		Case No.:	Q1447	SAS No.:	Q1447		
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number	
ICB22	Mercury	0.20	+/-0.20	U	0.20	CV	02/28/2025	10:40	LB134848	

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: G Environmental

SDG No.: Q1447

Contract: GENV01

Lab Code: CHEM

Case No.: Q1447

SAS No.: Q1447

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB35	Mercury	0.20	+/-0.20	U	0.20	CV	02/28/2025	10:47	LB134848
CCB36	Mercury	0.20	+/-0.20	U	0.20	CV	02/28/2025	11:18	LB134848
CCB37	Mercury	0.20	+/-0.20	U	0.20	CV	02/28/2025	11:48	LB134848
CCB38	Mercury	0.20	+/-0.20	U	0.20	CV	02/28/2025	12:09	LB134848

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: G Environmental

SDG No.: Q1447

Contract: GENV01

Lab Code: CHEM

Case No.: Q1447

SAS No.: Q1447

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB01	Arsenic	20.0	+/-20.0	U	20.0	P	03/03/2025	19:25	LB134881
	Barium	100	+/-100	U	100	P	03/03/2025	19:25	LB134881
	Cadmium	6.00	+/-6.00	U	6.00	P	03/03/2025	19:25	LB134881
	Chromium	10.0	+/-10.0	U	10.0	P	03/03/2025	19:25	LB134881
	Lead	12.0	+/-12.0	U	12.0	P	03/03/2025	19:25	LB134881
	Selenium	20.0	+/-20.0	U	20.0	P	03/03/2025	19:25	LB134881
	Silver	10.0	+/-10.0	U	10.0	P	03/03/2025	19:25	LB134881

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: G Environmental		SDG No.: Q1447							
Contract: GENV01	Lab Code: CHEM	Case No.: Q1447	SAS No.: Q1447						
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB01	Arsenic	20.0	+/-20.0	U	20.0	P	03/03/2025	20:01	LB134881
	Barium	100	+/-100	U	100	P	03/03/2025	20:01	LB134881
	Cadmium	6.00	+/-6.00	U	6.00	P	03/03/2025	20:01	LB134881
	Chromium	10.0	+/-10.0	U	10.0	P	03/03/2025	20:01	LB134881
	Lead	12.0	+/-12.0	U	12.0	P	03/03/2025	20:01	LB134881
	Selenium	20.0	+/-20.0	U	20.0	P	03/03/2025	20:01	LB134881
	Silver	2.51	+/-10.0	J	10.0	P	03/03/2025	20:01	LB134881
CCB02	Arsenic	20.0	+/-20.0	U	20.0	P	03/03/2025	21:16	LB134881
	Barium	100	+/-100	U	100	P	03/03/2025	21:16	LB134881
	Cadmium	6.00	+/-6.00	U	6.00	P	03/03/2025	21:16	LB134881
	Chromium	10.0	+/-10.0	U	10.0	P	03/03/2025	21:16	LB134881
	Lead	12.0	+/-12.0	U	12.0	P	03/03/2025	21:16	LB134881
	Selenium	20.0	+/-20.0	U	20.0	P	03/03/2025	21:16	LB134881
	Silver	10.0	+/-10.0	U	10.0	P	03/03/2025	21:16	LB134881
CCB03	Arsenic	20.0	+/-20.0	U	20.0	P	03/03/2025	22:24	LB134881
	Barium	100	+/-100	U	100	P	03/03/2025	22:24	LB134881
	Cadmium	6.00	+/-6.00	U	6.00	P	03/03/2025	22:24	LB134881
	Chromium	10.0	+/-10.0	U	10.0	P	03/03/2025	22:24	LB134881
	Lead	12.0	+/-12.0	U	12.0	P	03/03/2025	22:24	LB134881
	Selenium	20.0	+/-20.0	U	20.0	P	03/03/2025	22:24	LB134881
	Silver	10.0	+/-10.0	U	10.0	P	03/03/2025	22:24	LB134881
CCB04	Arsenic	20.0	+/-20.0	U	20.0	P	03/03/2025	23:11	LB134881
	Barium	100	+/-100	U	100	P	03/03/2025	23:11	LB134881
	Cadmium	6.00	+/-6.00	U	6.00	P	03/03/2025	23:11	LB134881
	Chromium	10.0	+/-10.0	U	10.0	P	03/03/2025	23:11	LB134881
	Lead	12.0	+/-12.0	U	12.0	P	03/03/2025	23:11	LB134881
	Selenium	20.0	+/-20.0	U	20.0	P	03/03/2025	23:11	LB134881
	Silver	10.0	+/-10.0	U	10.0	P	03/03/2025	23:11	LB134881
CCB05	Arsenic	20.0	+/-20.0	U	20.0	P	03/04/2025	00:08	LB134881
	Barium	100	+/-100	U	100	P	03/04/2025	00:08	LB134881
	Cadmium	6.00	+/-6.00	U	6.00	P	03/04/2025	00:08	LB134881
	Chromium	10.0	+/-10.0	U	10.0	P	03/04/2025	00:08	LB134881
	Lead	12.0	+/-12.0	U	12.0	P	03/04/2025	00:08	LB134881
	Selenium	20.0	+/-20.0	U	20.0	P	03/04/2025	00:08	LB134881
	Silver	10.0	+/-10.0	U	10.0	P	03/04/2025	00:08	LB134881

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: <u>G Environmental</u>		SDG No.: <u>Q1447</u>							
Contract: <u>GENV01</u>		Lab Code: <u>CHEM</u>		Case No.: <u>Q1447</u>		SAS No.: <u>Q1447</u>			
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB01	Arsenic	20.0	+/-20.0	U	20.0	P	03/05/2025	10:56	LB134904
	Barium	100	+/-100	U	100	P	03/05/2025	10:56	LB134904
	Cadmium	6.00	+/-6.00	U	6.00	P	03/05/2025	10:56	LB134904
	Chromium	10.0	+/-10.0	U	10.0	P	03/05/2025	10:56	LB134904
	Lead	12.0	+/-12.0	U	12.0	P	03/05/2025	10:56	LB134904
	Selenium	20.0	+/-20.0	U	20.0	P	03/05/2025	10:56	LB134904
	Silver	10.0	+/-10.0	U	10.0	P	03/05/2025	10:56	LB134904

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: <u>G Environmental</u>		SDG No.: <u>Q1447</u>							
Contract: <u>GENV01</u>	Lab Code: <u>CHEM</u>	Case No.: <u>Q1447</u>	SAS No.: <u>Q1447</u>						
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB01	Arsenic	20.0	+/-20.0	U	20.0	P	03/05/2025	11:17	LB134904
	Barium	100	+/-100	U	100	P	03/05/2025	11:17	LB134904
	Cadmium	6.00	+/-6.00	U	6.00	P	03/05/2025	11:17	LB134904
	Chromium	10.0	+/-10.0	U	10.0	P	03/05/2025	11:17	LB134904
	Lead	12.0	+/-12.0	U	12.0	P	03/05/2025	11:17	LB134904
	Selenium	20.0	+/-20.0	U	20.0	P	03/05/2025	11:17	LB134904
	Silver	10.0	+/-10.0	U	10.0	P	03/05/2025	11:17	LB134904
CCB02	Arsenic	20.0	+/-20.0	U	20.0	P	03/05/2025	11:59	LB134904
	Barium	100	+/-100	U	100	P	03/05/2025	11:59	LB134904
	Cadmium	6.00	+/-6.00	U	6.00	P	03/05/2025	11:59	LB134904
	Chromium	10.0	+/-10.0	U	10.0	P	03/05/2025	11:59	LB134904
	Lead	12.0	+/-12.0	U	12.0	P	03/05/2025	11:59	LB134904
	Selenium	20.0	+/-20.0	U	20.0	P	03/05/2025	11:59	LB134904
	Silver	10.0	+/-10.0	U	10.0	P	03/05/2025	11:59	LB134904
CCB03	Arsenic	20.0	+/-20.0	U	20.0	P	03/05/2025	12:49	LB134904
	Barium	100	+/-100	U	100	P	03/05/2025	12:49	LB134904
	Cadmium	6.00	+/-6.00	U	6.00	P	03/05/2025	12:49	LB134904
	Chromium	10.0	+/-10.0	U	10.0	P	03/05/2025	12:49	LB134904
	Lead	12.0	+/-12.0	U	12.0	P	03/05/2025	12:49	LB134904
	Selenium	20.0	+/-20.0	U	20.0	P	03/05/2025	12:49	LB134904
	Silver	10.0	+/-10.0	U	10.0	P	03/05/2025	12:49	LB134904
CCB04	Arsenic	20.0	+/-20.0	U	20.0	P	03/05/2025	14:11	LB134904
	Barium	100	+/-100	U	100	P	03/05/2025	14:11	LB134904
	Cadmium	6.00	+/-6.00	U	6.00	P	03/05/2025	14:11	LB134904
	Chromium	10.0	+/-10.0	U	10.0	P	03/05/2025	14:11	LB134904
	Lead	12.0	+/-12.0	U	12.0	P	03/05/2025	14:11	LB134904
	Selenium	20.0	+/-20.0	U	20.0	P	03/05/2025	14:11	LB134904
	Silver	10.0	+/-10.0	U	10.0	P	03/05/2025	14:11	LB134904